MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
STATE PUBLIC HEALTH LABORATORY
BREATH ALCOHOL PROGRAM

DATAMASTER MAINTENANCE REPORT

RECEIVED
By Carol Day at 3:14 pm, Mar 19, 2014

REPORT #6

Complete this report at the time of the regular monthly preventive maintenance check (not to exceed 35 days).
Complete this report whenever the instrument is serviced or repaired and whenever it is placed into service.
Retain the original and send a copy within 15 days to the Breath Alcohol Program, DHSS.

DATAMASTER SN	NAME OF AGENCY	DATE OF INSPECTION
201247	Eldon Police Department	03/06/2014
LOCATION OF INSTRUMENT (STREET AND CITY)		TIME OF INSPECTION
111 South Oak, Eldon Mo 65026		8:33 am

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

- [x] DIAGNOSTIC CHECK (PRINTOUT ATTACHED) — DATE AND TIME (from printout) 3/6/2014 08:33
  - [x] COMPUTER
  - [x] DETECTOR
  - [x] PROGRAM
  - [x] FILTERS
  - [x] HEATERS SAMPLE CHAMBER 48 °C
  - [x] QUARTZ STANDARD
  - [x] FLOW DETECTOR
  - [x] CALIBRATION
  - [x] PUMP HIGH SPEED
  - [x] PRINTER
- [x] INDICATOR LIGHTS
- [x] SIMULATOR SOLUTION SUPPLIER Guth LOT # 13280 EXP. DATE 10/16/2015
- [x] SIMULATOR TEMP (34°C ± 0.2°C) 34.0 °C SIMULATOR SN SD2999 EXP. DATE 09/09/2014
- [x] CALIBRATION CHECK – **(ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)**

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Mark the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

- [x] 0.100% STANDARD - MUST READ BETWEEN 0.095% AND 0.105% INCLUSIVE
- [ ] 0.080% STANDARD - MUST READ BETWEEN 0.076% AND 0.084% INCLUSIVE
- [ ] 0.040% STANDARD - MUST READ BETWEEN 0.038% AND 0.042% INCLUSIVE

TEST 1	TEST 2	TEST 3
.098	.100	.101

- [x] PERFORM R.F.I. TEST (PRINTOUT ATTACHED)

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS	(0-.04)	(.05-.09)	(.10-.14)	(.15-.19)	OVER .19
2			1		

LIST ANY NEW PARTS AND DESCRIBE ANY ALTERATION OR MODIFICATION THAT WAS MADE TO RESTORE THE INSTRUMENT TO OPERATE SATISFACTORILY AND WITHIN ESTABLISHED LIMITS (USE OTHER SIDE IF NECESSARY).

New Printer installed by Welsh at Safety Center,

INSPECTING OFFICER

SIGNATURE	PRINT FULL NAME
[signature]	Brian McLuel
TYPE II PERMIT NUMBER/EXPIRATION DATE	TELEPHONE NUMBER
12-11-15	280-0555

RETURN COMPLETED REPORT TO THE: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
2875 James Blvd.
Poplar Bluff, MO 63901

ACE THIS SIDE DOWN - THIS EDGE IN FIRST

BAC DataMaster
Evidence Ticket

STATE OF MISSOURI
ELDON POLICE DEPARTMENT

BAC DATAMASTER SERIAL NUMBER 201274
03/06/14

ARREST TIME: 08:00
SUBJECT NAME:
TEST/RFI
DOB: 03/06/14 SEX: M
STATE/D.L.: MO/123456
ARRESTING OFFICER:
KIDWELL/BRIAN
OFFICER I.D.: 806
TESTING OFFICER:
KIDWELL/BRIAN
OFFICER I.D.: 806
PERMIT NUMBER: 230304
EXPIRATION DATE: 12/11/15
MISCELLANEOUS DATA:

--- BREATH ANALYSIS ---

BLANK TEST	.000	08:44
INTERNAL STANDARD	VERIFIED	08:44
RADIO INTERFERENCE		

ATOR SIGNATURE

Stock No.

REORDER ALL SUPPLIES FROM N.P.A.S.
P.O. BOX 1435, MANSFIELD, OH 44901 PHONE 1-800-800-8143 (NPAS)

BAC DataMaster
Evidence Ticket

STATE OF MISSOURI
ELDON POLICE DEPARTMENT

BAC DATAMASTER SERIAL NUMBER 201274
03/06/14

ARREST TIME: 08:00
SUBJECT NAME:
KIDWELL/BRIAN
DOB: 12/12/13 SEX: M
STATE/D.L.: MO/123456
ARRESTING OFFICER:
KIDWELL/BRIAN
OFFICER I.D.: 806
TESTING OFFICER:
KIDWELL/BRIAN
OFFICER I.D.: 806
PERMIT NUMBER: 230304
EXPIRATION DATE: 12/11/15
MISCELLANEOUS DATA:

--- BREATH ANALYSIS ---

BLANK TEST	.000	08:47
INTERNAL STANDARD	VERIFIED	08:47
SUBJECT SAMPLE	.000	08:47
BLANK TEST	.000	08:48

OPERATOR SIGNATURE

Card Stock No.
60020

REORDER ALL SUPPLIES FROM N.P.A.S.
P.O. BOX 1435, MANSFIELD, OH 44901 PHONE 1-800-800-8143 (NPAS)

FACE THIS SIDE DOWN - THIS EDGE IN FIRST

BAC DataMaster
Evidence Ticket

STATE OF MISSOURI
ELDON POLICE DEPARTMENT

BAC DATAMASTER SERIAL NUMBER 201274
03/06/14
08:33

--- DIAGNOSTIC CHECK ---

COMPUTER:	OKAY
PROGRAM (04-07-2009):	OKAY
HEATERS SAMPLE CHAMBER:	48c
FLOW DETECTOR:	OKAY
PUMP HIGH SPEED:	OKAY
DETECTOR:	OKAY
FILTERS:	OKAY
QUARTZ STANDARD:	OKAY
CALIBRATION:	OKAY

PRINTER TEST
!"#$%&'()*+,-./0123456789:;<=>?@ABCDEFG
HIJKLMNOPQRSTUVWXYZ[\]^_`abcdefghijklmno
pqrstuvwxyz{|}→¯

OPERATOR SIGNATURE

Card Stock No.

FACE THIS SIDE DOWN - THIS EDGE IN FIRST

BAC DataMaster
Evidence Ticket

STATE OF MISSOURI
ELDON POLICE DEPARTMENT

BAC DATAMASTER SERIAL NUMBER 201274
03/06/14

TESTING OFFICER:
KIDWELL/BRIAN/D
OFFICER I.D.: 806
PERMIT NUMBER: 230304
EXPIRATION DATE: 12/11/15
MISCELLANEOUS DATA:

--- SUPERVISOR MODE ---

BLANK TEST	.000	08:36
INTERNAL STANDARD	VERIFIED	08:36
EXTERNAL STANDARD	.098	08:37
BLANK TEST	.000	08:37
EXTERNAL STANDARD	.100	08:38
BLANK TEST	.000	08:38
EXTERNAL STANDARD	.101	08:39
BLANK TEST	.000	08:39

N = 3
SIM. = .1
AVG. = .0996

OPERATOR SIGNATURE

Card Stock No.