



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
STATE PUBLIC HEALTH LABORATORY
BREATH ALCOHOL PROGRAM
DATAMASTER MAINTENANCE REPORT

RECEIVED

By Carol Day at 2:38 pm, Mar 31, 2014

Complete this report at the time of the regular monthly preventive maintenance check (not to exceed 35 days).
Complete this report whenever the instrument is serviced or repaired and whenever it is placed into service.
Retain the original and send a copy within 15 days to the Breath Alcohol Program, DHSS.

DATAMASTER SN 201253	NAME OF AGENCY MINER POLICE DEPARTMENT	DATE OF INSPECTION 03/26/2014			
LOCATION OF INSTRUMENT (STREET AND CITY) 103 STATE HIGHWAY H		TIME OF INSPECTION 10:29 am			
CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.					
<input checked="" type="checkbox"/> DIAGNOSTIC CHECK (PRINTOUT ATTACHED)		DATE AND TIME (from printout) 03-26-2014 10:32			
<input checked="" type="checkbox"/> COMPUTER	<input checked="" type="checkbox"/> DETECTOR				
<input checked="" type="checkbox"/> PROGRAM	<input checked="" type="checkbox"/> FILTERS				
<input checked="" type="checkbox"/> HEATERS SAMPLE CHAMBER	34.2 °C	<input checked="" type="checkbox"/> QUARTZ STANDARD			
<input checked="" type="checkbox"/> FLOW DETECTOR	<input checked="" type="checkbox"/> CALIBRATION				
<input checked="" type="checkbox"/> PUMP HIGH SPEED	<input checked="" type="checkbox"/> PRINTER				
<input checked="" type="checkbox"/> INDICATOR LIGHTS					
<input checked="" type="checkbox"/> SIMULATOR SOLUTION SUPPLIER REPCO		LOT # 13002 EXP. DATE 06/29/2015			
<input checked="" type="checkbox"/> SIMULATOR TEMP (34°C ± 0.2°C) 34.2 °C		SIMULATOR SN _____ EXP. DATE _____			
<input checked="" type="checkbox"/> CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)					
Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Mark the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)					
<input checked="" type="checkbox"/> 0.100% STANDARD - MUST READ BETWEEN 0.095% AND 0.105% INCLUSIVE					
<input type="checkbox"/> 0.080% STANDARD - MUST READ BETWEEN 0.076% AND 0.084% INCLUSIVE					
<input type="checkbox"/> 0.040% STANDARD - MUST READ BETWEEN 0.038% AND 0.042% INCLUSIVE					
TEST 1 → .104	TEST 2 → .103	TEST 3 → .102			
<input checked="" type="checkbox"/> PERFORM R.F.I. TEST (PRINTOUT ATTACHED)					
INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)					
REFUSALS 2	(0-.04)	(.05-.09) 1	(.10-.14)	(.15-.19)	OVER .19
LIST ANY NEW PARTS AND DESCRIBE ANY ALTERATION OR MODIFICATION THAT WAS MADE TO RESTORE THE INSTRUMENT TO OPERATE SATISFACTORILY AND WITHIN ESTABLISHED LIMITS (USE OTHER SIDE IF NECESSARY).					
CERTIFIED USING REPCO .010% SOLUTION					
LOT # 13002					
EXP DATE 06-29-2015					
INSPECTING OFFICER					
SIGNATURE 			PRINT FULL NAME CHRIS GRIGGS		
TYPE II PERMIT NUMBER/EXPIRATION DATE 220094 04/12/2014			TELEPHONE NUMBER (573) 471-8568		
RETURN COMPLETED REPORT TO THE: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office 2875 James Blvd. Poplar Bluff, MO 63901					

REP CO MARKETING INC.3101-188 STONVEROCK DRIVE
RALEIGH, N.C. 27604
919-876-8480**CERTIFICATE OF ANALYSIS**

MANUFACTURER AND SUPPLIER: RepCo Marketing, Inc.
LOT NUMBER: 13002
EXPIRATION DATE: June 19, 2015 at 11:59 p.m.

RepCo Marketing, Inc. certifies the following:


RepCo Marketing, Inc. manufactured, tested and supplied Lot Number 13002 of Alcohol Certified Solution for simulators. Random samples of said lot number were analyzed by an independent laboratory utilizing a gas chromatograph and found to contain .1217 gms/dl +/- .003 gms/dl wt/vol ethanol (95% Confidence).

The alcohol and distilled water used in the solution were found to be free of any interfering substance.

This solution will produce a vapor alcohol value of .100 +/- 3% gms/210L Breath when heated to 34 Degrees Celsius +/- 0.2 Degrees Celsius in a simulator (95% Confidence).

The date of manufacture for this lot number is June 20, 2013
The expiration date for this lot number is June 19, 2015 at 11:59 p.m.

This document is a true representation of the original Certificate of Analysis.



Cecil B. Garner, President
RepCo Marketing, Inc.

Face This Side Down - This Edge In First

**BAC DataMaster
Evidence Ticket**

STATE OF MISSOURI
MINER POLICE DEPARTMENT

BAC DATAMASTER SERIAL NUMBER 201253
03/26/14

TESTING OFFICER:
GRIGGS
OFFICER I.D.: 520
PERMIT NUMBER: 220094
EXPIRATION DATE: 04/12/14
MISCELLANEOUS DATA:
TST

--- SUPERVISOR MODE ---

BLANK TEST	.000	10:26
INTERNAL STANDARD	VERIFIED	10:26
EXTERNAL STANDARD	.104	10:26
BLANK TEST	.000	10:27
EXTERNAL STANDARD	.103	10:27
BLANK TEST	.000	10:28
EXTERNAL STANDARD	.102	10:29
BLANK TEST	.000	10:29

N = 3
SIM. = .1
AVG. = .103



Operator Signature

Face This Side Down - This Edge In First

**BAC DataMaster
Evidence Ticket**

STATE OF MISSOURI
MINER POLICE DEPARTMENT

BAC DATAMASTER SERIAL NUMBER 201253
03/26/14
10:32

--- DIAGNOSTIC CHECK ---

COMPUTER:	OKAY
PROGRAM (04-07-2009):	OKAY
HEATERS	
SAMPLE CHAMBER:	49c
FLOW DETECTOR:	OKAY
PUMP	
HIGH SPEED:	OKAY
DETECTOR:	OKAY
FILTERS:	OKAY
QUARTZ STANDARD:	OKAY
CALIBRATION:	OKAY

PRINTER TEST

!"#\$%&'()*+,-./0123456789:;{<=>?@ABCDEFGHIJK
LMNOPQRSTUVWXYZ[\]^_`abcdefghijklmnopqrstuvwxyz{|}~



Operator Signature

Face This Side Down - This Edge In First

**BAC DataMaster
Evidence Ticket**

STATE OF MISSOURI
MINER POLICE DEPARTMENT

BAC DATAMASTER SERIAL NUMBER 201253
03/26/14

ARREST TIME: 09:00
SUBJECT NAME:
JJ
DOB: 09/09/09 SEX: M
STATE/D.L.: JJ/8
ARRESTING OFFICER:
U
OFFICER I.D.: 8
TESTING OFFICER:
U
OFFICER I.D.: 8
PERMIT NUMBER: 999999
EXPIRATION DATE: 09/09/09
MISCELLANEOUS DATA:
RFI

--- BREATH ANALYSIS ---

RADIO INTERFERENCE



Operator Signature