



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
STATE PUBLIC HEALTH LABORATORY
BREATH ALCOHOL PROGRAM
DATAMASTER MAINTENANCE REPORT

RECEIVED

By Carol Day at 3:37 pm, Apr 28, 2014

REPORT #

Complete this report at the time of the regular monthly preventive maintenance check (not to exceed 35 days).
Complete this report whenever the instrument is serviced or repaired and whenever it is placed into service.
Retain the original and send a copy within 15 days to the Breath Alcohol Program, DHSS.

DATAMASTER SN 201253	NAME OF AGENCY MINER POLICE DEPT	DATE OF INSPECTION 04/27/2014
LOCATION OF INSTRUMENT (STREET AND CITY) 103 ST. HWY H, MINER, MO. 63801		TIME OF INSPECTION 10:14 am

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

<input checked="" type="checkbox"/> DIAGNOSTIC CHECK (PRINTOUT ATTACHED)	DATE AND TIME (from printout) <u>04/27/14 10:15</u>
<input checked="" type="checkbox"/> COMPUTER	<input checked="" type="checkbox"/> DETECTOR
<input checked="" type="checkbox"/> PROGRAM	<input checked="" type="checkbox"/> FILTERS
<input checked="" type="checkbox"/> HEATERS SAMPLE CHAMBER _____ 49°C	<input checked="" type="checkbox"/> QUARTZ STANDARD
<input checked="" type="checkbox"/> FLOW DETECTOR	<input checked="" type="checkbox"/> CALIBRATION
<input checked="" type="checkbox"/> PUMP HIGH SPEED	<input checked="" type="checkbox"/> PRINTER

<input checked="" type="checkbox"/> INDICATOR LIGHTS
<input checked="" type="checkbox"/> SIMULATOR SOLUTION SUPPLIER <u>REPCO MARKETING</u> LOT # <u>13002</u> EXP. DATE <u>06/19/2015</u>
<input checked="" type="checkbox"/> SIMULATOR TEMP (34°C ± 0.2°C) <u>34.0</u> °C SIMULATOR SN <u>SD2273</u> EXP. DATE <u>01/03/2015</u>

CALIBRATION CHECK - (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Mark the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

<input checked="" type="checkbox"/> 0.100% STANDARD - MUST READ BETWEEN 0.095% AND 0.105% INCLUSIVE
<input type="checkbox"/> 0.080% STANDARD - MUST READ BETWEEN 0.076% AND 0.084% INCLUSIVE
<input type="checkbox"/> 0.040% STANDARD - MUST READ BETWEEN 0.038% AND 0.042% INCLUSIVE

TEST 1 ← .103%	TEST 2 ← .104%	TEST 3 ← .104%
----------------	----------------	----------------

PERFORM R.F.I. TEST (PRINTOUT ATTACHED)

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:
(DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS	0	(0-.04)	0	(.05-.09)	0	(.10-.14)	1	(.15-.19)	1	OVER .19	0
----------	---	---------	---	-----------	---	-----------	---	-----------	---	----------	---

LIST ANY NEW PARTS AND DESCRIBE ANY ALTERATION OR MODIFICATION THAT WAS MADE TO RESTORE THE INSTRUMENT TO OPERATE SATISFACTORILY AND WITHIN ESTABLISHED LIMITS
(USE OTHER SIDE IF NECESSARY).
OPERATING WITHIN D.O.H.S.S STANDARDS.

INSPECTING OFFICER:	
SIGNATURE 	PRINT FULL NAME LANCE ASH
TYPE II PERMIT NUMBER/EXPIRATION DATE 240149 04/22/16	TELEPHONE NUMBER (573) 471-8568

RETURN COMPLETED REPORT TO THE:
Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
2875 James Blvd.
Poplar Bluff, MO 63901

REPCO MARKETING INC.3101-188 STONYBROOK DRIVE
RALEIGH, NC. 27604
919-876-8480**CERTIFICATE OF ANALYSIS****MANUFACTURER AND SUPPLIER: RepCo Marketing, Inc.**
LOT NUMBER: 13002
EXPIRATION DATE: June 19, 2015 at 11:59 p.m.

RepCo Marketing, Inc. certifies the following:


RepCo Marketing, Inc. manufactured, tested and supplied Lot Number 13002 of Alcohol Certified Solution for simulators. Random samples of said lot number were analyzed by an independent laboratory utilizing a gas chromatograph and found to contain .1217 gms/dl +/- .003 gms/dl wt/vol ethanol (95% Confidence).

The alcohol and distilled water used in the solution were found to be free of any interfering substance.

This solution will produce a vapor alcohol value of .100 +/- 3% gms/210L: Breath when heated to 34 Degrees Celsius +/- 0.2 Degrees Celsius in a simulator (95% Confidence).

The date of manufacture for this lot number is June 20, 2013
The expiration date for this lot number is June 19, 2015 at
11:59 p.m.

This document is a true representation of the original Certificate of Analysis.



Cecil B. Garner, President
RepCo Marketing, Inc.

Face This Side Down - This Edge In First

BAC DataMaster Evidence Ticket

STATE OF MISSOURI
MINER POLICE DEPARTMENT

BAC DATAMASTER SERIAL NUMBER 201253
04/27/14
10:15

---- DIAGNOSTIC CHECK ----

COMPUTER: OKAY

PROGRAM (04-07-2009): OKAY

HEATERS
SAMPLE CHAMBER: 49c

FLOW DETECTOR: OKAY

PUMP
HIGH SPEED: OKAY

DETECTOR: OKAY

FILTERS: OKAY

QUARTZ STANDARD: OKAY

CALIBRATION: OKAY

PRINTER TEST

!"#\$%&'()*+,-./0123456789:;<=>?@ABCDEFGH
IJKLMNOPQRSTUVWXYZ[\]^_`abcdefghijklmnop
qrstuvwxyz{|}~

Operator Signature



Face This Side Down - This Edge In First

Face This Side Down - This Edge In First

BAC DataMaster Evidence Ticket

BAC DataMaster Evidence Ticket

STATE OF MISSOURI
MINER POLICE DEPARTMENT

STATE OF MISSOURI
MINER POLICE DEPARTMENT

BAC DATAMASTER SERIAL NUMBER 201253
04/27/14

BAC DATAMASTER SERIAL NUMBER 201253
04/27/14

TESTING OFFICER:
ASH/LANCE
OFFICER I.D.: 526
PERMIT NUMBER: 240149
EXPIRATION DATE: 04/22/16
MISCELLANEOUS DATA:
N
N

ARREST TIME: 10:10
SUBJECT NAME:
RFITEST
DOB: 12/22/34 SEX: M
STATE/D.L.: NA/NA
ARRESTING OFFICER:
NA
OFFICER I.D.: 526
TESTING OFFICER:
ASH/LANCE
OFFICER I.D.: NA
PERMIT NUMBER: 240149
EXPIRATION DATE: 04/22/16
MISCELLANEOUS DATA:
N
N

--- SUPERVISOR MODE ---


BLANK TEST	.000	10:18
INTERNAL STANDARD	VERIFIED	10:18
EXTERNAL STANDARD	.103	10:19
BLANK TEST	.000	10:20
EXTERNAL STANDARD	.104	10:20
BLANK TEST	.000	10:21
EXTERNAL STANDARD	.104	10:21
BLANK TEST	.000	10:22

N = 3
SIM. = .1
AVG. = .1036


--- BREATH ANALYSIS ---

BLANK TEST	.000	10:32
INTERNAL STANDARD	VERIFIED	10:32
RADIO INTERFERENCE		

Operator Signature



Operator Signature





STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM



PERMIT
TYPE II
JAMES L ASH

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

DATAMASTER, INTOXILYZER 5000

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 4/22/2014

NUMBER 240149

EXPIRES 4/22/2016

MO 580-0771 (6-10)

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

LAB-4 (R6-10)

 **STATE OF MISSOURI**
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.



Operator ASH, JAMES
Permit No 240149
Date Issued 4/22/2014 Date Expires 4/22/2016