



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
 STATE PUBLIC HEALTH LABORATORY
 BREATH ALCOHOL PROGRAM
DATAMASTER MAINTENANCE REPORT

RECEIVED REPORT #6
 By Carol Day at 9:56 am, Dec 18, 2014

Complete this report at the time of the regular monthly preventive maintenance check (not to exceed 30 days).
 Complete this report whenever the instrument is serviced or repaired and whenever it is placed into service.
 Retain the original and send a copy within 15 days to the Breath Alcohol Program, DHSS.

| | | |
|---|--------------------------------|----------------------------------|
| DATA MASTER SN 201253 | NAME OF AGENCY MINER POLICE | DATE OF INSPECTION 12/11/2014 |
| LOCATION OF INSTRUMENT (STREET AND CITY) 103 ST. HWY H, MINER, MO. 63801 | | TIME OF INSPECTION 8:16 am |

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

| | |
|--|---|
| <input checked="" type="checkbox"/> DIAGNOSTIC CHECK (PRINTOUT ATTACHED) | DATE AND TIME (from printout) <u>12-11-2014 08:17</u> |
| <input checked="" type="checkbox"/> COMPUTER | <input checked="" type="checkbox"/> DETECTOR |
| <input checked="" type="checkbox"/> PROGRAM | <input checked="" type="checkbox"/> FILTERS |
| <input checked="" type="checkbox"/> HEATERS SAMPLE CHAMBER <u>49</u> °C | <input checked="" type="checkbox"/> QUARTZ STANDARD |
| <input checked="" type="checkbox"/> FLOW DETECTOR | <input checked="" type="checkbox"/> CALIBRATION |
| <input checked="" type="checkbox"/> PUMP HIGH SPEED | <input checked="" type="checkbox"/> PRINTER |

INDICATOR LIGHTS

SIMULATOR SOLUTION SUPPLIER REPCO MARKETING LOT # 13002 EXP. DATE 06/19/2015

SIMULATOR TEMP (34°C ± 0.2°C) 34.0 °C SIMULATOR SN SD2273 EXP. DATE 01/03/2015

CALIBRATION CHECK - (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Mark the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

0.100% STANDARD - MUST READ BETWEEN 0.095% AND 0.105% INCLUSIVE
 0.080% STANDARD - MUST READ BETWEEN 0.076% AND 0.084% INCLUSIVE
 0.040% STANDARD - MUST READ BETWEEN 0.038% AND 0.042% INCLUSIVE

| | | |
|--------------------|--------------------|--------------------|
| TEST 1 <u>.102</u> | TEST 2 <u>.103</u> | TEST 3 <u>.103</u> |
|--------------------|--------------------|--------------------|

PERFORM R.F.I. TEST (PRINTOUT ATTACHED)

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:
 (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

| | | | | | | | | | | | |
|----------|---|----------|---|-----------|---|-----------|---|-----------|---|----------|---|
| REFUSALS | 0 | (.0-.04) | 0 | (.05-.09) | 0 | (.10-.14) | 0 | (.15-.19) | 0 | OVER .19 | 0 |
|----------|---|----------|---|-----------|---|-----------|---|-----------|---|----------|---|

LIST ANY NEW PARTS AND DESCRIBE ANY ALTERATION OR MODIFICATION THAT WAS MADE TO RESTORE THE INSTRUMENT TO OPERATE SATISFACTORILY AND WITHIN ESTABLISHED LIMITS (USE OTHER SIDE IF NECESSARY).
OPERATING WITHIN DHSS STANDARDS

| | |
|--|------------------------------------|
| INSPECTING OFFICER | |
| SIGNATURE | PRINT FULL NAME LANCE ASH |
| TYPE II PERMIT NUMBER EXPIRATION DATE 240149 04/22/2016 | TELEPHONE NUMBER (573) 471-8568 |

RETURN COMPLETED REPORT TO THE: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
 2875 James Blvd.
 Poplar Bluff, MO 63901

CERTIFICATE OF ANALYSIS

MANUFACTURER AND SUPPLIER: RepCo Marketing, Inc.
LOT NUMBER: 13002
EXPIRATION DATE: June 19, 2015 at 11:59 p.m.

RepCo Marketing, Inc. certifies the following:

RepCo Marketing, Inc. manufactured, tested and supplied Lot Number 13002 of Alcohol Certified Solution for simulators. Random samples of said lot number were analyzed by an independent laboratory utilizing a gas chromatograph and found to contain .1217 gms/dl +/- .003 gms/dl wt/vol ethanol (95% Confidence).

The alcohol and distilled water used in the solution were found to be free of any interfering substance.

This solution will produce a vapor alcohol value of .100 +/- 3% gms/210L Breath when heated to 34 Degrees Celsius +/- 0.2 Degrees Celsius in a simulator (95% Confidence).

The date of manufacture for this lot number is June 20, 2013
The expiration date for this lot number is June 19, 2015 at
11:59 p.m.

This document is a true representation of the original Certificate of Analysis.



Cecil B. Garner, President
RepCo Marketing, Inc.



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM



PERMIT
TYPE II
JAMES L ASH

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field services and repairs, and operate the following breath analyzer(s):

DATAMASTER, INTOXILYZER 5000

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.024, RSMo and 306.111 through 306.119 RSMo.

DATE: 4/22/2014

NUMBER: 240149

EXPIRES: 4/22/2016

HC 580-0771 (06/01)

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

CAE-40933D



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the ethanol content in breath form of expired air in Missouri.



Operator ASH, JAMES
 Permit No 240149
 Date Issued 4/22/2014 Date Expires 4/22/2016

BAC DataMaster
Evidence Ticket

STATE OF MISSOURI
MINER POLICE DEPARTMENT

BAC DATAMASTER SERIAL NUMBER 201233
12/11/14

ARREST TIME: 08:10
SUBJECT NAME:
DOE/JOHN
DOB: 12/22/34 SEX: M
STATE/D.L.: MO/123456789
ARRESTING OFFICER:
ASH
OFFICER I.D.: 526
TESTING OFFICER:
ASH/LANCE
OFFICER I.D.: 326
PERMIT NUMBER: 240149
EXPIRATION DATE: 04/22/16
MISCELLANEOUS DATA:
N
N

--- BREATH ANALYSIS ---

RADIO INTERFERENCE

Operator Signature

Ash #526

**BAC DataMaster
Evidence Ticket**

STATE OF MISSOURI
MINER POLICE DEPARTMENT

BAC DATAMASTER SERIAL NUMBER 201253
12/11/14
08:17

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY
PROGRAM (04-07-2009): OKAY
EATERS
SAMPLE CHAMBER: 49c
LOW DETECTOR: OKAY
UMP
HIGH SPEED: OKAY
DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY

PRINTER TEST

!"#\$%&'()*+,-./0123456789:;<=>?@ABCDEFGHIJKL
MNOPQRSTUVWXYZ[\]^_`abcdefghijklmnopqr
stuvwxyz{|}~`

Signature

Ash #526

2208-02

**BAC DataMaster
Evidence Ticket**

STATE OF MISSOURI
MINER POLICE DEPARTMENT

BAC DATAMASTER SERIAL NUMBER 201253
12/11/14

TESTING OFFICER:
ASH/LANCE
OFFICER I.D.: 526
PERMIT NUMBER: 240149
EXPIRATION DATE: 04/22/18
MISCELLANEOUS DATA:
N

--- SUPERVISOR MODE ---

| | | |
|-------------------|----------|-------|
| BLANK TEST | .000 | 08:22 |
| INTERNAL STANDARD | VERIFIED | 08:22 |
| EXTERNAL STANDARD | .102 | 08:22 |
| BLANK TEST | .000 | 08:23 |
| EXTERNAL STANDARD | .103 | 08:23 |
| BLANK TEST | .000 | 08:24 |
| EXTERNAL STANDARD | .103 | 08:25 |
| BLANK TEST | .000 | 08:26 |

N = 3
SIM. = .1
AVG. = .1026

Operator Signature

Ash #526

2208-02