



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
 STATE PUBLIC HEALTH LABORATORY
 BREATH ALCOHOL PROGRAM
DATAMASTER MAINTENANCE REPORT

RECEIVED REPORT #6
 By Carol Day at 8:26 am, Sep 03, 2014

Complete this report at the time of the regular monthly preventive maintenance check (not to exceed 35 days).
 Complete this report whenever the instrument is serviced or repaired and whenever it is placed into service.
 Retain the original and send a copy within 15 days to the Breath Alcohol Program, DHSS.

DATAMASTER SN 201255	NAME OF AGENCY NEOSTO POLICE	DATE OF INSPECTION 8-30-2014
LOCATION OF INSTRUMENT (STREET AND CITY) 60 @ DARRIDGE DR (CHECKPOINT)		TIME OF INSPECTION 21:10

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

<input checked="" type="checkbox"/> DIAGNOSTIC CHECK (PRINTOUT ATTACHED)	DATE AND TIME (from printout) 8/30/14 21:11
<input checked="" type="checkbox"/> COMPUTER	<input checked="" type="checkbox"/> DETECTOR
<input checked="" type="checkbox"/> PROGRAM	<input checked="" type="checkbox"/> FILTERS
<input checked="" type="checkbox"/> HEATERS SAMPLE CHAMBER 49 °C	<input checked="" type="checkbox"/> QUARTZ STANDARD
<input type="checkbox"/> FLOW DETECTOR	<input type="checkbox"/> CALIBRATION
<input checked="" type="checkbox"/> PUMP HIGH SPEED	<input checked="" type="checkbox"/> PRINTER

INDICATOR LIGHTS

SIMULATOR SOLUTION SUPPLIER **REPCO MARKETING** LOT # **13002** EXP. DATE **06-19-15**

SIMULATOR TEMP (34°C ± 0.2°C) **34.0** °C SIMULATOR SN **SD2237** EXP. DATE **8-14-15**

CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Mark the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

0.100% STANDARD - MUST READ BETWEEN 0.095% AND 0.105% INCLUSIVE
 0.080% STANDARD - MUST READ BETWEEN 0.076% AND 0.084% INCLUSIVE
 0.040% STANDARD - MUST READ BETWEEN 0.038% AND 0.042% INCLUSIVE

TEST 1 • .097 %	TEST 2 • .099 %	TEST 3 • .099 %
------------------------	------------------------	------------------------

PERFORM R.F.I. TEST (PRINTOUT ATTACHED)

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS 0 (0-.04)	0 (.05-.09)	0 (.10-.14)	0 (.15-.19)	0 OVER .19
---------------------------	--------------------	--------------------	--------------------	-------------------

LIST ANY NEW PARTS AND DESCRIBE ANY ALTERATION OR MODIFICATION THAT WAS MADE TO RESTORE THE INSTRUMENT TO OPERATE SATISFACTORILY AND WITHIN ESTABLISHED LIMITS (USE OTHER SIDE IF NECESSARY).

OPERATING WITHIN DOT SPECIFICATIONS

INSPECTING OFFICER

SIGNATURE 	PRINT FULL NAME DAVID KENNEDY
TYPE II PERMIT NUMBER/EXPIRATION DATE 230242 10/24/15	TELEPHONE NUMBER 417 451-8012

RETURN COMPLETED REPORT TO THE:
 Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
 2875 James Blvd.
 Poplar Bluff, MO 63901

CERTIFICATE OF ANALYSIS

MANUFACTURER AND SUPPLIER: RepCo Marketing, Inc.

LOT NUMBER: 13002

EXPIRATION DATE: June 19, 2015 at 11:59 p.m.

RepCo Marketing, Inc. certifies the following:


RepCo Marketing, Inc. manufactured, tested and supplied Lot Number 13002 of Alcohol Certified Solution for simulators. Random samples of said lot number were analyzed by an independent laboratory utilizing a gas chromatograph and found to contain .1217 gms/dl +/-0.003 gms/dl wt/vol ethanol (95% Confidence).

The alcohol and distilled water used in the solution were found to be free of any interfering substance.

This solution will produce a vapor alcohol value of .100 +/-3% gms/210L Breath when heated to 34 Degrees Celsius +/-0.2 Degrees Celsius in a simulator (95% Confidence).

The date of manufacture for this lot number is June 20, 2013
The expiration date for this lot number is June 19, 2015 at
11:59 p.m.

This document is a true representation of the original Certificate of Analysis.



Cecil B. Garner, President
RepCo Marketing, Inc.

BAC DataMaster Evidence Ticket

STATE OF MISSOURI
NEOSHO POLICE DEPARTMENT

OFFICER: [Name] / 3541HQ NUMBER: 201255
AR/30 11

DATE: 11/20/10 08:10
SUBJECT: [Name]

DRIVER: [Name]
VEHICLE: [Make/Model/Year]
LICENSE: [Number]

TEST TYPE: [Type]
TESTER: [Name]

TEST RESULTS: [List of results]

TESTER SIGNATURE: [Signature]
DATE: [Date]

Officer Signature: [Signature] / 1-3

[Signature]

[Signature]

HW = 0
SIR = 11
AW = 10980

BLANK TEST	1.00	1.00
EXTERNAL STIMULUS	1.00	1.00
BLANK TEST	1.00	1.00
EXTERNAL STIMULUS	1.00	1.00
BLANK TEST	1.00	1.00
EXTERNAL STIMULUS	1.00	1.00
INTERNAL STIMULUS	1.00	1.00
BLANK TEST	1.00	1.00

DRY	0.00	0.00
EXTERNAL STIMULUS	0.00	0.00
DETECTOR	0.00	0.00
HIGH SPEED PUMP	0.00	0.00
LOW DETECTOR	0.00	0.00
INTERNAL STIMULUS	0.00	0.00
EXTERNAL STIMULUS	0.00	0.00
COMPLEMENT	0.00	0.00

NEOSHO POLICE DEPARTMENT
STATE OF MISSOURI
OFFICER: [Name]

STATE OF MISSOURI
NEOSHO POLICE DEPARTMENT
OFFICER: [Name]



Simulator Calibration Report

This calibration report is to certify the alcohol reference simulator listed below has been examined and tested using standards traceable to the National Institute of Standards and Technology (NIST) in accordance with the standards set by the Missouri Department of Health and Senior Services Rules and Regulations:
19 CSR 25-30.051 (4).

Checked: 08/14/2014 Expires: 08/14/2015
Digital Therm. SN:094948
MSC Tech:DRL Temp:33.96
Agency: Neosho Police Dept
SD 2237



Technician Printed Name: DAN LUCAS

Technician Signature: *[Handwritten Signature]*

Date: 8/14/2014

Contact: Missouri Safety Center

Breath-Alcohol Instrument Training Program

660-543-4834



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM



PERMIT
TYPE II
DAVID KENNEDY

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

DATAMASTER, INTOX DMT

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 10/24/2013

NUMBER 230242

EXPIRES 10/24/2015

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

acting director

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MO 580-0771 (6-10)

LAB-4 (R6-10)



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.



Operator KENNEDY, DAVID
 Permit No 230242
 Date Issued 10/24/2013 Date Expires 10/24/2015