MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
STATE PUBLIC HEALTH LABORATORY
BREATH ALCOHOL PROGRAM

DATAMASTER MAINTENANCE REPORT

REVIEWED REPORT #6
By Carol Day at 10:01 am, Aug 28, 2014

Complete this report at the time of the regular monthly preventive maintenance check (not to exceed 35 days).
Complete this report whenever the instrument is serviced or repaired and whenever it is placed into service.
Retain the original and send a copy within 15 days to the Breath Alcohol Program, DHSS.

DATAMASTER SN	NAME OF AGENCY	DATE OF INSPECTION
201256	Winfield P.D.	07/27/2014

LOCATION OF INSTRUMENT (STREET AND CITY)	TIME OF INSPECTION
51 Harry's Way Winfield, Missouri	3:58 am

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

☑ DIAGNOSTIC CHECK (PRINTOUT ATTACHED) DATE AND TIME (from printout) 07/27/2014 @03:58

☑ COMPUTER ☑ DETECTOR

☑ PROGRAM ☑ FILTERS

☑ HEATERS SAMPLE CHAMBER 48 °C ☑ QUARTZ STANDARD

☑ FLOW DETECTOR ☑ CALIBRATION

☑ PUMP HIGH SPEED ☑ PRINTER

☑ INDICATOR LIGHTS

☑ SIMULATOR SOLUTION SUPPLIER Guth Laboratories LOT # 13290 EXP. DATE 10/29/2015

☑ SIMULATOR TEMP (34°C ± 0.2°C) 34 °C SIMULATOR SN SD 3000 EXP. DATE 12/17/2014

☑ CALIBRATION CHECK – **(ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)**

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Mark the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

☑ 0.100% STANDARD - MUST READ BETWEEN 0.095% AND 0.105% INCLUSIVE
☐ 0.080% STANDARD - MUST READ BETWEEN 0.076% AND 0.084% INCLUSIVE
☐ 0.040% STANDARD - MUST READ BETWEEN 0.038% AND 0.042% INCLUSIVE

TEST 1	TEST 2	TEST 3
.098	.098	.099

☑ PERFORM R.F.I. TEST (PRINTOUT ATTACHED)

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS	(0-.04)	(.05-.09)	(.10-.14)	(.15-.19)	OVER .19
0	1	3	2	2	1

LIST ANY NEW PARTS AND DESCRIBE ANY ALTERATION OR MODIFICATION THAT WAS MADE TO RESTORE THE INSTRUMENT TO OPERATE SATISFACTORILY AND WITHIN ESTABLISHED LIMITS (USE OTHER SIDE IF NECESSARY).

Instrument Operating Within Established Limits.

INSPECTING OFFICER

SIGNATURE	PRINT FULL NAME
Sgt. M. McCollister 12502	Sgt. M. McCollister, DSN: 2502
TYPE II PERMIT NUMBER/EXPIRATION DATE	**TELEPHONE NUMBER**
240304 07/22/2016	(636) 566-6936

RETURN COMPLETED REPORT TO THE: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
2875 James Blvd.
Poplar Bluff, MO 63901

GUTH LABORATORIES, INC.
590 NORTH 67th STREET • HARRISBURG, PA 17111-4611 • TELEPHONE: 717-564-6470

CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number 13290 of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on October 31, 2013, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain 0.1202% (w/vol) ethyl alcohol. The expiration date for this lot number is October 29, 2015 at 11:59 PM.

When used in a calibrated Simulator, operating at 34°C +/- .2°C, this solution will give a breath alcohol analysis instrument reading of 0.100 g/210L +/- 3%.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President
GUTH LABORATORIES, INC.

NIST Traceability:
Testing was conducted using Certiliant Reference Standard lot number FN122211-02 whose values are traceable to NIST.
All balances are calibrated annually by an outside agency using NIST traceable weights
Calibration verification is done prior to each use utilizing NIST traceable weights

STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

2

PERMIT
TYPE II
MARC MCCOLLISTER

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

DATAMASTER

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 7/22/2014

NUMBER 240304

EXPIRES 7/22/2016

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MO 580-0771 (6-10)

LAB-4 (R6-10)

STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator MCCOLLISTER, MARC
Permit No 240304
Date Issued 7/22/2014 **Date Expires** 7/22/2016

Face This Side Down – This Edge In First

BAC DataMaster
Evidence Ticket

STATE OF MISSOURI
BAC DATAMASTER SERIAL NUMBER 201256
07/27/14

TESTING OFFICER:
MCCOLLISTER/MARC/A
OFFICER I.D.: 2502
PERMIT NUMBER: 240304
EXPIRATION DATE: 07/22/16
MISCELLANEOUS DATA:
N/A
N/A

--- SUPERVISOR MODE ---

BLANK TEST	.000	04:01
INTERNAL STANDARD	VERIFIED	04:01
EXTERNAL STANDARD	.098	04:01
BLANK TEST	.000	04:02
EXTERNAL STANDARD	.098	04:02
BLANK TEST	.000	04:03
EXTERNAL STANDARD	.099	04:04
BLANK TEST	.000	04:04

N = 3
SIM. = .1
AVG. = .0983

tor Signature M. Millis

2208-02

Face This Side Down – This Edge In First

BAC DataMaster
Evidence Ticket

STATE OF MISSOURI
BAC DATAMASTER SERIAL NUMBER 201256
07/27/14
03:58

--- DIAGNOSTIC CHECK ---

COMPUTER:	OKAY
PROGRAM (04-07-2009):	OKAY
HEATERS SAMPLE CHAMBER:	48c
FLOW DETECTOR:	OKAY
PUMP HIGH SPEED:	OKAY
DETECTOR:	OKAY
FILTERS:	OKAY
QUARTZ STANDARD:	OKAY
CALIBRATION:	OKAY

PRINTER TEST
!"#$%&'()*+,-./0123456789:;<=>?@ABCDEFG
HIJKLMNOPQRSTUVWXYZ[\]^_`abcdefghijklmno
pqrstuvwxyz{|}~

Operator Signature M. Millis

2208-02

Face This Side Down – This Edge In First

BAC DataMaster
Evidence Ticket

STATE OF MISSOURI
BAC DATAMASTER SERIAL NUMBER 201256
07/27/14

ARREST TIME: 23:00
SUBJECT NAME:
DOE/JOHN/A
DOB: 09/09/69 SEX: M
STATE/D.L.: MO/00000000069
ARRESTING OFFICER:
MCCOLLISTER/MARC/A
OFFICER I.D.: 2502
TESTING OFFICER:
MCCOLLISTER/MARC/A
OFFICER I.D.: 2502
PERMIT NUMBER: 240304
EXPIRATION DATE: 07/22/16
MISCELLANEOUS DATA:
RFI TEST
N/A

--- BREATH ANALYSIS ---

BLANK TEST	.000	04:08
INTERNAL STANDARD	VERIFIED	04:09
RADIO INTERFERENCE		

Operator Signature M. McCollister

2208-02