

DATAMASTER MAINTENANCE REPORT

RECEIVED REPORT #8
By Carol Day at 8:42 am, Jan 15, 2014

State this report at the time of the regular monthly preventive maintenance check (not to exceed 35 days).
State this report whenever the instrument is serviced or repaired and whenever it is placed into service.
Retain the original and send a copy within 15 days to the Breath Alcohol Program, DHSS.

TAMASTER SN 201259	NAME OF AGENCY Boonville Police Dept	DATE OF INSPECTION 01/14/2014
LOCATION OF INSTRUMENT (STREET AND CITY) 401 E. Morgan, Boonville		TIME OF INSPECTION 2304 hrs.

HECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

<input checked="" type="checkbox"/> DIAGNOSTIC CHECK (PRINTOUT ATTACHED)	DATE AND TIME (from printout) 2313 hrs
<input checked="" type="checkbox"/> COMPUTER	<input checked="" type="checkbox"/> DETECTOR
<input checked="" type="checkbox"/> PROGRAM	<input checked="" type="checkbox"/> FILTERS
<input checked="" type="checkbox"/> HEATERS SAMPLE CHAMBER 48 °C	<input type="checkbox"/> QUARTZ STANDARD
<input checked="" type="checkbox"/> FLOW DETECTOR	<input checked="" type="checkbox"/> CALIBRATION
<input checked="" type="checkbox"/> PUMP HIGH SPEED	<input type="checkbox"/> PRINTER
<input checked="" type="checkbox"/> INDICATOR LIGHTS	
<input checked="" type="checkbox"/> SIMULATOR SOLUTION SUPPLIER Rep Co Marketing LOT # 13001 EXP. DATE 03/07/2015	
<input checked="" type="checkbox"/> SIMULATOR TEMP (34°C ± 0.2°C) 34.0 °C SIMULATOR SN SD3501 EXP. DATE 8/8/2014	

CALIBRATION CHECK - (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Mark the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

0.100% STANDARD - MUST READ BETWEEN 0.095% AND 0.105% INCLUSIVE

0.080% STANDARD - MUST READ BETWEEN 0.078% AND 0.084% INCLUSIVE

0.040% STANDARD - MUST READ BETWEEN 0.038% AND 0.042% INCLUSIVE


TEST 1 .097	TEST 2 .097	TEST 3 .098
--------------------	--------------------	--------------------

PERFORM R.F.I. TEST (PRINTOUT ATTACHED) **2303 hrs.**

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS 1	(.0-.04) 21	(.05-.09) -	(.10-.14) -	(.15-.19) -	OVER .19 -
-------------------	--------------------	--------------------	--------------------	--------------------	-------------------

LIST ANY NEW PARTS AND DESCRIBE ANY ALTERATION OR MODIFICATION THAT WAS MADE TO RESTORE THE INSTRUMENT TO OPERATE SATISFACTORILY AND WITHIN ESTABLISHED LIMITS (USE OTHER SIDE IF NECESSARY).

INSPECTING OFFICER	
SIGNATURE 	PRINT FULL NAME JAMES DECKER
TYPE II PERMIT NUMBER EXPIRATION DATE 220215 09/04/14	TELEPHONE NUMBER 660-882-2727

RETURN COMPLETED REPORT TO THE: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
2876 James Blvd.
Ponchar Ruff MO 63001

3101-188 STONYBROOK DRIVE
RALEIGH, N.C. 27604
919-876-5480

CERTIFICATE OF ANALYSIS

MANUFACTURER AND SUPPLIER: RepCo Marketing, Inc.

LOT NUMBER: 13001

EXPIRATION DATE: March 7, 2015 at 11:59 p.m.

RepCo Marketing, Inc. certifies the following:

RepCo Marketing, Inc. manufactured, tested and supplied Lot Number 13001 of Alcohol Certified Solution for simulators. Random samples of said lot number were analyzed by an independent laboratory utilizing a gas chromatograph and found to contain .1215 gms/dl +/- .003 gms/dl wt/vol ethanol (95% Confidence).


The alcohol and distilled water used in the solution were found to be free of any interfering substance.

This solution will produce a vapor alcohol value of .100 +/- 3% gms/210L Breath when heated to 34 Degrees Celsius +/- 0.2 Degrees Celsius in a simulator (95% Confidence).

The date of manufacture for this lot number is March 8, 2013

The expiration date for this lot number is March 7, 2015 at 11:59 p.m.

This document is a true representation of the original Certificate of Analysis.



Cecil B. Garner, President
RepCo Marketing, Inc.



STATE OF MISSOURI
BOONVILLE POLICE DEPARTMENT

BACID MASTER SERIAL NUMBER 201259

01/14/14
23:04

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY

PROGRAM (04-07-2009): OKAY

HEATERS: 4SC

SAMPLE CHAMBER: 4SC

FLOW DETECTOR: OKAY

PUMP: OKAY

HIGH SPEED: OKAY

DETECTOR: OKAY

FILTERS: OKAY

QUARTZ STANDARD: OKAY

CALIBRATION: OKAY

PRINTER TEST:

/0123456789:;<=>?@A|BCDEFG
HIJKLMNOPQRSTUVWXYZ\]^_`abcd|efghijklmno
pqrstuvwxyzzz1234

Operator's Signature

A large, stylized handwritten signature in black ink.



STATE OF MISSOURI
BOONVILLE POLICE DEPARTMENT

BAC DATA MASTER SERIAL NUMBER 201253
01/14/14

PREST TIME: 22:00

SUBJECT NAME:

TEST

DOB: 01/01/80

SEX: M

STATE/D.L.: OK/000000

RESTING OFFICER:

TEST

OFFICER I.D.: 0

TESTING OFFICER:

DECKARD, JAMES M

OFFICER I.D.: 07

UNIT NUMBER: 220215

EXPIRATION DATE: 09/04/14

ISCALENEEDS DATA:

--- BREATH ANALYSIS ---

BLANK TEST

INTERNAL STANDARD

NO ID INTERFERENCE

.000 23:03
VERIFIED 23:03

Operator Signature



STATE OF MISSOURI
BOONVILLE POLICE DEPARTMENT

BAC DATA MASTER SERIAL NUMBER 201255
01/14/14

TESTING OFFICER:
DECKERD/JAMES/M
OFFICER I.T. : 07
PERMIT NUMBER: 226215
EXPIRATION DATE: 03/04/14
MISCELLANEOUS DATA:

--- SUPERVISOR MODE: ---

BLANK TEST	.000	23:13
INTERNAL STANDARD	VERIFIED	23:13
EXTERNAL STANDARD	.097	23:13
BLANK TEST	.000	23:14
EXTERNAL STANDARD	.097	23:14
BLANK TEST	.000	23:15
EXTERNAL STANDARD	.096	23:15
BLANK TEST	.000	23:16

N = 3
SIM. = .1
QW. = .0973

Operator Signature



PERMIT TYPE II



JAMES DECKARD

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field repairs, and operate the following breath analyzer(s):

DATAMASTER

for the determination of the alcoholic content of blood from a sample of expired (alveolar) air. Issued under the provisions of sections 577.020 through 577.041, RSMo 1986.

Date 09/04/2012

Number 220215

Expires 09/04/2014

MO 680-0771 (7-88)

Director of State Public Health Laboratory

Director, Department of Health

Lab. 4 (R7-88)