



BREATH ALCOHOL PROGRAM
DATAMASTER MAINTENANCE REPORT

RECEIVED
 REPORT #
 By Carol Day at 9:59 am, Jul 14, 2014

Complete this report at the time of the regular monthly preventive maintenance check (not to exceed 35 days).
 Complete this report whenever the instrument is serviced or repaired and whenever it is placed into service.
 Retain the original and send a copy within 15 days to the Breath Alcohol Program, DHSS.

DATAMASTER SN <u>201259</u>	NAME OF AGENCY <u>Boonville PD</u>	DATE OF INSPECTION <u>07-10-2014</u>
LOCATION OF INSTRUMENT (STREET AND CITY) <u>401 E Morgan St. Boonville</u>		TIME OF INSPECTION <u>0526</u>

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

<input checked="" type="checkbox"/> DIAGNOSTIC CHECK (PRINTOUT ATTACHED)	DATE AND TIME (from printout) <u>0522</u>
<input checked="" type="checkbox"/> COMPUTER	<input checked="" type="checkbox"/> DETECTOR
<input checked="" type="checkbox"/> PROGRAM	<input checked="" type="checkbox"/> FILTERS
<input checked="" type="checkbox"/> HEATERS SAMPLE CHAMBER <u>48</u> °C	<input checked="" type="checkbox"/> QUARTZ STANDARD
<input checked="" type="checkbox"/> FLOW DETECTOR	<input checked="" type="checkbox"/> CALIBRATION
<input checked="" type="checkbox"/> PUMP HIGH SPEED	<input checked="" type="checkbox"/> PRINTER

INDICATOR LIGHTS

SIMULATOR SOLUTION SUPPLIER Guth LOT # 13280 EXP. DATE 10-16-15

SIMULATOR TEMP (34°C ± 0.2°C) 34.0 °C SIMULATOR SN SD3501 EXP. DATE 08-08-14

CALIBRATION CHECK - (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Mark the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

0.100% STANDARD - MUST READ BETWEEN 0.095% AND 0.105% INCLUSIVE
 0.080% STANDARD - MUST READ BETWEEN 0.076% AND 0.084% INCLUSIVE
 0.040% STANDARD - MUST READ BETWEEN 0.038% AND 0.042% INCLUSIVE

TEST 1 = <u>.095</u>	TEST 2 = <u>.096</u>	TEST 3 = <u>.095</u>
----------------------	----------------------	----------------------

PERFORM R.F.I. TEST (PRINTOUT ATTACHED)

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:
 (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS <u>0</u>	(0-.04) <u>3</u>	(.05-.09) <u>1</u>	(.10-.14) <u>1</u>	(.15-.19) <u>1</u>	OVER .19 <u>0</u>
-------------------	------------------	--------------------	--------------------	--------------------	-------------------

LIST ANY NEW PARTS AND DESCRIBE ANY ALTERATION OR MODIFICATION THAT WAS MADE TO RESTORE THE INSTRUMENT TO OPERATE SATISFACTORILY AND WITHIN ESTABLISHED LIMITS (USE OTHER SIDE IF NECESSARY).

INSPECTING OFFICER

SIGNATURE 	PRINT FULL NAME <u>Eric L. Moss</u>
---------------	--

TYPE II PERMIT NUMBER/EXPIRATION DATE <u>240294 6/27/2016</u>	TELEPHONE NUMBER <u>660-882-2727</u>
--	---

RETURN COMPLETED REPORT TO THE: **Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office**
 2875 James Blvd.
 Parker Bluff MO 63901



GUTH LABORATORIES, INC.

690 NORTH 87th STREET • HARRISBURG, PA 17111-4611 • TELEPHONE: 717-664-6470

CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number 13280 of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on October 18, 2013, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain 0.1217% (w/vol) ethyl alcohol. The expiration date for this lot number is October 16, 2015 at 11:59 PM.

When used in a calibrated Simulator, operating at 34°C +/- .2°C, this solution will give a breath alcohol analysis instrument reading of 0.100 g/210L +/- 3%.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President
GUTH LABORATORIES, INC.

NIST Traceability:

Testing was conducted using Cerilliant Reference Standard lot number FN122211-02 whose values are traceable to NIST.

All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.

BAC DataMaster
Evidence Marked

STATE OF MISSOURI
POLICE DEPARTMENT

BAC DATAMASTER SERIAL NUMBER 201259

07/10/14
05:22

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY

PROGRAM (04-07-2009): OKAY

HEATERS
SAMPLE CHAMBER: 48C

FLOW DETECTOR: OKAY

PUMP
HIGH SPEED: OKAY

DETECTOR: OKAY

FILTERS: OKAY

GURRITZ STANDARD: OKAY

CALIBRATION: OKAY

PRINTER TEST

!"#\$%&'()*+,-./0123456789:;<=>?@ABCDEF
HIJKLMNOPQRSTUVWXYZ[\]^_`abcdefgijklmno
pqrstuvwxyza{1234567890

Signature: *[Handwritten Signature]*

BACD DataMaster
Evidence Marker

STATE OF MISSOURI
JEFFERSONVILLE POLICE DEPARTMENT

ABC DRINKMASTER SERIAL NUMBER 240294
07/18/14

ARREST TIME: 05:00

SEX: M
HERNIT NUMBER: 240294

DATE: 06/27/16

INTERNAL STANDARD
EXTERNAL STANDARD
MISCELLANEOUS DATA

STRENGTH ANALYSIS

LABORATORY INTERFERENCE

Signature

BACD DataMaster
Evidence Marker

STATE OF MISSOURI
JEFFERSONVILLE POLICE DEPARTMENT

ABC DRINKMASTER SERIAL NUMBER 201259
07/18/14

TESTING OFFICER:

MOSS/ERTEL

OFFICER: J.D. 10
HERNIT NUMBER: 240294

DATE: 06/27/16
MISCELLANEOUS DATA:

SUPERVISOR MODE

STRENGTH ANALYSIS	INTERNAL STANDARD	EXTERNAL STANDARD	VERIFIED
0.000	0.000	0.000	0.000
0.000	0.000	0.000	0.000
0.000	0.000	0.000	0.000
0.000	0.000	0.000	0.000
0.000	0.000	0.000	0.000

DATE: 06/27/16
NO. 10953

Signature



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM



PERMIT
TYPE II
ERIC L MOSS

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

DATAMASTER

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 6/27/2014

NUMBER 240294

EXPIRES 6/27/2016

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator MOSS, ERIC
 Permit No 240294
 Date Issued 6/27/2014 Date Expires 6/27/2016