DATAMASTER MAINTENANCE REPORT

RECEIVED
By Carol Day at 11:46 am, Oct 01, 2014

Complete this report at the time of the regular monthly preventive maintenance check (not to exceed 35 days).
Complete this report whenever the instrument is serviced or repaired and whenever it is placed into service.
Retain the original and send a copy within 15 days to the Breath Alcohol Program, DHSS.

DATAMASTER SN	NAME OF AGENCY	DATE OF INSPECTION
201259	Boonville Police Department	10/01/2014
LOCATION OF INSTRUMENT (STREET AND CITY)		TIME OF INSPECTION
401 E. Morgan Street, Boonville MO 65233		0144

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

- ☑ DIAGNOSTIC CHECK (PRINTOUT ATTACHED) — DATE AND TIME (from printout) 10/01/14 01:33
  - ☑ COMPUTER
  - ☑ DETECTOR
  - ☑ PROGRAM
  - ☑ FILTERS
  - ☑ HEATERS SAMPLE CHAMBER 48 °C
  - ☑ QUARTZ STANDARD
  - ☑ FLOW DETECTOR
  - ☑ CALIBRATION
  - ☑ PUMP HIGH SPEED
  - ☑ PRINTER
- ☑ INDICATOR LIGHTS
- ☑ SIMULATOR SOLUTION SUPPLIER Guth Labs LOT # 13280 EXP. DATE 10/18/2015
- ☑ SIMULATOR TEMP (34°C ± 0.2°C) 34.0 °C SIMULATOR SN SD3501 EXP. DATE ~~[illegible]~~ 9/30/15
- ☑ CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Mark the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

- ☑ 0.100% STANDARD - MUST READ BETWEEN 0.095% AND 0.105% INCLUSIVE
- ☐ 0.080% STANDARD - MUST READ BETWEEN 0.076% AND 0.084% INCLUSIVE
- ☐ 0.040% STANDARD - MUST READ BETWEEN 0.038% AND 0.042% INCLUSIVE

TEST 1	TEST 2	TEST 3
.095	.096	.096

- ☑ PERFORM R.F.I. TEST (PRINTOUT ATTACHED)

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS	(0-.04)	(.05-.09)	(.10-.14)	(.15-.19)	OVER .19
Ø	Ø	Ø	Ø	Ø	Ø

LIST ANY NEW PARTS AND DESCRIBE ANY ALTERATION OR MODIFICATION THAT WAS MADE TO RESTORE THE INSTRUMENT TO OPERATE SATISFACTORILY AND WITHIN ESTABLISHED LIMITS (USE OTHER SIDE IF NECESSARY).

INSPECTING OFFICER

SIGNATURE	PRINT FULL NAME
J. Deckard	James Deckard
TYPE II PERMIT NUMBER/EXPIRATION DATE	TELEPHONE NUMBER
220215 09/04/2014	(660) 882-2727

RETURN COMPLETED REPORT TO THE: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
2875 James Blvd.
Poplar Bluff, MO 63901

AN EQUAL OPPORTUNITY/AFFIRMATIVE ACTION EMPLOYER
services provided on a nondiscriminatory basis

GUTH LABORATORIES, INC.
550 NORTH 47th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-554-5475

CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number 13280 of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on October 18, 2013, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain 0.1217% (w/vol) ethyl alcohol. The expiration date for this lot number is October 16, 2015 at 11:59 PM.

When used in a calibrated Simulator, operating at 34°C +/- .2°C, this solution will give a breath alcohol analysis instrument reading of 0.100 g/210L +/- 3%.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President
GUTH LABORATORIES, INC.

NIST Traceability:
Testing was conducted using Cerilliant Reference Standard lot number FN122211-02 whose values are traceable to NIST.
All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.

BAC DataMaster
Evidence Ticket

STATE OF MISSOURI
BOONVILLE POLICE DEPARTMENT

BAC DATAMASTER SERIAL NUMBER 201259
10/01/14

TESTING OFFICER:
DECKARD/JAMES/M
OFFICER I.D.: 07
PERMIT NUMBER: 240291
EXPIRATION DATE: 06/27/16
MISCELLANEOUS DATA:

---- SUPERVISOR MODE ----

BLANK TEST	.000	01:44
INTERNAL STANDARD	VERIFIED	01:44
EXTERNAL STANDARD	.095	01:45
BLANK TEST	.000	01:45
EXTERNAL STANDARD	.096	01:46
BLANK TEST	.000	01:46
INTERNAL STANDARD	.096	01:47
BLANK TEST	.000	01:47

N = 3
MIN. = .1
AVG. = .0956

Operator Signature

BAC DataMaster
Evidence Ticket

STATE OF MISSOURI
BOONVILLE POLICE DEPARTMENT

BAC DATAMASTER SERIAL NUMBER 201259
10/01/14
01:33

--- DIAGNOSTIC CHECK ---

COMPUTER:	OKAY
PROGRAM (04-07-2009):	OKAY
HEATERS SAMPLE CHAMBER:	48c
FLOW DETECTOR:	OKAY
PUMP HIGH SPEED:	OKAY
DETECTOR:	OKAY
FILTERS:	OKAY
QUARTZ STANDARD:	OKAY
CALIBRATION:	OKAY

PRINTER TEST
!"#$%&'()*+,-./0123456789:;<=>?@ABCDEFG
HIJKLMNOPQRSTUVWXYZ[\]^_`abcdefghijklmno
pqrstuvwxyz{|}~

Signature

BAC DataMaster
Evidence Ticket

STATE OF MISSOURI
BUCKVILLE POLICE DEPARTMENT

BAC DATAMASTER SERIAL NUMBER 201259
10/01/14

ARREST TIME: 00:10
SUBJECT NAME:
TEST
DOB: 01/01/90 SEX: F
STATE/D.L.: NV/000000
ARRESTING OFFICER:
TEST
OFFICER I.D.: 00
TESTING OFFICER:
DECKARD/JAMES/M
OFFICER I.D.: 07
PERMIT NUMBER: 240291
EXPIRATION DATE: 06/27/16
MISCELLANEOUS DATA:

--- BREATH ANALYSIS ---

BLANK TEST	.000	01:37
INTERNAL STANDARD	VERIFIED	01:37
RADIO INTERFERENCE		

Operator Signature

STATE OF MIS

DEPARTMENT OF HEALTH

BREATH ALCOHO

2

PERMIT

TYPE II

JAMES M DECKARD

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

DATAMASTER

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 6/27/2014

NUMBER 240291

EXPIRES 6/27/2016

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MO 580-0771 (6-10)

LAB-4 (R6-10)

STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator DECKARD, JAMES
Permit No 240291
Date Issued 6/27/2014 Date Expires 6/27/2016