



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
 STATE PUBLIC HEALTH LABORATORY
 BREATH ALCOHOL PROGRAM
DATAMASTER MAINTENANCE REPORT

RECEIVED
 By Carol Day at 2:27 pm, Jan 23, 2014

REPORT #6

Complete this report at the time of the regular monthly preventive maintenance check (not to exceed 35 days).
 Complete this report whenever the instrument is serviced or repaired and whenever it is placed into service.
 Retain the original and send a copy within 15 days to the Breath Alcohol Program, DHSS.

DATAMASTER SN <u>201268</u>	NAME OF AGENCY <u>VANDALIA Police Dept.</u>	DATE OF INSPECTION <u>01-16-2014</u>
LOCATION OF INSTRUMENT (STREET AND CITY) <u>201 E. Park St. Vandalia Mo. 63382</u>		TIME OF INSPECTION <u>14:18</u>

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

<input checked="" type="checkbox"/> DIAGNOSTIC CHECK (PRINTOUT ATTACHED)	DATE AND TIME (from printout) <u>01-16-2014 14:18</u>
<input checked="" type="checkbox"/> COMPUTER	<input checked="" type="checkbox"/> DETECTOR
<input checked="" type="checkbox"/> PROGRAM	<input checked="" type="checkbox"/> FILTERS
<input checked="" type="checkbox"/> HEATERS SAMPLE CHAMBER <u>49</u> °C	<input checked="" type="checkbox"/> QUARTZ STANDARD
<input checked="" type="checkbox"/> FLOW DETECTOR	<input checked="" type="checkbox"/> CALIBRATION
<input checked="" type="checkbox"/> PUMP HIGH SPEED	<input checked="" type="checkbox"/> PRINTER

<input checked="" type="checkbox"/> INDICATOR LIGHTS
<input checked="" type="checkbox"/> SIMULATOR SOLUTION SUPPLIER <u>Guth Labs</u> LOT # <u>13010</u> EXP. DATE <u>01-09-2015</u>
<input checked="" type="checkbox"/> SIMULATOR TEMP (34°C ± 0.2°C) <u>34.0</u> °C SIMULATOR SN <u>SD 2280</u> EXP. DATE <u>04-19-2014</u>

CALIBRATION CHECK - (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Mark the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

0.100% STANDARD - MUST READ BETWEEN 0.095% AND 0.105% INCLUSIVE
 0.080% STANDARD - MUST READ BETWEEN 0.076% AND 0.084% INCLUSIVE
 0.040% STANDARD - MUST READ BETWEEN 0.038% AND 0.042% INCLUSIVE

TEST 1 * <u>.096</u>	TEST 2 * <u>.097</u>	TEST 3 * <u>.097</u>
----------------------	----------------------	----------------------

PERFORM R.F.I. TEST (PRINTOUT ATTACHED)

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:
 (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS	(.0-.04) <u>1</u>	(.05-.09) <u>—</u>	(.10-.14) <u>—</u>	(.15-.19) <u>—</u>	OVER .19 <u>—</u>
----------	-------------------	--------------------	--------------------	--------------------	-------------------

LIST ANY NEW PARTS AND DESCRIBE ANY ALTERATION OR MODIFICATION THAT WAS MADE TO RESTORE THE INSTRUMENT TO OPERATE SATISFACTORILY AND WITHIN ESTABLISHED LIMITS (USE OTHER SIDE IF NECESSARY).

INSPECTING OFFICER	
SIGNATURE <u>Raymond A. Laird</u>	PRINT FULL NAME <u>Raymond A. Laird</u>
TYPE II PERMIT NUMBER/EXPIRATION DATE <u>220392</u>	TELEPHONE NUMBER <u>573-594-6186</u>

RETURN COMPLETED REPORT TO THE: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
2875 James Blvd.
Poplar Bluff, MO 63901



GUTH LABORATORIES, INC.

590 NORTH 6TH STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-504-5470

CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number 13010 of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on January 14, 2013, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain 0.1218% (w/vol) ethyl alcohol. The expiration date for this lot number is January 9, 2015 at 11:59 PM.

When used in a calibrated Simulator, operating at 34°C +/- .2°C, this solution will give a breath alcohol analysis instrument reading of 0.100 g/210L +/- 3%.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President
GUTH LABORATORIES, INC.

NIST Traceability:

Testing was conducted using Cerilliant Reference Standard lot number FN122211-02 whose values are traceable to NIST.

All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.

Face This Side Down - This Edge In First

BAC DataMaster
Evidence Ticket

STATE OF MISSOURI
VANDALIA POLICE DEPARTMENT


BAC DATAMASTER SERIAL NUMBER 201268
01/16/14

ARREST TIME: 01:01
SUBJECT NAME:
N/A
DOB: 01/01/01 SEX: M
STATE/D.L.: MO/N/A
ARRESTING OFFICER:
N/A
OFFICER I.D.: N/A
TESTING OFFICER:
LAIRD/RAYMOND/A
OFFICER I.D.: 202
PERMIT NUMBER: 220392
EXPIRATION DATE: 11/19/14
MISCELLANEOUS DATA:
RFI/TEST.

--- BREATH ANALYSIS ---

BLANK TEST	.000	15:11
INTERNAL STANDARD	VERIFIED	15:11
RADIO INTERFERENCE		

Operator Signature



2209-02

Face This Side Down - This Edge In First

**BAC DataMaster
Evidence Ticket**

STATE OF MISSOURI
VANDALIA POLICE DEPARTMENT

BAC DATAMASTER SERIAL NUMBER 201268
01/18/14

TESTING OFFICER:
LAIRD/RAYMOND/A
OFFICER I.D.: 202
PERMIT NUMBER: 220392
EXPIRATION DATE: 11/19/14
MISCELLANEOUS DATA:

--- SUPERVISOR MODE ---

BLANK TEST	.000	15:01
INTERNAL STANDARD	VERIFIED	15:01
EXTERNAL STANDARD	.096	15:01
BLANK TEST	.000	15:02
EXTERNAL STANDARD	.097	15:02
BLANK TEST	.000	15:03
EXTERNAL STANDARD	.097	15:03
BLANK TEST	.000	15:04

N = 3
SIM. = .1
AVG. = .0966

Operator Signature 

2208-02

Face This Side Down - This Edge In First

BAC DataMaster
Evidence Ticket

STATE OF MISSOURI
VANDALIA POLICE DEPARTMENT

BAC DATAMASTER SERIAL NUMBER 201268

01/16/14

14:18

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY

PROGRAM (04-07-2009): OKAY

HEATERS

SAMPLE CHAMBER: 49c

FLOW DETECTOR: OKAY

PUMP

HIGH SPEED: OKAY

DETECTOR: OKAY

FILTERS: OKAY


QUARTZ STANDARD: OKAY

CALIBRATION: OKAY

PRINTER TEST

!"#\$%&'()*+,-./0123456789:;<=>?@ABCDEFGHIJ
KLMNOPQRSTUVWXYZ[\]^_`abcdefghijklmnopq
rstuvwxyz{|}~

Operator Signature



2208-02

State of Missouri
DEPARTMENT OF HEALTH



P E R M I T
TYPE II



RAYMOND A LAIRD

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field repairs, and operate the following breath analyzer(s):

DATAMASTER

for the determination of the alcoholic content of blood from a sample of expired (alveolar) air. Issued under the provisions of sections 577.020 through 577.041, RSMo 1988.

Date 11/19/2012

Number 220392

Expires 11/19/2014

MO 680-0771 (7-88)

Director of State Public Health Laboratory

Director, Department of Health

Lab. 4 (97-88)