



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
 STATE PUBLIC HEALTH LABORATORY
 BREATH ALCOHOL PROGRAM
 DATAMASTER MAINTENANCE REPORT

RECEIVED
 By Carol Day at 2:43 pm, Nov 18, 2014

Complete this report at the time of the regular monthly preventive maintenance check (not to exceed 35 days).
 Complete this report whenever the instrument is serviced or repaired and whenever it is placed into service.
 Retain the original and send a copy within 15 days to the Breath Alcohol Program, DHSS.

DATAMASTER SN 201270	NAME OF AGENCY La Grange Police Department	DATE OF INSPECTION 11/06/2014
LOCATION OF INSTRUMENT (STREET AND CITY) 204 West Washington Street, La Grange		TIME OF INSPECTION 00:22 am

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

<input checked="" type="checkbox"/> DIAGNOSTIC CHECK (PRINTOUT ATTACHED)	DATE AND TIME (from printout) 11/06/14 00:22
<input checked="" type="checkbox"/> COMPUTER	<input checked="" type="checkbox"/> DETECTOR
<input checked="" type="checkbox"/> PROGRAM	<input checked="" type="checkbox"/> FILTERS
<input checked="" type="checkbox"/> HEATERS SAMPLE CHAMBER 49°C	<input checked="" type="checkbox"/> QUARTZ STANDARD
<input checked="" type="checkbox"/> FLOW DETECTOR	<input checked="" type="checkbox"/> CALIBRATION
<input checked="" type="checkbox"/> PUMP HIGH SPEED	<input checked="" type="checkbox"/> PRINTER

INDICATOR LIGHTS

SIMULATOR SOLUTION SUPPLIER Guth Laboratories Inc LOT # 14110 EXP. DATE 05/01/2016

SIMULATOR TEMP (34°C ± 0.2°C) 34 °C SIMULATOR SN DR 5388 EXP. DATE 04/29/2015

CALIBRATION CHECK - (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Mark the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

- 0.100% STANDARD - MUST READ BETWEEN 0.095% AND 0.105% INCLUSIVE
- 0.080% STANDARD - MUST READ BETWEEN 0.076% AND 0.084% INCLUSIVE
- 0.040% STANDARD - MUST READ BETWEEN 0.038% AND 0.042% INCLUSIVE

TEST 1 → .099	TEST 2 → .099	TEST 3 → .099
---------------	---------------	---------------

PERFORM R.F.I. TEST (PRINTOUT ATTACHED)

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:
 (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS	1	(0-.04)	10	(.05-.09)	0	(.10-.14)	0	(.15-.19)	0	OVER .19	0
----------	---	---------	----	-----------	---	-----------	---	-----------	---	----------	---

LIST ANY NEW PARTS AND DESCRIBE ANY ALTERATION OR MODIFICATION THAT WAS MADE TO RESTORE THE INSTRUMENT TO OPERATE SATISFACTORILY AND WITHIN ESTABLISHED LIMITS (USE OTHER SIDE IF NECESSARY).
 Datamaster 201270 complies with the Department of Health and Senior Services rules and regulations

INSPECTING OFFICER

SIGNATURE 	PRINT FULL NAME Don L Cibert
---------------	---------------------------------

TYPE II PERMIT NUMBER/EXPIRATION DATE 240364 10/17/2016	TELEPHONE NUMBER (673) 655-4099
--	------------------------------------

RETURN COMPLETED REPORT TO THE: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
 2875 James Blvd.
 Poplar Bluff, MO 63901



GUTH LABORATORIES, INC.

690 NORTH 87th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-564-5470

CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number 14110 of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on May 5, 2014, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain 0.1206% (w/vol) ethyl alcohol. The expiration date for this lot number is May 1, 2016 at 11:59 PM.

When used in a calibrated Simulator, operating at 34°C +/- .2°C, this solution will give a breath alcohol analysis instrument reading of 0.100 g/210L +/- 3%.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President
GUTH LABORATORIES, INC.

NIST Traceability:

Testing was conducted using Certified Reference Standard lot number FN122211-02 whose values are traceable to NIST.

All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.

BAC DataMaster Evidence Ticket

STATE OF MISSOURI
BAC DATAMASTER SERIAL NUMBER 201270
11/06/14
00:22

----- DIAGNOSTIC CHECK -----

COMPUTER: OKAY

PROGRAM (04-07-2009): OKAY

HEATERS
SAMPLE CHAMBER: 49c

FLOW DETECTOR: OKAY

PUMP
HIGH SPEED: OKAY

DETECTOR: OKAY

FILTERS: OKAY


----- QUARTZ STANDARD: OKAY

----- CALIBRATION: OKAY

PRINTER TEST

!"#\$%&'()*+,-./0123456789:; <=>?@ABCDEFGHIJ
KLMNOPQRSTUVWXYZ[\]^_`abcdefghijklmnop
qrstuvwxyz{|}~"

OPERATOR SIGNATURE



Card Stock No.
60021

RECORDER ALL SUPPLIES FROM N.P.A.S.
P.O. BOX 1435, MANSFIELD, OH 44901

**BAC DataMaster
Evidence Ticket**

STATE OF MISSOURI
BAC DATAMASTER SERIAL NUMBER 201270
11/06/14


TESTING OFFICER:
CYBERT/DUNN
OFFICER I. D. # 112
PERMIT NUMBER: 240364
EXPIRATION DATE: 10/17/16
MISCELLANEOUS DATA:
MONTHLY MAINTENANCE

----- SUPERVISOR MODE -----

BLANK TEST	.000	00:25
INTERNAL STANDARD	VERIFIED	00:25
EXTERNAL STANDARD	.099	00:25
BLANK TEST	.000	00:26
EXTERNAL STANDARD	.099	00:26
BLANK TEST	.000	00:27
EXTERNAL STANDARD	.099	00:27
BLANK TEST	.000	00:28

----- N = 3
----- SIM. = .1
----- AVG. = .099

OPERATOR SIGNATURE



Card Stock No.
60021

REORDER ALL SUPPLIES FROM N.P.A.S.
P.O. BOX 1435, MANSFIELD, OH 44901

BAC DataMaster
Evidence Ticket

STATE OF MISSOURI
BAC DATAMASTER SERIAL NUMBER 201270
11/06/14

ARREST TIME: 00:00
SUBJECT NAME:
RFI/TEST
DOB: 11/11/11 SEX: M
STATE/D.L.#: MO/11111111
ARRESTING OFFICER:
N/A
OFFICER I.D.#: N/A
TESTING OFFICER:
CIBERT/DON/L
OFFICER I.D.#: 112
PERMIT NUMBER: 240364
EXPIRATION DATE: 10/17/16
MISCELLANEOUS DATA:
RFI TEST
MONTHLY MAINTENANCE

----- BREATH ANALYSIS -----

_____	BLANK TEST	.000	00:30
_____	INTERNAL STANDARD	VERIFIED	00:30
_____	RADIO INTERFERENCE		

OPERATOR SIGNATURE



Card Stock No.
60021

REORDER ALL SUPPLIES FROM N.P.A.S.
P.O. BOX 1435, MANSFIELD, OH 44901



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM



PERMIT
TYPE II
DON L CIBERT

Is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

DATAMASTER

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 10/17/2014

NUMBER 240364

EXPIRES 10/17/2016

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

acting director

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES