



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
 STATE PUBLIC HEALTH LABORATORY
 BREATH ALCOHOL PROGRAM
DATAMASTER MAINTENANCE REPORT

RECEIVED 2/14/14-CD
 REPORT #6

Complete this report at the time of the regular monthly preventive maintenance check (not to exceed 30 days).
 Complete this report whenever the instrument is serviced or repaired and whenever it is placed into service.
 Retain the original and send a copy within 15 days to the Breath Alcohol Program, DHSS.

REVIEWED
 By Carol Day at 9:39 am, Mar 14, 2014

DATAMASTER SN 201271	NAME OF AGENCY Pleasant Hill Police Department	DATE OF INSPECTION 02/06/2014
-------------------------	---	----------------------------------

LOCATION OF INSTRUMENT (STREET AND CITY) 300 Commercial Street Pleasant Hill, MO 64080	TIME OF INSPECTION 10:42 pm
---	--------------------------------

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

- DIAGNOSTIC CHECK (PRINTOUT ATTACHED) DATE AND TIME (from printout) 02/06/14 at 22:42
- COMPUTER DETECTOR
- PROGRAM FILTERS
- HEATERS SAMPLE CHAMBER 49 °C QUARTZ STANDARD
- FLOW DETECTOR CALIBRATION
- PUMP HIGH SPEED PRINTER

INDICATOR LIGHTS

SIMULATOR SOLUTION SUPPLIER Repco LOT # 13001 EXP. DATE 03/07/2015

SIMULATOR TEMP (34°C ± 0.2°C) 34.0 °C SIMULATOR SN SD1967 EXP. DATE 12/12/2014

CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Mark the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

- 0.100% STANDARD - MUST READ BETWEEN 0.095% AND 0.105% INCLUSIVE
- 0.080% STANDARD - MUST READ BETWEEN 0.076% AND 0.084% INCLUSIVE
- 0.040% STANDARD - MUST READ BETWEEN 0.038% AND 0.042% INCLUSIVE

TEST 1 • .099	TEST 2 • .099	TEST 3 • .100
---------------	---------------	---------------

PERFORM R.F.I. TEST (PRINTOUT ATTACHED)

**INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:
 (DO NOT INCLUDE SELF-ADMINISTERED TESTS)**

REFUSALS	0	(.0-.04)	0	(.05-.09)	0	(.10-.14)	0	(.15-.19)	0	OVER .19	0
----------	---	----------	---	-----------	---	-----------	---	-----------	---	----------	---

LIST ANY NEW PARTS AND DESCRIBE ANY ALTERATION OR MODIFICATION THAT WAS MADE TO RESTORE THE INSTRUMENT TO OPERATE SATISFACTORILY AND WITHIN ESTABLISHED LIMITS (USE OTHER SIDE IF NECESSARY).

Instrument is operating with DOH guidelines.

INSPECTING OFFICER

SIGNATURE <i>Joe Miller</i>	PRINT FULL NAME Joe Miller
--------------------------------	-------------------------------

TYPE II PERMIT NUMBER/EXPIRATION DATE 230196 09/17/2015	TELEPHONE NUMBER (816) 540-9109
--	------------------------------------

RETURN COMPLETED REPORT TO THE: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
 2875 James Blvd.
 Poplar Bluff, MO 63901

CERTIFICATE OF ANALYSIS

MANUFACTURER AND SUPPLIER: RepCo Marketing, Inc.

LOT NUMBER: 13001

EXPIRATION DATE: March 7, 2015 at 11:59 p.m.

RepCo Marketing, Inc. certifies the following:

RepCo Marketing, Inc. manufactured, tested and supplied Lot Number 13001 of Alcohol Certified Solution for simulators. Random samples of said lot number were analyzed by an independent laboratory utilizing a gas chromatograph and found to contain .1215 gms/dl +/- .003 gms/dl wt/vol ethanol (95% Confidence).

The alcohol and distilled water used in the solution were found to be free of any interfering substance.

This solution will produce a vapor alcohol value of .100 +/- 3% gms/210L Breath when heated to 34 Degrees Celsius +/- 0.2 Degrees Celsius in a simulator (95% Confidence).

The date of manufacture for this lot number is March 8, 2013
The expiration date for this lot number is March 7, 2015 at
11:59 p.m.

This document is a true representation of the original Certificate of Analysis.



Cecil B. Garner, President
RepCo Marketing, Inc.

FACE THIS SIDE DOWN - THIS EDGE IN FIRST

BAC DataMaster
Evidence Ticket

STATE OF MISSOURI
PLEASANT HILL POLICE DEPARTMENT

BAC DATAMASTER SERIAL NUMBER 201271
02/06/14

TESTING OFFICER:
MILLER/JOE/D
OFFICER I.D.: 129
PERMIT NUMBER: 230196
EXPIRATION DATE: 09/17/15
MISCELLANEOUS DATA:

--- SUPERVISOR MODE ---

BLANK TEST	.000	23:19
INTERNAL STANDARD	VERIFIED	23:19
EXTERNAL STANDARD	.099	23:19
BLANK TEST	.000	23:20
EXTERNAL STANDARD	.099	23:20
BLANK TEST	.000	23:21
EXTERNAL STANDARD	.100	23:22
BLANK TEST	.000	23:22

N = 3
SIM. = .1
AVG. = .0993



OPERATOR SIGNATURE

Card Stock No.
60021
REORDER ALL SUPPLIES FROM N.P.A.S.
P.O. BOX 1435, MANSFIELD, OH 44901

FACE THIS SIDE DOWN - THIS EDGE IN FIRST

BAC DataMaster
Evidence Ticket

STATE OF MISSOURI
PLEASANT HILL POLICE DEPARTMENT

BAC DATAMASTER SERIAL NUMBER 201271
02/06/14
22:42

--- DIAGNOSTIC CHECK ---

COMPUTER:	OKAY
PROGRAM (04-07-2009):	OKAY
HEATERS	
SAMPLE CHAMBER:	49c
FLOW DETECTOR:	OKAY
PUMP	
HIGH SPEED:	OKAY
DETECTOR:	OKAY
FILTERS:	OKAY
QUARTZ STANDARD:	OKAY
CALIBRATION:	OKAY

PRINTER TEST

!"#\$%&'()*+,-./0123456789:;<=>?@ABCDEFGHIJ
KLMNOPQRSTUVWXYZ[\]^_`abcdefgijklmnop
qrstuvwxyz{|}~



OPERATOR SIGNATURE

Card Stock No.
60021
REORDER ALL SUPPLIES FROM N.P.A.S.
P.O. BOX 1435, MANSFIELD, OH 44901

FACE THIS SIDE DOWN - THIS EDGE IN FIRST

BAC DataMaster

Evidence Ticket

STATE OF MISSOURI
PLEASANT HILL POLICE DEPARTMENT

BAC DATAMASTER SERIAL NUMBER 201271
02/06/14

ARREST TIME: 22:00

SUBJECT NAME:

RFI/TEST

DOB: 02/06/14 SEX: M

STATE/D.L.: NA/N/A

ARRESTING OFFICER:

N/A

OFFICER I.D.: N/A

TESTING OFFICER:

MILLER/JOE/D

OFFICER I.D.: 129

PERMIT NUMBER: 230196

EXPIRATION DATE: 09/17/15


MISCELLANEOUS DATA:

N

--- BREATH ANALYSIS ---

BLANK TEST	.000	22:47
INTERNAL STANDARD	VERIFIED	22:47
RADIO INTERFERENCE		

OPERATOR SIGNATURE



Card Stock No.
60021

REORDER ALL SUPPLIES FROM N.P.A.S.
P.O. BOX 1435, MANSFIELD, OH 44901

State of Missouri
DEPARTMENT OF HEALTH



P E R M I T
TYPE II



JOE MILLER

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field repairs, and operate the following breath analyzer(s):

DATAMASTER

for the determination of the alcoholic content of blood from a sample of expired (alveolar) air. Issued under the provisions of sections 577.020 through 577.041, RSMo 1986.

Date 09/08/2011

Number 210224

Expires 09/08/2013

Director of State Public Health Laboratory

Director, Department of Health