



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 STATE PUBLIC HEALTH LABORATORY  
 BREATH ALCOHOL PROGRAM  
**DATAMASTER MAINTENANCE REPORT**

**RECEIVED**  
 By Carol Day at 9:25 am, Feb 06, 2014

Complete this report at the time of the regular monthly preventive maintenance check (not to exceed 35 days).  
 Complete this report whenever the instrument is serviced or repaired and whenever it is placed into service.  
 Retain the original and send a copy within 15 days to the Breath Alcohol Program, DHSS.

DATAMASTER SN <b>201273</b>	NAME OF AGENCY <b>Mountain View Police Department</b>	DATE OF INSPECTION <b>02/05/2014</b>
LOCATION OF INSTRUMENT (STREET AND CITY) <b>233 W. Second St. Mountain View, MO</b>		TIME OF INSPECTION <b>2:05 am</b>

**CHECKLIST:** Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

<input checked="" type="checkbox"/> DIAGNOSTIC CHECK (PRINTOUT ATTACHED)	DATE AND TIME (from printout) <b>02/05/2014 02:32</b>
<input checked="" type="checkbox"/> COMPUTER	<input checked="" type="checkbox"/> DETECTOR
<input checked="" type="checkbox"/> PROGRAM	<input checked="" type="checkbox"/> FILTERS
<input checked="" type="checkbox"/> HEATERS SAMPLE CHAMBER <b>47°C</b>	<input checked="" type="checkbox"/> QUARTZ STANDARD
<input checked="" type="checkbox"/> FLOW DETECTOR	<input checked="" type="checkbox"/> CALIBRATION
<input checked="" type="checkbox"/> PUMP HIGH SPEED	<input checked="" type="checkbox"/> PRINTER
<input checked="" type="checkbox"/> INDICATOR LIGHTS	
<input checked="" type="checkbox"/> SIMULATOR SOLUTION SUPPLIER <b>RepCo Marketing</b> LOT # <b>13002</b> EXP. DATE <b>06/19/2015</b>	
<input checked="" type="checkbox"/> SIMULATOR TEMP (34°C ± 0.2°C) <b>34.0</b> °C SIMULATOR SN <b>SD2673</b> EXP. DATE <b>08/06/2014</b>	

**CALIBRATION CHECK - (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)**

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Mark the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

0.100% STANDARD - MUST READ BETWEEN 0.095% AND 0.105% INCLUSIVE  
 0.080% STANDARD - MUST READ BETWEEN 0.076% AND 0.084% INCLUSIVE  
 0.040% STANDARD - MUST READ BETWEEN 0.038% AND 0.042% INCLUSIVE

TEST 1 <b>.100</b>	TEST 2 <b>.102</b>	TEST 3 <b>.103</b>
--------------------	--------------------	--------------------

**PERFORM R.F.I. TEST (PRINTOUT ATTACHED)**

**INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)**

REFUSALS	1	(.0-.04)	1	(.05-.09)	0	(.10-.14)	1	(.15-.19)	0	OVER .19	0
----------	---	----------	---	-----------	---	-----------	---	-----------	---	----------	---

LIST ANY NEW PARTS AND DESCRIBE ANY ALTERATION OR MODIFICATION THAT WAS MADE TO RESTORE THE INSTRUMENT TO OPERATE SATISFACTORILY AND WITHIN ESTABLISHED LIMITS (USE OTHER SIDE IF NECESSARY).

This instrument passed the standards set forth by the Missouri Department of Health and Senior Services. Solution is from RepCo Marketing with Lot No. 13002 Bottle No. 1438.

<b>INSPECTING OFFICER</b>	
SIGNATURE <i>[Signature]</i> <b>DN: 173</b>	PRINT FULL NAME <b>David Karr/ 123</b>
TYPE II PERMIT NUMBER/EXPIRATION DATE <b>230072 04/26/2015</b>	TELEPHONE NUMBER <b>(417) 934-2525</b>

RETURN COMPLETED REPORT TO THE: **Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office  
 2875 James Blvd.  
 Poplar Bluff, MO 63901**

**REPCO MARKETING INC.**3101-189 STONYBROOK DRIVE  
RALEIGH, N.C. 27604  
919-876-5480**CERTIFICATE OF ANALYSIS****MANUFACTURER AND SUPPLIER: RepCo Marketing, Inc.****LOT NUMBER: 13002****EXPIRATION DATE: June 19, 2015 at 11:59 p.m.**

RepCo Marketing, Inc. certifies the following:

RepCo Marketing, Inc. manufactured, tested and supplied Lot Number 13002 of Alcohol Certified Solution for simulators. Random samples of said lot number were analyzed by an independent laboratory utilizing a gas chromatograph and found to contain .1217 gms/dl +/- .003 gms/dl wt/vol ethanol (95% Confidence).

The alcohol and distilled water used in the solution were found to be free of any interfering substance.

This solution will produce a vapor alcohol value of .100 +/- 3% gms/210L Breath when heated to 34 Degrees Celsius +/- 0.2 Degrees Celsius in a simulator (95% Confidence).

The date of manufacture for this lot number is June 20, 2013  
The expiration date for this lot number is June 19, 2015 at  
11:59 p.m.

This document is a true representation of the original Certificate of Analysis.



Cecil B. Garner, President  
RepCo Marketing, Inc.

# BAC DataMaster Evidence Ticket

STATE OF MISSOURI  
MOUNTAIN VIEW POLICE DEPARTMENT


BAC DATA MASTER SERIAL NUMBER 201218  
02/05/14

ARREST TIME: 02:15  
SUBJECT NAME:  
FLOW/TEST  
DOB: 01/02/03 SEX: M  
STATE/D.L.: MO/0987654321  
ARRESTING OFFICER:  
NA  
OFFICER I.D.:  
TESTING OFFICER:  
KARR/D/M  
OFFICER I.D.: 123  
PERMIT NUMBER: 200072  
EXPIRATION DATE: 04/28/15  
MISCELLANEOUS DATA:

--- BREATH ANALYSIS ---

BLANK TEST	.000	02:47
INTERNAL STANDARD	VERIFIED	02:47
SUBJECT SAMPLE	.104	02:48
BLANK TEST	.000	02:48

Operator Signature



# BAC DataMaster Evidence Ticket

STATE OF MISSOURI  
MOUNTAIN VIEW POLICE DEPARTMENT

BAC DATAMASTER SERIAL NUMBER 801273

02/05/14

02:32

### --- DIAGNOSTIC CHECK ---

COMPLIANCE:	OKAY
PROGRAM (04-07-2009):	OKAY
HEATERS	
SAMPLE CHAMBER:	47E
FLOW DETECTOR:	OKAY
PUMP	
HIGH SPEED:	OKAY
DETECTOR:	OKAY
FILTERS:	OKAY
QUARTZ STANDARD:	OKAY
CALIBRATION:	OKAY

### PRINTER TEST

!"#\$%&'()\*+,-./0123456789:;<=>?@ABCDEFGHI  
 HIJKLMNOPQRSTUVWXYZ[\]^\_`abcdefghijklmnop  
 qrstuvwxyz{|}~

Operator Signature



173

# BAC DataMaster Evidence Ticket

STATE OF MISSOURI  
MOUNTAIN VIEW POLICE DEPARTMENT

BAC DATA MASTER SERIAL NUMBER 201279  
02/05/14

TESTING OFFICER:  
KARR/D/M  
OFFICER I.D.: 123  
PERMIT NUMBER: 230072  
EXPIRATION DATE: 04/26/15  
MISCELLANEOUS DATA:

--- SUPERVISOR MODE ---

BLANK TEST	.000	02:35
INTERNAL STANDARD	VERIFIED	02:35
EXTERNAL STANDARD	.100	02:36
BLANK TEST	.000	02:36
EXTERNAL STANDARD	.102	02:37
BLANK TEST	.000	02:37
EXTERNAL STANDARD	.103	02:38
BLANK TEST	.000	02:38

N = 3  
SIM. = .1  
AVG. = .1016

Operator Signature



179

# BAC DataMaster Evidence Ticket

STATE OF MISSOURI  
MOUNTAIN VIEW POLICE DEPARTMENT

BAC DATAMASTER SERIAL NUMBER 201273  
02/05/14

ARREST TIME: 02:15  
SUBJECT NAME:  
RFI/TEST  
DOB: 01/02/83 SEX: M  
STATE/D.L.: MO/8987654321  
ARRESTING OFFICER:  
NA  
OFFICER I.D.:  
TESTING OFFICER:  
KARR/D/M  
OFFICER I.D.: 123  
PERMIT NUMBER: 230072  
EXPIRATION DATE: 04/26/15  
MISCELLANEOUS DATA:

--- BREATH ANALYSIS ---

BLANK TEST .000 02:41  
INTERNAL STANDARD VERIFIED 02:41  
RADIO INTERFERENCE

Operator Signature



123



STATE OF MISSOURI  
DEPARTMENT OF HEALTH AND SENIOR SERVICES  
BREATH ALCOHOL PROGRAM



**PERMIT**  
**TYPE II**

**DAVID KARR**

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

**DATAMASTER**

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 04/26/2013

NUMBER 230072

EXPIRES 04/26/2015

MO 590-0771 (9-10)

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

Acting Director

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

LAB-4 (Re-10)