



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
 STATE PUBLIC HEALTH LABORATORY
 BREATH ALCOHOL PROGRAM
DATAMASTER MAINTENANCE REPORT

RECEIVED
 By Carol Day at 10:02 am, Jul 07, 2014

REPORT #6

Complete this report at the time of the regular monthly preventive maintenance check (not to exceed 35 days).
 Complete this report whenever the instrument is serviced or repaired and whenever it is placed into service.
 Retain the original and send a copy within 15 days to the Breath Alcohol Program, DHSS.

DATAMASTER SN 201273	NAME OF AGENCY Mountain View Police Department	DATE OF INSPECTION 07/03/2014
LOCATION OF INSTRUMENT (STREET AND CITY) 233 W. 2nd St. Mountain View, MO		TIME OF INSPECTION 8:45 pm

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

<input checked="" type="checkbox"/> DIAGNOSTIC CHECK (PRINTOUT ATTACHED)	DATE AND TIME (from printout) <u>07/03/2014 20:48</u>
<input checked="" type="checkbox"/> COMPUTER	<input checked="" type="checkbox"/> DETECTOR
<input checked="" type="checkbox"/> PROGRAM	<input checked="" type="checkbox"/> FILTERS
<input checked="" type="checkbox"/> HEATERS SAMPLE CHAMBER <u>47</u> °C	<input checked="" type="checkbox"/> QUARTZ STANDARD
<input checked="" type="checkbox"/> FLOW DETECTOR	<input checked="" type="checkbox"/> CALIBRATION
<input checked="" type="checkbox"/> PUMP HIGH SPEED	<input checked="" type="checkbox"/> PRINTER

INDICATOR LIGHTS

SIMULATOR SOLUTION SUPPLIER RepCo Marketing Inc LOT # 13002 EXP. DATE 06/19/2015

SIMULATOR TEMP (34°C ± 0.2°C) 34.0 °C SIMULATOR SN 094948 EXP. DATE 06/05/2015

CALIBRATION CHECK - (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Mark the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

- 0.100% STANDARD - MUST READ BETWEEN 0.095% AND 0.105% INCLUSIVE
- 0.080% STANDARD - MUST READ BETWEEN 0.076% AND 0.084% INCLUSIVE
- 0.040% STANDARD - MUST READ BETWEEN 0.038% AND 0.042% INCLUSIVE

TEST 1 .099	TEST 2 .100	TEST 3 .101
--------------	--------------	--------------

PERFORM R.F.I. TEST (PRINTOUT ATTACHED)

**INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:
 (DO NOT INCLUDE SELF-ADMINISTERED TESTS)**

REFUSALS	0	(.0-.04)	1	(.05-.09)	0	(.10-.14)	2	(.15-.19)	0	OVER .19	2
----------	---	----------	---	-----------	---	-----------	---	-----------	---	----------	---

LIST ANY NEW PARTS AND DESCRIBE ANY ALTERATION OR MODIFICATION THAT WAS MADE TO RESTORE THE INSTRUMENT TO OPERATE SATISFACTORILY AND WITHIN ESTABLISHED LIMITS (USE OTHER SIDE IF NECESSARY).

Instrument is operating within the guidelines set forth by the Missouri Department of Health and Senior Services. RepCo Marketing Inc. Lot No. 13002; Bottle No. 1411; Expiration Date 06-19-2015. Previous Attempt on 07/03/2014 resulted in Simulator difficulty. Changed simulator and worked properly.

INSPECTING OFFICER

SIGNATURE 	PRINT FULL NAME David Karr/ DSN: 123
---------------	---

TYPE II PERMIT NUMBER/EXPIRATION DATE 230072 04/26/2015	TELEPHONE NUMBER (417) 934-2525
--	------------------------------------

RETURN COMPLETED REPORT TO THE: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
 2875 James Blvd.
 Poplar Bluff, MO 63901



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

**PERMIT
TYPE II**

DAVID KARR

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

DATAMASTER

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 04/26/2013
NUMBER 230072
EXPIRES 04/26/2015

W. H. Karr

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

Shae Vankersdyk

Acting Director

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES
LAB-4 (R6-10)

MO 580-0771 (6-10)



CERTIFICATE OF ANALYSIS

MANUFACTURER AND SUPPLIER: RepCo Marketing, Inc.
LOT NUMBER: 13002
EXPIRATION DATE: June 19, 2015 at 11:59 p.m.

RepCo Marketing, Inc. certifies the following:

RepCo Marketing, Inc. manufactured, tested and supplied Lot Number 13002 of Alcohol Certified Solution for simulators. Random samples of said lot number were analyzed by an independent laboratory utilizing a gas chromatograph and found to contain .1217 gms/dl +/- .003 gms/dl wt/vol ethanol (95% Confidence).

The alcohol and distilled water used in the solution were found to be free of any interfering substance.

This solution will produce a vapor alcohol value of .100 +/- 3% gms/210L Breath when heated to 34 Degrees Celsius +/- 0.2 Degrees Celsius in a simulator (95% Confidence).

The date of manufacture for this lot number is June 20, 2013
The expiration date for this lot number is June 19, 2015 at 11:59 p.m.

This document is a true representation of the original Certificate of Analysis.



Cecil B. Garner, President
RepCo Marketing, Inc.

Face This Side Down - This Edge In First

BAC DataMaster
Evidence Ticket

STATE OF MISSOURI
MOUNTAIN VIEW POLICE DEPARTMENT

BAC DATAMASTER SERIAL NUMBER 201273
07/03/14
20:40

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY
PROGRAM (84-07-2000): OKAY
HEATERS
SAMPLE CHAMBER: 47c
FLOW DETECTOR: OKAY
PUMP
HIGH SPEED: OKAY
DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY

PRINTER TEST

!"#\$%&'()*+,-./0123456789:; <=>?@ABCDEFGHIJ
KLMNOPQRSTUVWXYZ[\]^_`abcdefghijklmnopq
rstuvwxyz{|}~

Operator Signature



123

Face This Side Down - This Edge In First

BAC DataMaster Evidence Ticket

STATE OF MISSOURI
MOUNTAIN VIEW POLICE DEPARTMENT

BAC DATAMASTER SERIAL NUMBER 201273
07/03/14

TESTING OFFICER:
KARR/D/M
OFFICER I.D.: 123
PERMIT NUMBER: 230072
EXPIRATION DATE: 04/26/15
MISCELLANEOUS DATA:

--- SUPERVISOR MODE ---

BLANK TEST	.000	20:50
INTERNAL STANDARD	VERIFIED	20:50
EXTERNAL STANDARD	.099	20:50
BLANK TEST	.000	20:51
EXTERNAL STANDARD	.100	20:51
BLANK TEST	.000	20:52
EXTERNAL STANDARD	.101	20:52
BLANK TEST	.000	20:53

N = 3
SIM. = .1
AVG. = .1

Operator Signature



2208-02

Face This Side Down - This Edge In First

BAC DataMaster Evidence Ticket

STATE OF MISSOURI
MOUNTAIN VIEW POLICE DEPARTMENT

BAC DATAMASTER SERIAL NUMBER 201273
07/03/14

ARREST TIME: 20:30
SUBJECT NAME:
TEST

DOB: 04/03/01 SEX: M

STATE/D.L.: MO/1

ARRESTING OFFICER:
NR

OFFICER I.D.:

TESTING OFFICER:

KARR/D/T

OFFICER I.D.: 123

PERMIT NUMBER: 230072

EXPIRATION DATE: 04/26/13

MISCELLANEOUS DATA:

--- BREATH ANALYSIS ---

BLANK TEST	.000	20:56
INTERNAL STANDARD	VERIFIED	20:56
RADIO INTERFERENCE		

Operator Signature



123

2208-02