



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
 STATE PUBLIC HEALTH LABORATORY
 BREATH ALCOHOL PROGRAM
DATAMASTER MAINTENANCE REPORT

RECEIVED 2/14/14 CD REPORT #6

REVIEWED
 By Carol Day at 10:17 am, Mar 20, 2014

Complete this report at the time of the regular monthly preventive maintenance check (not to exceed 3 months).
 Complete this report whenever the instrument is serviced or repaired and whenever it is placed into service.
 Retain the original and send a copy within 15 days to the Breath Alcohol Program, DHSS.

DATAMASTER SN 201275	NAME OF AGENCY Potosi PD	DATE OF INSPECTION 02/07/2014
LOCATION OF INSTRUMENT (STREET AND CITY) One Police Plaza Potosi, MO 63664		TIME OF INSPECTION 4:11 pm

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

<input checked="" type="checkbox"/> DIAGNOSTIC CHECK (PRINTOUT ATTACHED)	DATE AND TIME (from printout) 02/07/2014 1611
<input checked="" type="checkbox"/> COMPUTER	<input checked="" type="checkbox"/> DETECTOR
<input checked="" type="checkbox"/> PROGRAM	<input checked="" type="checkbox"/> FILTERS
<input checked="" type="checkbox"/> HEATERS SAMPLE CHAMBER 49 °C	<input checked="" type="checkbox"/> QUARTZ STANDARD
<input checked="" type="checkbox"/> FLOW DETECTOR	<input checked="" type="checkbox"/> CALIBRATION
<input checked="" type="checkbox"/> PUMP HIGH SPEED	<input checked="" type="checkbox"/> PRINTER
<input checked="" type="checkbox"/> INDICATOR LIGHTS	
<input checked="" type="checkbox"/> SIMULATOR SOLUTION SUPPLIER RepCo	LOT # 13002 EXP. DATE 06/19/2015
<input checked="" type="checkbox"/> SIMULATOR TEMP (34°C ± 0.2°C) 34.00 °C	SIMULATOR SN SD1312 EXP. DATE 07/10/2014

CALIBRATION CHECK - (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Mark the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

0.100% STANDARD - MUST READ BETWEEN 0.095% AND 0.105% INCLUSIVE
 0.080% STANDARD - MUST READ BETWEEN 0.076% AND 0.084% INCLUSIVE
 0.040% STANDARD - MUST READ BETWEEN 0.038% AND 0.042% INCLUSIVE

TEST 1 → .097	TEST 2 → .097	TEST 3 → .097
---------------	---------------	---------------

PERFORM R.F.I. TEST (PRINTOUT ATTACHED)

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS	0	(0-.04)	1	(.05-.09)	0	(.10-.14)	0	(.15-.19)	1	OVER .19	1
----------	---	---------	---	-----------	---	-----------	---	-----------	---	----------	---

LIST ANY NEW PARTS AND DESCRIBE ANY ALTERATION OR MODIFICATION THAT WAS MADE TO RESTORE THE INSTRUMENT TO OPERATE SATISFACTORILY AND WITHIN ESTABLISHED LIMITS (USE OTHER SIDE IF NECESSARY).

This instrument operates within established limits

INSPECTING OFFICER

SIGNATURE Lt. Michael Gum DSN 102	PRINT FULL NAME Lt. Michael Gum DSN 102
TYPE II PERMIT NUMBER/EXPIRATION DATE 220359 10/11/2014	TELEPHONE NUMBER (573) 438-5468

RETURN COMPLETED REPORT TO THE: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
 2875 James Blvd.
 Poplar Bluff, MO 63901

CERTIFICATE OF ANALYSIS

MANUFACTURER AND SUPPLIER: RepCo Marketing, Inc.

LOT NUMBER: 13002

EXPIRATION DATE: June 19, 2015 at 11:59 p.m.

RepCo Marketing, Inc. certifies the following:


RepCo Marketing, Inc. manufactured, tested and supplied Lot Number 13002 of Alcohol Certified Solution for simulators. Random samples of said lot number were analyzed by an independent laboratory utilizing a gas chromatograph and found to contain .1217 gms/dl +/- .003 gms/dl wt/vol ethanol (95% Confidence).

The alcohol and distilled water used in the solution were found to be free of any interfering substance.

This solution will produce a vapor alcohol value of .100 +/- 3% gms/210L: Breath when heated to 34 Degrees Celsius +/- 0.2 Degrees Celsius in a simulator (95% Confidence).

The date of manufacture for this lot number is June 20, 2013
The expiration date for this lot number is June 19, 2015 at 11:59 p.m.

This document is a true representation of the original Certificate of Analysis.




Cecil B. Garner, President
RepCo Marketing, Inc.

State of Missouri
DEPARTMENT OF HEALTH



P E R M I T
TYPE II



WYMICHELAN, GUY

is hereby authorized to instruct and supervise operators, train instructors, inspect
callouts, perform tests, repairs, and operate the following breath analyzer(s):

DATAMASTER

for the transportation of the alcoholic content of blood from a sample of expired (air/poler)
air tested under the provisions of sections 57.020 through 57.041, RSMo 1986.

Exp. 10/11/2012

Number 220859

Expires 10/11/2014

MOSSMITH, C-00

W. A. ...

Director of State Public Health Laboratory

[Signature]

Director, State Public Health Laboratory

MO-4 (10/1/00)

FACE THIS SIDE DOWN - THIS EDGE IN FIRST

BAC DataMaster

Evidence Ticket

STATE OF OHIO
POLICE DEPARTMENT

DATA MASTER SERIAL NUMBER 281275
02-07-14

ARREST DATE: 10/10/10

SUBJECT NAME:

DOE, JOHN

DOB: 10/10/10 SEX: M

STATE ID: 10-123456789

ARRESTING OFFICER:

TEST

OFFICER I.D.:

TESTING OFFICER:

000-000

OFFICER I.D. #: 100

PERMIT NUMBER: 220000

EXPIRES DATE: 10-11-14

MISCELLANEOUS DATA:

R.F.I. TEST

TEST RESULTS ANALYSIS

SUBJECT TEST	1000	10:25
INTERNAL STANDARD	VEH 100	10:26
RADIO INTERFERENCE		

OPERATOR SIGNATURE

CARD STK #
60036

REORDER ALL SUPPLIES FROM N.P.A.S.
2260 NORTH MAIN, MANSFIELD, OH 44903 419-526-6727 (NPAS)