



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
 STATE PUBLIC HEALTH LABORATORY
 BREATH ALCOHOL PROGRAM
DATAMASTER MAINTENANCE REPORT

REPORT #6

RECEIVED
 By Carol Day at 11:54 am, Sep 10, 2014

Complete this report at the time of the regular monthly preventive maintenance check (not to exceed 35 days since last maintenance check).
 Complete this report whenever the instrument is serviced or repaired and whenever it is placed into service.
 Retain the original and send a copy within 15 days to the Breath Alcohol Program, DHSS.

DATAMASTER SN 201277	NAME OF AGENCY Carterville Police Department	DATE OF INSPECTION 09/03/2014
LOCATION OF INSTRUMENT (STREET AND CITY) 1200 E. 1st Street Carterville, MO 64835		TIME OF INSPECTION 10:22 pm

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

<input checked="" type="checkbox"/> DIAGNOSTIC CHECK (PRINTOUT ATTACHED)	DATE AND TIME (from printout) <u>09/03/2014 / 2222 hrs</u>
<input checked="" type="checkbox"/> COMPUTER	<input checked="" type="checkbox"/> DETECTOR
<input checked="" type="checkbox"/> PROGRAM	<input checked="" type="checkbox"/> FILTERS
<input checked="" type="checkbox"/> HEATERS SAMPLE CHAMBER <u>49</u> °C	<input checked="" type="checkbox"/> QUARTZ STANDARD
<input checked="" type="checkbox"/> FLOW DETECTOR	<input checked="" type="checkbox"/> CALIBRATION
<input checked="" type="checkbox"/> PUMP HIGH SPEED	<input checked="" type="checkbox"/> PRINTER

<input checked="" type="checkbox"/> INDICATOR LIGHTS
<input checked="" type="checkbox"/> SIMULATOR SOLUTION SUPPLIER <u>Guth Laboratories Inc.</u> LOT # <u>14110</u> EXP. DATE <u>05/01/2016</u>
<input checked="" type="checkbox"/> SIMULATOR TEMP (34°C ± 0.2°C) <u>34.0</u> °C SIMULATOR SN <u>094948</u> EXP. DATE <u>07/14/2015</u>

CALIBRATION CHECK - (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Mark the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

0.100% STANDARD - MUST READ BETWEEN 0.095% AND 0.105% INCLUSIVE
 0.080% STANDARD - MUST READ BETWEEN 0.076% AND 0.084% INCLUSIVE
 0.040% STANDARD - MUST READ BETWEEN 0.038% AND 0.042% INCLUSIVE

TEST 1 • <u>.097</u>	TEST 2 • <u>.097</u>	TEST 3 • <u>.097</u>
----------------------	----------------------	----------------------

PERFORM R.F.I. TEST (PRINTOUT ATTACHED)

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:
 (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS	1	(0-.04)	0	(.05-.09)	1	(.10-.14)	0	(.15-.19)	2	OVER .19	0
----------	---	---------	---	-----------	---	-----------	---	-----------	---	----------	---

LIST ANY NEW PARTS AND DESCRIBE ANY ALTERATION OR MODIFICATION THAT WAS MADE TO RESTORE THE INSTRUMENT TO OPERATE SATISFACTORILY AND WITHIN ESTABLISHED LIMITS (USE OTHER SIDE IF NECESSARY).

Instrument is operating within MODHSS specifications

INSPECTING OFFICER	
SIGNATURE 	PRINT FULL NAME Christopher Shonk
TYPE II PERMIT NUMBER/EXPIRATION DATE 220432 / 12/27/2014	TELEPHONE NUMBER (417) 673-2616

RETURN COMPLETED REPORT TO THE:
 Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
 2875 James Blvd.
 Poplar Bluff, MO 63901

BAC DataMaster Evidence Ticket

STATE OF MISSOURI
CARTERVILLE POLICE DEPARTMENT

BAC DATAMASTER SERIAL NUMBER 201277
09/03/14


TESTING OFFICER:
SHONK/CHRISTOPHER/M
OFFICER I.D.: 504
PERMIT NUMBER: 220432
EXPIRATION DATE: 12/27/14
MISCELLANEOUS DATA:

--- SUPERVISOR MODE ---

BLANK TEST	.000	22:25
INTERNAL STANDARD	VERIFIED	22:25
EXTERNAL STANDARD	.097	22:25
BLANK TEST	.000	22:26
EXTERNAL STANDARD	.097	22:26
BLANK TEST	.000	22:27
EXTERNAL STANDARD	.097	22:27
BLANK TEST	.000	22:28

N = 3
SIM. = .1
RWG. = .097

Operator Signature

 #504

BAC DataMaster Evidence Ticket

STATE OF MISSOURI
CARTERVILLE POLICE DEPARTMENT

BAC DATAMASTER SERIAL NUMBER 201277
09/03/14
22:22

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY
PROGRAM (04-07-2009): OKAY
HEATERS
SAMPLE CHAMBER: 49c
FLOW DETECTOR: OKAY
PUMP
HIGH SPEED: OKAY
DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY

PRINTER TEST
!##%&'()*+,-./0123456789:;<=>?@ABCDEFGHIJ
KLMNOPQRSTUVWXYZ[\]^_`abcde fghijklmno
pqrstuvwxyz{|}~"

Operator Signature

 #504

BAC DataMaster
Evidence Ticket

STATE OF MISSOURI
CARTERSVILLE POLICE DEPARTMENT

BAC DATAMASTER SERIAL NUMBER 201277
09/03/14

ARREST TIME: 13:00
SUBJECT NAME:
SOBER/SAMPLE
DOB: 12/12/12 SEX: M
STATE/D.L.: MO/2
ARRESTING OFFICER:
SHONK/C/M
OFFICER I.D.: 504
TESTING OFFICER:
SHONK/CHRISTOPHER/M
OFFICER I.D.: 504
PERMIT NUMBER: 220432
EXPIRATION DATE: 12/27/14
MISCELLANEOUS DATA:

--- BREATH ANALYSIS ---

BLANK TEST	.000	22:33
INTERNAL STANDARD	VERIFIED	22:33
SUBJECT SAMPLE	.000	22:33
BLANK TEST	.000	22:34

Operator Signature

 #504

BAC DataMaster
Evidence Ticket

STATE OF MISSOURI
CARTERSVILLE POLICE DEPARTMENT

BAC DATAMASTER SERIAL NUMBER 201277
09/03/14

ARREST TIME: 12:00
SUBJECT NAME:
RFI/CHECK
DOB: 12/12/12 SEX: M
STATE/D.L.: MO/1
ARRESTING OFFICER:
SHONK/C/M
OFFICER I.D.: 504
TESTING OFFICER:
SHONK/CHRISTOPHER/M
OFFICER I.D.: 504
PERMIT NUMBER: 220432
EXPIRATION DATE: 12/27/14
MISCELLANEOUS DATA:

--- BREATH ANALYSIS ---

BLANK TEST	.000	22:30
INTERNAL STANDARD	VERIFIED	22:30
RADIO INTERFERENCE		

Operator Signature

 #504



GUTH LABORATORIES, INC.

590 NORTH 67TH STREET • HARRISBURG PA 17111-4511 • TELEPHONE 717-564-5470

CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number **14110** of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on **May 5, 2014**, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain **0.1206%** (w/vol) ethyl alcohol. The expiration date for this lot number is **May 1, 2016** at 11:59 PM.

When used in a calibrated Simulator, operating at $34^{\circ}\text{C} \pm .2^{\circ}\text{C}$, this solution will give a breath alcohol analysis instrument reading of $0.100 \text{ g}/210\text{L} \pm 3\%$.

The alcohol and water used in this solution were free of test interfering substances.


Ted L. Pauley, President
GUTH LABORATORIES, INC.

NIST Traceability:


Testing was conducted using Cerilliant Reference Standard lot number FNI22211-02 whose values are traceable to NIST.

All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.

State of Missouri
DEPARTMENT OF HEALTH



P E R M I T
TYPE II



CHRISTOPHER SHONK

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field repairs, and operate the following breath analyzer(s):

DATA MASTER

for the determination of the alcoholic content of blood from a sample of expired (alveolar) air. Issued under the provisions of sections 577.020 through 577.041, RSMo 1986.

Date 12/27/2012

Number 220432

Expires 12/27/2014

Director of State Public Health Laboratory

ACTING DIRECTOR

Director, Department of Health