



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
 STATE PUBLIC HEALTH LABORATORY
 BREATH ALCOHOL PROGRAM
DATAMASTER MAINTENANCE REPORT

RECEIVED 6/11/14-CD REPORT #6

REVIEWED
 By Carol Day at 1:53 pm, Jun 26, 2014

Complete this report at the time of the regular monthly preventive maintenance check (not to exceed 35 days).
 Complete this report whenever the instrument is serviced or repaired and whenever it is placed into service.
 Retain the original and send a copy within 15 days to the Breath Alcohol Program, DHSS.

DATAMASTER SN 201282	NAME OF AGENCY Carthage Police	DATE OF INSPECTION 06/03/2014
LOCATION OF INSTRUMENT (STREET AND CITY) 310 W 4th St. Carthage		TIME OF INSPECTION 1:00 pm

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

<input checked="" type="checkbox"/> DIAGNOSTIC CHECK (PRINTOUT ATTACHED)	DATE AND TIME (from printout) 06/03/2014 13:03
<input checked="" type="checkbox"/> COMPUTER	<input checked="" type="checkbox"/> DETECTOR
<input checked="" type="checkbox"/> PROGRAM	<input checked="" type="checkbox"/> FILTERS
<input checked="" type="checkbox"/> HEATERS SAMPLE CHAMBER 49 °C	<input checked="" type="checkbox"/> QUARTZ STANDARD
<input checked="" type="checkbox"/> FLOW DETECTOR	<input checked="" type="checkbox"/> CALIBRATION
<input checked="" type="checkbox"/> PUMP HIGH SPEED	<input checked="" type="checkbox"/> PRINTER

<input checked="" type="checkbox"/> INDICATOR LIGHTS
<input checked="" type="checkbox"/> SIMULATOR SOLUTION SUPPLIER RepCo LOT # 13002 EXP. DATE 06/19/2015
<input checked="" type="checkbox"/> SIMULATOR TEMP (34°C ± 0.2°C) 34.0 °C SIMULATOR SN SD2225 EXP. DATE 08/01/2014

CALIBRATION CHECK - (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Mark the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

0.100% STANDARD - MUST READ BETWEEN 0.095% AND 0.105% INCLUSIVE
 0.080% STANDARD - MUST READ BETWEEN 0.076% AND 0.084% INCLUSIVE
 0.040% STANDARD - MUST READ BETWEEN 0.038% AND 0.042% INCLUSIVE

TEST 1 .098	TEST 2 .099	TEST 3 .098
-------------	-------------	-------------

PERFORM R.F.I. TEST (PRINTOUT ATTACHED)

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:
 (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS 0	(0-.04) 0	(.05-.09) 0	(.10-.14) 0	(.15-.19) 0	OVER .19 0
------------	-----------	-------------	-------------	-------------	------------

LIST ANY NEW PARTS AND DESCRIBE ANY ALTERATION OR MODIFICATION THAT WAS MADE TO RESTORE THE INSTRUMENT TO OPERATE SATISFACTORILY AND WITHIN ESTABLISHED LIMITS (USE OTHER SIDE IF NECESSARY).

INSTRUMENT OPERATING WITHIN DOH GUIDELINES.
 RepCo Bottle #0448 Exp. 06/19/2015

INSPECTING OFFICER	
SIGNATURE 	PRINT FULL NAME Tim Wilson
TYPE II PERMIT NUMBER/EXPIRATION DATE 230172 08/14/2014	TELEPHONE NUMBER (417) 358-8177

RETURN COMPLETED REPORT TO THE:
 Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
 2875 James Blvd.
 Poplar Bluff, MO 63901

CERTIFICATE OF ANALYSIS

MANUFACTURER AND SUPPLIER: RepCo Marketing, Inc.

LOT NUMBER: 13002

EXPIRATION DATE: June 19, 2015 at 11:59 p.m.

RepCo Marketing, Inc. certifies the following:


RepCo Marketing, Inc. manufactured, tested and supplied Lot Number 13002 of Alcohol Certified Solution for simulators. Random samples of said lot number were analyzed by an independent laboratory utilizing a gas chromatograph and found to contain .1217 gms/dl +/- .003 gms/dl wt/vol ethanol (95% Confidence).

The alcohol and distilled water used in the solution were found to be free of any interfering substance.

This solution will produce a vapor alcohol value of .100 +/- 3% gms/210L Breath when heated to 34 Degrees Celsius +/- 0.2 Degrees Celsius in a simulator (95% Confidence).

The date of manufacture for this lot number is June 20, 2013
The expiration date for this lot number is June 19, 2015 at
11:59 p.m.

This document is a true representation of the original Certificate of Analysis.



Cecil B. Garner, President
RepCo Marketing, Inc.

BAC DataMaster
Evidence Ticket

STATE OF MISSOURI
CARTHAGE POLICE DEPARTMENT

BAC DATAMASTER SERIAL NUMBER 201282
06/03/14

ARREST TIME: 00:00

SUBJECT NAME:

WILSON

DOB: 11/11/11

SEX: M

STRTE/O.L.# NO/11111

ARRESTING OFFICER:

WILSON

OFFICER I.D.# 102

TESTING OFFICER:

WILSON

OFFICER I.D.# 102

PERMIT NUMBER: 230172

EXPIRATION DATE: 08/14/15

MISCELLANEOUS DATA:

REF TEST

--- BREATH ANALYSIS ---

BLANK TEST

.000

INTERNAL STANDARD

.000

EXTERNAL STANDARD

.000

BLANK TEST

.000

EXTERNAL STANDARD

.000

INTERNAL STANDARD

.000

EXTERNAL STANDARD

.000

BLANK TEST

.000

Operator Signature



BAC DataMaster
Evidence Ticket

STATE OF MISSOURI
CARTHAGE POLICE DEPARTMENT

BAC DATAMASTER SERIAL NUMBER 201282
06/03/14

TESTING OFFICER:

WILSON

OFFICER I.D.# 102

PERMIT NUMBER: 230172

EXPIRATION DATE: 08/14/15

MISCELLANEOUS DATA:

MAINTENANCE

--- SUPERVISOR MODE ---

BLANK TEST

.000

INTERNAL STANDARD

.000

EXTERNAL STANDARD

.000

BLANK TEST

.000

EXTERNAL STANDARD

.000

INTERNAL STANDARD

.000

EXTERNAL STANDARD

.000

BLANK TEST

.000

EXTERNAL STANDARD

.000

INTERNAL STANDARD

.000

EXTERNAL STANDARD

.000

BLANK TEST

.000

Operator Signature



Face This Side Down - This Edge In First

BAC DataMaster
Evidence Ticket

STATE OF MISSOURI
CHARGE POLICE DEPARTMENT

BAC DATAMASTER SERIAL NUMBER 201282
06/03/14
19:03

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY

PROGRAM (04-07-2009): OKAY

HEATERS
SAMPLE CHAMBER: 45C

FLOW DETECTOR: OKAY

PUMP
HIGH SPEED: OKAY

DETECTOR: OKAY

FILTERS: OKAY

QUARTZ STANDARD: OKAY

CALIBRATION: OKAY

PRINTER TEST
!"#\$%&'()*+,-./0123456789:;<=>?@AB CDEFG
HIJKLMNOPQRSTUVWXYZ[\]^_`abcde+ghijklmnop
qrsuvwxyz{|}~+-

Operator Signature 