MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
STATE PUBLIC HEALTH LABORATORY
BREATH ALCOHOL PROGRAM

DATAMASTER MAINTENANCE REPORT

RECEIVED REPORT #6
By Carol Day at 2:15 pm, Feb 06, 2014

Complete this report at the time of the regular monthly preventive maintenance check (not to exceed 36 days).
Complete this report whenever the instrument is serviced or repaired and whenever it is placed into service.
Retain the original and send a copy within 15 days to the Breath Alcohol Program, DHSS.

DATAMASTER SN	NAME OF AGENCY	DATE OF INSPECTION
201286	Mtn Grove Police	01/30/2014
LOCATION OF INSTRUMENT (STREET AND CITY)		TIME OF INSPECTION
102 E. STATE ST Mtn Grove.		0814

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

- [x] DIAGNOSTIC CHECK (PRINTOUT ATTACHED) — DATE AND TIME (from printout) 01/30/2014 0814
- [x] COMPUTER
- [x] DETECTOR
- [x] PROGRAM
- [x] FILTERS
- [x] HEATERS SAMPLE CHAMBER 48 °C
- [x] QUARTZ STANDARD
- [x] FLOW DETECTOR
- [x] CALIBRATION
- [x] PUMP HIGH SPEED
- [x] PRINTER
- [x] INDICATOR LIGHTS
- [x] SIMULATOR SOLUTION SUPPLIER GLETH LABS LOT # 12170 EXP. DATE 9/5/14
- [x] SIMULATOR TEMP (34°C ± 0.2°C) 34.0 °C SIMULATOR SN 094948 EXP. DATE 01/29/15
- [x] CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Mark the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

- [x] 0.100% STANDARD - MUST READ BETWEEN 0.095% AND 0.105% INCLUSIVE
- [ ] 0.080% STANDARD - MUST READ BETWEEN 0.076% AND 0.084% INCLUSIVE
- [ ] 0.040% STANDARD - MUST READ BETWEEN 0.038% AND 0.042% INCLUSIVE

TEST 1	TEST 2	TEST 3
.097	.097	.097

- [x] PERFORM R.F.I. TEST (PRINTOUT ATTACHED)

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:
(DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS	(0-.04)	(.05-.09)	(.10-.14)	(.15-.19)	OVER .19
			8	0	0

LIST ANY NEW PARTS AND DESCRIBE ANY ALTERATION OR MODIFICATION THAT WAS MADE TO RESTORE THE INSTRUMENT TO OPERATE SATISFACTORILY AND WITHIN ESTABLISHED LIMITS (USE OTHER SIDE IF NECESSARY).

INSPECTING OFFICER

SIGNATURE	PRINT FULL NAME
Danny Bledsoe	Danny Bledsoe
TYPE II PERMIT NUMBER/EXPIRATION DATE	TELEPHONE NUMBER
2300060 04/10/2015	417-926-5181

RETURN COMPLETED REPORT TO THE: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
2875 James Blvd.
Poplar Bluff, MO 63901

GUTH LABORATORIES, INC.
600 NORTH 67th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-564-5470

CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number **12170** of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on **September 11, 2012,** using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain **0.1215%** (w/vol) ethyl alcohol. The expiration date for this lot number is **September 5, 2014** at 11:59 PM.

When used in a calibrated Simulator, operating at 34°C +/- .2°C, this solution will give a breath alcohol analysis instrument reading of **0.100 g/210L +/- 3%.**

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President
GUTH LABORATORIES, INC.

NIST Traceability:
Testing was conducted using Cerilliant Reference Standard lot number FNI22211-02 whose values are traceable to NIST.
All balances are calibrated annually by an outside agency using NIST traceable weights.
Calibration verification is done prior to each use utilizing NIST traceable weights.

Face This Side Down - This Edge In First

BAC DataMaster
Evidence Ticket

STATE OF MISSOURI
MOUNTAIN GROVE POLICE DEPARTMENT

BAC DATAMASTER SERIAL NUMBER 201286
01/30/14

ARREST TIME: 08:00
SUBJECT NAME:
RFI/TEST
DOB: 01/01/01 SEX: M
STATE/D.L.: MO/000000000
ARRESTING OFFICER:
BLEDSOE/DANNY
OFFICER I.D.: 1201
TESTING OFFICER:
BLEDSOE
OFFICER I.D.: 1201
PERMIT NUMBER: 230060
EXPIRATION DATE: 04/10/15
MISCELLANEOUS DATA:

--- BREATH ANALYSIS ---

BLANK TEST	.000	08:34
INTERNAL STANDARD	VERIFIED	08:34
SUBJECT SAMPLE	.000	08:34
BLANK TEST	.000	08:35

Operator Signature Wally Bledsoe

Face This Side Down - This Edge In First

BAC DataMaster
Evidence Ticket

STATE OF MISSOURI
MOUNTAIN GROVE POLICE DEPARTMENT

BAC DATAMASTER SERIAL NUMBER 201286
01/30/14

ARREST TIME: 08:00
SUBJECT NAME:
RFI/TEST
DOB: 01/01/01 SEX: M
STATE/D.L.: MO/000000000
ARRESTING OFFICER:
BLEDSOE/DANNY
OFFICER I.D.: 1201
TESTING OFFICER:
BLEDSOE
OFFICER I.D.: 1201
PERMIT NUMBER: 230060
EXPIRATION DATE: 04/10/15
MISCELLANEOUS DATA:

--- BREATH ANALYSIS ---

BLANK TEST	.000	08:32
INTERNAL STANDARD	VERIFIED	08:32
RADIO INTERFERENCE		

Operator Signature Wally Bledsoe

Face This Side Down - This Edge In First

BAC DataMaster
Evidence Ticket

STATE OF MISSOURI
MOUNTAIN GROVE POLICE DEPARTMENT

BAC DATAMASTER SERIAL NUMBER 201286
01/30/14

TESTING OFFICER:
BLEDSOE
OFFICER I.D.: 1201
PERMIT NUMBER: 230060
EXPIRATION DATE: 04/10/15
MISCELLANEOUS DATA:

--- SUPERVISOR MODE ---

BLANK TEST	.000	08:26
INTERNAL STANDARD	VERIFIED	08:26
EXTERNAL STANDARD	.097	08:27
BLANK TEST	.000	08:27
EXTERNAL STANDARD	.097	08:28
BLANK TEST	.000	08:28
EXTERNAL STANDARD	.097	08:29
BLANK TEST	.000	08:29

N = 3
STM. = .1
AVG. = .097

Operator Signature Wally Bledsoe

Face This Side Down - This Edge In First

BAC DataMaster
Evidence Ticket

STATE OF MISSOURI
MOUNTAIN GROVE POLICE DEPARTMENT

BAC DATAMASTER SERIAL NUMBER 201286
01/30/14
08:14

--- DIAGNOSTIC CHECK ---

COMPUTER:	OKAY
PROGRAM (04-07-2009):	OKAY
HEATERS SAMPLE CHAMBER:	48c
FLOW DETECTOR:	OKAY
PUMP HIGH SPEED:	OKAY
DETECTOR:	OKAY
FILTERS:	OKAY
QUARTZ STANDARD:	OKAY
CALIBRATION:	OKAY

PRINTER TEST
!"#$%&'()*+,-./0123456789:;<=>?@ABCDEFG
HIJKLMNOPQRSTUVWXYZ[\]^_`abcdefghijklmno
pqrstuvwxyz{|}~

Operator Signature Wally Bledsoe

2

STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

PERMIT
TYPE II

DANIEL R BLEDSOE JR

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

DATAMASTER

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 04/10/2013

NUMBER 230060

EXPIRES 04/10/2015

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

Acting Director

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MO 580-0771 (5-10)

LAB-4 (R5-10)