



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
 STATE PUBLIC HEALTH LABORATORY
 BREATH ALCOHOL PROGRAM
DATAMASTER MAINTENANCE REPORT

received 7/1/14-cd

REVIEWED REPORT #0
 By Carol Day at 9:43 am, Jul 14, 2014

Complete this report at the time of the regular monthly preventive maintenance check (not to exceed 35 days).
 Complete this report whenever the instrument is serviced or repaired and whenever it is placed into service.
 Retain the original and send a copy within 15 days to the Breath Alcohol Program, DHSS.

DATAMASTER SN 201288	NAME OF AGENCY CAMPBELL PD	DATE OF INSPECTION 06/27/2014
LOCATION OF INSTRUMENT (STREET AND CITY) 2875 JAMES BLVD., POPLAR BLUFF, MO 63901		TIME OF INSPECTION 3:55 pm

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

<input checked="" type="checkbox"/> DIAGNOSTIC CHECK (PRINTOUT ATTACHED)	DATE AND TIME (from printout) 15:55
<input checked="" type="checkbox"/> COMPUTER	<input checked="" type="checkbox"/> DETECTOR
<input checked="" type="checkbox"/> PROGRAM	<input checked="" type="checkbox"/> FILTERS
<input checked="" type="checkbox"/> HEATERS SAMPLE CHAMBER 49 °C	<input checked="" type="checkbox"/> QUARTZ STANDARD
<input checked="" type="checkbox"/> FLOW DETECTOR	<input checked="" type="checkbox"/> CALIBRATION
<input checked="" type="checkbox"/> PUMP HIGH SPEED	<input checked="" type="checkbox"/> PRINTER

INDICATOR LIGHTS

SIMULATOR SOLUTION SUPPLIER GUTH LOT # 14110 EXP. DATE 03/11/2015

SIMULATOR TEMP (34°C ± 0.2°C) 34.0 °C SIMULATOR SN DR5786 EXP. DATE 05/15/2015

CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Mark the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

- 0.100% STANDARD - MUST READ BETWEEN 0.095% AND 0.105% INCLUSIVE
- 0.080% STANDARD - MUST READ BETWEEN 0.076% AND 0.084% INCLUSIVE
- 0.040% STANDARD - MUST READ BETWEEN 0.038% AND 0.042% INCLUSIVE

TEST 1 → 0.098	TEST 2 → 0.099	TEST 3 → 0.098
----------------	----------------	----------------

PERFORM R.F.I. TEST (PRINTOUT ATTACHED)

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:
 (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS	0	(0-.04)	0	(.05-.09)	0	(.10-.14)	0	(.15-.19)	0	OVER .19	0
----------	---	---------	---	-----------	---	-----------	---	-----------	---	----------	---

LIST ANY NEW PARTS AND DESCRIBE ANY ALTERATION OR MODIFICATION THAT WAS MADE TO RESTORE THE INSTRUMENT TO OPERATE SATISFACTORILY AND WITHIN ESTABLISHED LIMITS (USE OTHER SIDE IF NECESSARY).

REPLACED CHIP. RECALIBRATED. INSTRUMENT AND SIMULATOR OPERATING WITHIN DHSS SPECIFICATIONS.

INSPECTING OFFICER

SIGNATURE 	PRINT FULL NAME BRIAN M. LUTMER
---------------	------------------------------------

TYPE II PERMIT NUMBER/EXPIRATION DATE 230031/02-10-2015	TELEPHONE NUMBER (573) 840-9140
--	------------------------------------

RETURN COMPLETED REPORT TO THE: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
 2875 James Blvd.
 Poplar Bluff, MO 63901



GUTH LABORATORIES, INC.

690 NORTH 67th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-564-5470

CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number 14110 of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on May 5, 2014, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain 0.1206% (w/vol) ethyl alcohol. The expiration date for this lot number is May 1, 2016 at 11:59 PM.

When used in a calibrated Simulator, operating at 34°C +/- .2°C, this solution will give a breath alcohol analysis instrument reading of 0.100 g/210L +/- 3%.


The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President
GUTH LABORATORIES, INC.


NIST Traceability:

Testing was conducted using Cerilliant Reference Standard lot number FN122211-02 whose values are traceable to NIST.

All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM



**PERMIT
TYPE II**


BRIAN M LUTMER

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

**DATAMASTER, DATAMASTER DMT, INTOXILYZER 5000,
INTOXILYZER 8000, ALCO-SENSOR IV WITH PRINTER, EC/IR II**

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 02/10/2013
NUMBER 230031
EXPIRES 02/10/2015


DIRECTOR OF STATE PUBLIC HEALTH LABORATORY
Acting Director

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES
LAB-4 (R6-10)

SIDE DOWN - THIS EDGE IN FIRST

BAC DataMaster
Evidence Ticket

1. NAME OF SUBJECT
2. DATE OF TEST
3. TIME OF TEST
4. LOCATION
5. TESTER
6. INSTRUMENT
7. MODEL
8. SERIAL NO.
9. LOT NO.
10. VOLUME
11. ALCOHOL
12. BAC
13. COMMENTS

14. SIGNATURE OF TESTER
15. DATE OF REPORT

B. J.

ORDER FROM N.P.A.S.
FIELD OH 44901 PHONE 1-800-800-8143 (NPAS)

FACE THIS SIDE DOWN - THIS EDGE IN FIRST

BAC DataMaster
Evidence Ticket

1. NAME OF SUBJECT
2. DATE OF TEST
3. TIME OF TEST
4. LOCATION
5. TESTER
6. INSTRUMENT
7. MODEL
8. SERIAL NO.
9. LOT NO.
10. VOLUME
11. ALCOHOL
12. BAC
13. COMMENTS

14. SIGNATURE OF TESTER
15. DATE OF REPORT

OPERATOR SIGNATURE

B. J.

Card Stock No.
60021

REORDER ALL SUPPLIES FROM N.P.A.S.
P.O. BOX 1435, MANSFIELD OH 44901 PHONE 1-800-800-8143 (NPAS)

FACE THIS SIDE DOWN - THIS EDGE IN FIRST

BAC DataMaster
Evidence Ticket

STATE OF MISSISSIPPI
CORRECTIONAL INSTITUTIONS DEPARTMENT

BAC DATA MASTER SERVICE NUMBER 25912000
06/27/19

INSTITUTION NAME
ATLANTA/ATLANTA

INSTITUTION NUMBER 2590001

EXPIRATION DATE 06/30/19

STANDARD NUMBER
001-14110

--- SUPERVISOR NAME ---


CLASS TEST .0000 16180
CORRECTIONAL INSTITUTION VERIFIED 16180

CLASS
CORRECTIONAL INSTITUTION .0000 16180
CLASS TEST .0000 16180

CLASS
CORRECTIONAL INSTITUTION .0000 16180
CORRECTIONAL INSTITUTION .0000 16180

CLASS
CORRECTIONAL INSTITUTION .0000 16180
CORRECTIONAL INSTITUTION .0000 16180

OPERATOR SIGNATURE



Card Stock No.
60021

REORDER ALL SUPPLIES FROM N.P.A.S.
P.O. BOX 1435, MANSFIELD OH 44901 PHONE 1-800-800-8143 (NPAS)