



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
 STATE PUBLIC HEALTH LABORATORY
 BREATH ALCOHOL PROGRAM
DATAMASTER MAINTENANCE REPORT

RECEIVED
 By Carol Day at 1:43 pm, Sep 08, 2014

REPORT #6

Complete this report at the time of the regular monthly preventive maintenance check (not to exceed 35 days).
 Complete this report whenever the instrument is serviced or repaired and whenever it is placed into service.
 Retain the original and send a copy within 15 days to the Breath Alcohol Program, DHSS.

DATAMASTER SN 201288	NAME OF AGENCY Campbell Police Dept.	DATE OF INSPECTION 09/01/2014
LOCATION OF INSTRUMENT (STREET AND CITY) 204 W. Grand Campbell, MO 63933		TIME OF INSPECTION 11:00 am

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

<input checked="" type="checkbox"/> DIAGNOSTIC CHECK (PRINTOUT ATTACHED)	DATE AND TIME (from printout) 09/01/14
<input checked="" type="checkbox"/> COMPUTER	<input checked="" type="checkbox"/> DETECTOR
<input checked="" type="checkbox"/> PROGRAM	<input checked="" type="checkbox"/> FILTERS
<input checked="" type="checkbox"/> HEATERS SAMPLE CHAMBER 49°C	<input checked="" type="checkbox"/> QUARTZ STANDARD
<input checked="" type="checkbox"/> FLOW DETECTOR	<input checked="" type="checkbox"/> CALIBRATION
<input checked="" type="checkbox"/> PUMP HIGH SPEED	<input checked="" type="checkbox"/> PRINTER

INDICATOR LIGHTS

SIMULATOR SOLUTION SUPPLIER RepCo LOT # 13002 EXP. DATE 06/19/2015

SIMULATOR TEMP (34°C ± 0.2°C) 34 °C SIMULATOR SN SD1743 EXP. DATE 01/16/2015

CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Mark the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

0.100% STANDARD - MUST READ BETWEEN 0.095% AND 0.105% INCLUSIVE
 0.080% STANDARD - MUST READ BETWEEN 0.076% AND 0.084% INCLUSIVE
 0.040% STANDARD - MUST READ BETWEEN 0.038% AND 0.042% INCLUSIVE

TEST 1 .096%	TEST 2 .097%	TEST 3 .096%
--------------	--------------	--------------

PERFORM R.F.I. TEST (PRINTOUT ATTACHED)

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:
 (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS	0	(0-.04)	5	(.05-.09)	0	(.10-.14)	1	(.15-.19)	0	OVER .19	0
----------	---	---------	---	-----------	---	-----------	---	-----------	---	----------	---

LIST ANY NEW PARTS AND DESCRIBE ANY ALTERATION OR MODIFICATION THAT WAS MADE TO RESTORE THE INSTRUMENT TO OPERATE SATISFACTORILY AND WITHIN ESTABLISHED LIMITS (USE OTHER SIDE IF NECESSARY).
 Instrument is within D.O.H. specifications. Simulator solution RepCp lot#13002 exp. date June 19th 2015. Vapor alcohol value .100+/-3%

INSPECTING OFFICER

SIGNATURE 	PRINT FULL NAME Weldon Wallace
TYPE II PERMIT NUMBER/EXPIRATION DATE 220240 09/07/2014	TELEPHONE NUMBER (573) 246-2511

RETURN COMPLETED REPORT TO THE: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
 2875 James Blvd.
 Poplar Bluff, MO 63901

CERTIFICATE OF ANALYSIS

MANUFACTURER AND SUPPLIER: RepCo Marketing, Inc.
 LOT NUMBER: 13002
 EXPIRATION DATE: June 19, 2015 at 11:59 p.m.

RepCo Marketing, Inc. certifies the following:

RepCo Marketing, Inc. manufactured, tested and supplied Lot Number 13002 of Alcohol Certified Solution for simulators. Random samples of said lot number were analyzed by an independent laboratory utilizing a gas chromatograph and found to contain .1217 gms/dl +/- .003 gms/dl wt/vol ethanol (95% Confidence).

The alcohol and distilled water used in the solution were found to be free of any interfering substance.

This solution will produce a vapor alcohol value of .100 +/- 3% gms/210L Breath when heated to 34 Degrees Celsius +/- 0.2 Degrees Celsius in a simulator (95% Confidence).

The date of manufacture for this lot number is June 20, 2013
 The expiration date for this lot number is June 19, 2015 at 11:59 p.m.

This document is a true representation of the original Certificate of Analysis.



Cecil B. Garner, President
 RepCo Marketing, Inc.

State of Missouri
DEPARTMENT OF HEALTH



P E R M I T
TYPE II



WELDON WALLACE

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field repairs, and operate the following breath analyzer(s):

DATAMASTER

for the determination of the alcoholic content of blood from a sample of expired (alveolar) air. Issued under the provisions of sections 577.020 through 577.041, RSMo 1986.

Date 09/07/2012

Number 220240

Expires 09/07/2014

Director of State Public Health Laboratory

Director, Department of Health

FACE THIS SIDE DOWN - THIS EDGE IN FIRST

BAC DataMaster
Evidence Ticket

STATE OF MISSOURI
CAMPBELL POLICE DEPARTMENT

BAC DATAMASTER SERIAL NUMBER 201208
09/01/14

TESTING OFFICER:
MULLACE/MELDICH/L
OFFICER I.D.# 23
PERMIT NUMBER: 220249
EXPIRATION DATE: 09/07/14
MISCELLANEOUS DATA:
MAINTENANCE CHECK

--- SUPERVISOR MODE ---

BLANK TEST	.000	10:40
INTERNAL STANDARD	VERIFIED	10:40
EXTERNAL STANDARD	.096	10:41
BLANK TEST	.000	10:41
EXTERNAL STANDARD	.097	10:42
BLANK TEST	.000	10:42
EXTERNAL STANDARD	.096	10:43
BLANK TEST	.000	10:43

N = 3
STL = .1
HV% = .8963

OPERATOR SIGNATURE

Card Stock No. 60021
REORDER ALL SUPPLIES FROM N.P.A.S.
P.O. BOX 1435, MANSFIELD, OH 44901

FACE THIS SIDE DOWN - THIS EDGE IN FIRST

BAC DataMaster
Evidence Ticket

STATE OF MISSOURI
CAMPBELL POLICE DEPARTMENT

BAC DATAMASTER SERIAL NUMBER 201208
09/01/14
10:28

--- DIAGNOSTIC CHECK ---

COMPUTER:	OKAY
PROGRAM (04-87-2009):	OKAY
VENTER:	
SAMPLE CHAMBERS:	49c
FLOW DETECTOR:	OKAY
PUMP	
RPM SPEED:	OKAY
DETECTOR:	OKAY
FILTERS:	OKAY
QUARTZ STANDARD:	OKAY
CALIBRATION:	OKAY

PRINTER TEST

!@#\$%&'()*+,-./0123456789:;<=>?@ABCDEF6
HIJKL MNOPQRSTU VWXYZ[\]^_`abcd efghijklmno
pqrstuvwxy z{|}~*

OPERATOR SIGNATURE

Card Stock No. 60021
REORDER ALL SUPPLIES FROM N.P.A.S.
P.O. BOX 1435, MANSFIELD, OH 44901

FACE THIS SIDE DOWN - THIS EDGE IN FIRST

BAC DataMaster
Evidence Ticket

STATE OF MISSOURI
CAMPBELL POLICE DEPARTMENT


BAC DATAMASTER SERIAL NUMBER 281298
89-01-14

ARREST TIME: 10:00
SUBJECT NAME:
JOE/JON/H
DOB: 10/10/70 SEX: M
STATE/D.L.: MO/123456
ARRESTING OFFICER:
WALLACE/WELDON/L
OFFICER I.D.: 28
TESTING OFFICER:
WALLACE/WELDON/L
OFFICER I.D.: 75
PERMIT NUMBER: 228240
EXPIRATION DATE: 89-07-14
MISCELLANEOUS DATA:
MAINTENANCE CHECK

--- BREATH ANALYSIS ---

BLANK TEST	.000	10:47
INTERNAL STANDARD	VERIFIED	10:47
RADIO INTERFERENCE		

OPERATOR SIGNATURE



Card Stock No.
60021

REORDER ALL SUPPLIES FROM N.P.A.S.
P.O. BOX 1435, MANSFIELD, OH 44901