



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
 STATE PUBLIC HEALTH LABORATORY
 BREATH ALCOHOL PROGRAM
DATAMASTER MAINTENANCE REPORT

RECEIVED
 By Carol Day at 1:09 pm, Feb 07, 2014
 REPORT #6

Complete this report at the time of the regular monthly preventive maintenance check (not to exceed 35 days).
 Complete this report whenever the instrument is serviced or repaired and whenever it is placed into service.
 Retain the original and send a copy within 15 days to the Breath Alcohol Program, DHSS.

DATAMASTER SN 201291	NAME OF AGENCY Missouri State Highway Patrol	DATE OF INSPECTION 01/25/2014
LOCATION OF INSTRUMENT (STREET AND CITY) Cabool Police Department, Cabool, Missouri		TIME OF INSPECTION 12:49 pm

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

<input checked="" type="checkbox"/> DIAGNOSTIC CHECK (PRINTOUT ATTACHED)	DATE AND TIME (from printout) 01/25/2014 12:49pm
<input checked="" type="checkbox"/> COMPUTER	<input checked="" type="checkbox"/> DETECTOR
<input checked="" type="checkbox"/> PROGRAM	<input checked="" type="checkbox"/> FILTERS
<input checked="" type="checkbox"/> HEATERS SAMPLE CHAMBER _____ 48 °C	<input checked="" type="checkbox"/> QUARTZ STANDARD
<input checked="" type="checkbox"/> FLOW DETECTOR	<input checked="" type="checkbox"/> CALIBRATION
<input checked="" type="checkbox"/> PUMP HIGH SPEED	<input checked="" type="checkbox"/> PRINTER

<input checked="" type="checkbox"/> INDICATOR LIGHTS
<input checked="" type="checkbox"/> SIMULATOR SOLUTION SUPPLIER <u>Guth Laboratories, Inc.</u> LOT # <u>13001</u> EXP. DATE <u>03/07/2015</u>
<input checked="" type="checkbox"/> SIMULATOR TEMP (34°C ± 0.2°C) <u>33.99</u> °C SIMULATOR SN <u>1324</u> EXP. DATE <u>01/07/2015</u>

CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Mark the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

0.100% STANDARD - MUST READ BETWEEN 0.095% AND 0.105% INCLUSIVE
 0.080% STANDARD - MUST READ BETWEEN 0.076% AND 0.084% INCLUSIVE
 0.040% STANDARD - MUST READ BETWEEN 0.038% AND 0.042% INCLUSIVE

TEST 1 .101%	TEST 2 .102%	TEST 3 .103%
---------------	---------------	---------------

PERFORM R.F.I. TEST (PRINTOUT ATTACHED)

**INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:
 (DO NOT INCLUDE SELF-ADMINISTERED TESTS)**

REFUSALS	0	(0-.04)	0	(.05-.09)	1	(.10-.14)	2	(.15-.19)	0	OVER .19	0
----------	---	---------	---	-----------	---	-----------	---	-----------	---	----------	---

LIST ANY NEW PARTS AND DESCRIBE ANY ALTERATION OR MODIFICATION THAT WAS MADE TO RESTORE THE INSTRUMENT TO OPERATE SATISFACTORILY AND WITHIN ESTABLISHED LIMITS (USE OTHER SIDE IF NECESSARY).

This instrument is operating satisfactorily within the guidelines set by the Missouri Department of Health, Guth Laboratories Inc., 100% solution, Lot #13001, Bottle # 0416, expiration date 03/07/2015.

INSPECTING OFFICER

SIGNATURE 	PRINT FULL NAME Corporal Dale B. Pounds
---------------	--

TYPE II PERMIT NUMBER/EXPIRATION DATE 220325 09/28/2014	TELEPHONE NUMBER (417) 469-3121
--	------------------------------------

RETURN COMPLETED REPORT TO THE: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
 2875 James Blvd.
 Poplar Bluff, MO 63901

CERTIFICATE OF ANALYSIS

MANUFACTURER AND SUPPLIER: RepCo Marketing, Inc.
LOT NUMBER: 13001
EXPIRATION DATE: March 7, 2015 at 11:59 p.m.

RepCo Marketing, Inc. certifies the following:

RepCo Marketing, Inc. manufactured, tested and supplied Lot Number 13001 of Alcohol Certified Solution for simulators. Random samples of said lot number were analyzed by an independent laboratory utilizing a gas chromatograph and found to contain .1215 gms/dl +/- .003 gms/dl wt/vol ethanol (95% Confidence).

The alcohol and distilled water used in the solution were found to be free of any interfering substance.

This solution will produce a vapor alcohol value of .100 +/- 3% gms/210L Breath when heated to 34 Degrees Celsius +/- 0.2 Degrees Celsius in a simulator (95% Confidence).

The date of manufacture for this lot number is March 8, 2013
The expiration date for this lot number is March 7, 2015 at 11:59 p.m.

This document is a true representation of the original Certificate of Analysis.



Cecil B. Garner, President
RepCo Marketing, Inc.

State of Missouri
DEPARTMENT OF HEALTH



P E R M I T
T Y P E II



DALE B POUNDS

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field repairs, and operate the following breath analyzer(s):

DATAMASTER

for the determination of the alcoholic content of blood from a sample of expired (alveolar) air. Issued under the provisions of sections 577.020 through 577.041, RSMo 1986.

Date 09/28/2012

Number 220325

Expires 09/28/2014

MO 580-0771 (7-88)

Director of State Public Health Laboratory

Director, Department of Health

Lab. 4 (R7-88)

**BAC DataMaster
Evidence Ticket**

STATE OF MISSOURI
CABOOL POLICE DEPARTMENT

BAC DATAMASTER SERIAL NUMBER 201291
01/25/14
12:49

DIAGNOSTIC CHECK

0000 OKAY
0001 OKAY
0002 OKAY
0003 OKAY
0004 OKAY
0005 OKAY
0006 OKAY
0007 OKAY
0008 OKAY
0009 OKAY
0010 OKAY

PRINTER TEST
0000 OKAY
0001 OKAY
0002 OKAY
0003 OKAY
0004 OKAY
0005 OKAY
0006 OKAY
0007 OKAY
0008 OKAY
0009 OKAY
0010 OKAY

Signature Cpl D.B. Pounds

**BAC DataMaster
Evidence Ticket**

STATE OF MISSOURI
CABOOL POLICE DEPARTMENT

BAC DATAMASTER SERIAL NUMBER 201291
01/25/14

ARREST TIME: 12:15
SUBJECT NAME:
NELSON/TERRY/L
DOB: 02/02/62 SEX: M
STATE/D.L.: MO/183456789
ARRESTING OFFICER:
POUNDS/D/B

OFFICER I.D.: 984
TESTING OFFICER:
SAME
OFFICER I.D.: SAME
PERMIT NUMBER: 226325
EXPIRATION DATE: 09/28/14
MISCELLANEOUS DATA:

----- BREATH ANALYSIS -----

BLANK TEST .000
INTERNAL STANDARD VERIFIED
RADIO INTERFERENCE

12:15
12:15
12:15
12:15
12:15
12:15
12:15
12:15

N = 3
SIM. = .1
AVG. = .102

Operator Signature Cpl D.B. Pounds

Operator Signature Cpl D.B. Pounds