



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 STATE PUBLIC HEALTH LABORATORY  
 BREATH ALCOHOL PROGRAM  
**DATAMASTER MAINTENANCE REPORT**

RECEIVED 3/14/14-CD

**REVIEWED**  
 By Carol Day at 2:18 pm, Mar 31, 2014 REPORT #6

Complete this report at the time of the regular monthly preventive maintenance check (not to exceed 35 days).  
 Complete this report whenever the instrument is serviced or repaired and whenever it is placed into service.  
 Retain the original and send a copy within 15 days to the Breath Alcohol Program, DHSS.

DATAMASTER SN <b>201292</b>	NAME OF AGENCY <b>Owensville Police Department</b>	DATE OF INSPECTION <b>03-04-2014</b>
LOCATION OF INSTRUMENT (STREET AND CITY) <b>109 North Second Street Owensville</b>		TIME OF INSPECTION <b>1841</b>

**CHECKLIST:** Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

<input checked="" type="checkbox"/> DIAGNOSTIC CHECK (PRINTOUT ATTACHED)	DATE AND TIME (from printout) <b>03/04/14 1841</b>
<input checked="" type="checkbox"/> COMPUTER	<input checked="" type="checkbox"/> DETECTOR
<input checked="" type="checkbox"/> PROGRAM	<input checked="" type="checkbox"/> FILTERS
<input checked="" type="checkbox"/> HEATERS SAMPLE CHAMBER <b>49</b> °C	<input checked="" type="checkbox"/> QUARTZ STANDARD
<input checked="" type="checkbox"/> FLOW DETECTOR	<input checked="" type="checkbox"/> CALIBRATION
<input checked="" type="checkbox"/> PUMP HIGH SPEED	<input checked="" type="checkbox"/> PRINTER

INDICATOR LIGHTS

SIMULATOR SOLUTION SUPPLIER **RepCo Marketing, Inc** LOT # **13002** EXP. DATE **06-19-2015**

SIMULATOR TEMP (34°C ± 0.2°C) **34.0°C** °C SIMULATOR SN **502300** EXP. DATE **01-07-2015**

CALIBRATION CHECK - (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Mark the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

0.100% STANDARD - MUST READ BETWEEN 0.095% AND 0.105% INCLUSIVE  
 0.080% STANDARD - MUST READ BETWEEN 0.076% AND 0.084% INCLUSIVE  
 0.040% STANDARD - MUST READ BETWEEN 0.038% AND 0.042% INCLUSIVE

TEST 1 • <b>.101</b>	TEST 2 • <b>.102</b>	TEST 3 • <b>.102</b>
----------------------	----------------------	----------------------

PERFORM R.F.I. TEST (PRINTOUT ATTACHED)

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:  
 (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS <b>1</b>	(0-.04) <b>0</b>	(.05-.09) <b>1</b>	(.10-.14) <b>1</b>	(.15-.19) <b>0</b>	OVER .19 <b>0</b>
-------------------	------------------	--------------------	--------------------	--------------------	-------------------

LIST ANY NEW PARTS AND DESCRIBE ANY ALTERATION OR MODIFICATION THAT WAS MADE TO RESTORE THE INSTRUMENT TO OPERATE SATISFACTORILY AND WITHIN ESTABLISHED LIMITS (USE OTHER SIDE IF NECESSARY).

*Instrument operating properly within Department of Health and Senior Services guidelines and standards.*

INSPECTING OFFICER	
SIGNATURE <i>J. Griffith</i>	PRINT FULL NAME <b>JONATHAN SCOTT GRIFFITH</b>
TYPE II PERM NUMBER/EXPIRATION DATE <b>220077 / 03-23-2014</b>	TELEPHONE NUMBER <b>573-437-2195 or 573-437-7770</b>

RETURN COMPLETED REPORT TO THE:  
 Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office  
 2875 James Blvd.  
 Poplar Bluff, MO 63901

**CERTIFICATE OF ANALYSIS**

**MANUFACTURER AND SUPPLIER: RepCo Marketing, Inc.**

**LOT NUMBER: 13002**

**EXPIRATION DATE: June 19, 2015 at 11:59 p.m.**

RepCo Marketing, Inc. certifies the following:


RepCo Marketing, Inc. manufactured, tested and supplied Lot Number 13002 of Alcohol Certified Solution for simulators. Random samples of said lot number were analyzed by an independent laboratory utilizing a gas chromatograph and found to contain .1217 gms/dl +/- .003 gms/dl wt/vol ethanol (95% Confidence).

The alcohol and distilled water used in the solution were found to be free of any interfering substance.

This solution will produce a vapor alcohol value of .100 +/- 3% gms/210L Breath when heated to 34 Degrees Celsius +/- 0.2 Degrees Celsius in a simulator (95% Confidence).

The date of manufacture for this lot number is June 20, 2013  
The expiration date for this lot number is June 19, 2015 at 11:59 p.m.

This document is a true representation of the original Certificate of Analysis.



Cecil B. Garner, President  
RepCo Marketing, Inc.

Face This Side Down - This Edge In First

# BAC DataMaster Evidence Ticket

STATE OF MISSOURI  
OWENSVILLE POLICE DEPARTMENT

BAC DATAMASTER SERIAL NUMBER 201292  
03/04/14  
16:41

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY

PROGRAM (04-07-2009): OKAY

HEATERS

SAMPLE CHAMBER: 49c

FLOW DETECTOR: OKAY

PUMP

HIGH SPEED: OKAY

DETECTOR: OKAY

FILTERS: OKAY

QUARTZ STANDARD: OKAY

CALIBRATION: OKAY

PRINTER TEST

!"#\$%&'()\*+,-./0123456789:;<=>?@ABCDEF  
HIJKLMNOPQRS:TUVWXYZ[\]^\_`abcdefg|hijklmno  
pqrstuvwxyz{|}~"

Operator Signature

2208-02

Face This Side Down - This Edge In First

# BAC DataMaster Evidence Ticket

STATE OF MISSOURI  
OWENSVILLE POLICE DEPARTMENT

BAC DATAMASTER SERIAL NUMBER 201292  
03/04/14

TESTING OFFICER:

GRIFFITH/JONATHAN/SCOTT  
OFFICER I.D.: 104  
PERMIT NUMBER: 220077  
EXPIRATION DATE: 03/23/14  
MISCELLANEOUS DATA:

--- SUPERVISOR MODE ---

BLANK TEST	.000	18:44
INTERNAL STANDARD	VERIFIED	18:44
EXTERNAL STANDARD	.101	18:44
BLANK TEST	.000	18:45
EXTERNAL STANDARD	.102	18:46
BLANK TEST	.000	18:46
EXTERNAL STANDARD	.102	18:47
BLANK TEST	.000	18:47

N = 3  
SIM. = .1  
AVG. = .1016

Operator Signature

2208-02

Face This Side Down - This Edge In First

**BAC DataMaster  
Evidence Ticket**

STATE OF MISSOURI  
OWENSVILLE POLICE DEPARTMENT

BAC DATA MASTER SERIAL NUMBER 201292  
03/04/14

ARREST TIME: 06:00  
SUBJECT NAME:  
RFI/TEST  
DOB: 09/09/09 SEX: M  
STATE/D.L.: NA/NA  
ARRESTING OFFICER:  
NA  
OFFICER I.D.: NA  
TESTING OFFICER:  
GRIFFITH/JONATHAN/SCOTT  
OFFICER I.D.: 104  
PERMIT NUMBER: 220077  
EXPIRATION DATE: 03/23/14  
MISCELLANEOUS DATA:  
RFI TEST  
RFI TEST

--- BREATH ANALYSIS ---

BLANK TEST	.000	18:51
INTERNAL STANDARD	VERIFIED	18:51
RADIO INTERFERENCE		

Operator Signature

*[Handwritten Signature]* 104

2208-02