



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
STATE PUBLIC HEALTH LABORATORY
BREATH ALCOHOL PROGRAM

DATAMASTER MAINTENANCE REPORT

received 8/14/14-cd

REPORT #6

Complete this report at the time of the regular monthly preventive maintenance check (not to exceed ^{25 days})
Complete this report whenever the instrument is serviced or repaired and whenever it is placed into
Retain the original and send a copy within 15 days to the Breath Alcohol Program, DHSS.

REVIEWED
By Carol Day at 12:37 pm, Aug 28, 2014

DATAMASTER SN 201292	NAME OF AGENCY OWENSVILLE POLICE DEPARTMENT	DATE OF INSPECTION 08-09-2014
--------------------------------	---	---

LOCATION OF INSTRUMENT (STREET AND CITY) 109 NORTH SECOND STREET OWENSVILLE	TIME OF INSPECTION 0035
---	-----------------------------------

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

<input checked="" type="checkbox"/> DIAGNOSTIC CHECK (PRINTOUT ATTACHED)	DATE AND TIME (from printout) 08/09/14 00:35
<input checked="" type="checkbox"/> COMPUTER	<input checked="" type="checkbox"/> DETECTOR
<input checked="" type="checkbox"/> PROGRAM	<input checked="" type="checkbox"/> FILTERS
<input checked="" type="checkbox"/> HEATERS SAMPLE CHAMBER 49 °C	<input checked="" type="checkbox"/> QUARTZ STANDARD
<input checked="" type="checkbox"/> FLOW DETECTOR	<input checked="" type="checkbox"/> CALIBRATION
<input checked="" type="checkbox"/> PUMP HIGH SPEED	<input checked="" type="checkbox"/> PRINTER

INDICATOR LIGHTS

SIMULATOR SOLUTION SUPPLIER **Replco Marketing, Inc.** LOT # **13002** EXP. DATE **06-19-2015**

SIMULATOR TEMP (34°C ± 0.2°C) **34.0°C** °C SIMULATOR SN **SD 2300** EXP. DATE **01-07-2015**

CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Mark the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

- 0.100% STANDARD - MUST READ BETWEEN 0.095% AND 0.105% INCLUSIVE
- 0.080% STANDARD - MUST READ BETWEEN 0.076% AND 0.084% INCLUSIVE
- 0.040% STANDARD - MUST READ BETWEEN 0.038% AND 0.042% INCLUSIVE

TEST 1 • .102	TEST 2 • .102	TEST 3 • .103
----------------------	----------------------	----------------------

PERFORM R.F.I. TEST (PRINTOUT ATTACHED)

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:
(DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS 0	(.0-.04) 0	(.05-.09) 0	(.10-.14) 2	(.15-.19) 0	OVER .19 0
-------------------	-------------------	--------------------	--------------------	--------------------	-------------------

LIST ANY NEW PARTS AND DESCRIBE ANY ALTERATION OR MODIFICATION THAT WAS MADE TO RESTORE THE INSTRUMENT TO OPERATE SATISFACTORILY AND WITHIN ESTABLISHED LIMITS (USE OTHER SIDE IF NECESSARY).

INSTRUMENT OPERATING PROPERLY WITHIN DEPARTMENT OF HEALTH AND SENIOR SERVICES guidelines and standards.

INSPECTING OFFICER

SIGNATURE <i>[Signature]</i>	PRINT FULL NAME JONATHAN SCOTT GRIFFITH
TYPE II PERMIT NUMBER/EXPIRATION DATE 240130 / 04-03-2016	TELEPHONE NUMBER 573-437-2195

RETURN COMPLETED REPORT TO THE: **Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
2875 James Blvd.
Poplar Bluff, MO 63901**

CERTIFICATE OF ANALYSIS

MANUFACTURER AND SUPPLIER: RepCo Marketing, Inc.
LOT NUMBER: 13002
EXPIRATION DATE: June 19, 2015 at 11:59 p.m.

RepCo Marketing, Inc. certifies the following:

RepCo Marketing, Inc. manufactured, tested and supplied Lot Number 13002 of Alcohol Certified Solution for simulators. Random samples of said lot number were analyzed by an independent laboratory utilizing a gas chromatograph and found to contain .1217 gms/dl +/- .003 gms/dl wt/vol ethanol (95% Confidence).

The alcohol and distilled water used in the solution were found to be free of any interfering substance.

This solution will produce a vapor alcohol value of .100 +/- 3% gms/210L Breath when heated to 34 Degrees Celsius +/- 0.2 Degrees Celsius in a simulator (95% Confidence).

The date of manufacture for this lot number is June 20, 2013
The expiration date for this lot number is June 19, 2015 at 11:59 p.m.

This document is a true representation of the original Certificate of Analysis.



Cecil B. Garner, President
RepCo Marketing, Inc.

Face This Side Down - This Edge In First

**BAC DataMaster
Evidence Ticket**

STATE OF MISSOURI
OWENSVILLE POLICE DEPARTMENT

BAC DATAMASTER SERIAL NUMBER 201292
08/09/14
08:35

--- DIAGNOSTIC CHECK ---

COMPUTER:	OKAY
PROGRAM (04-07-2009):	OKAY
HEATERS	
SAMPLE CHAMBER:	49c
FLOW DETECTOR:	OKAY
PUMP	
HIGH SPEED:	OKAY
DETECTOR:	OKAY
FILTERS:	OKAY
QUARTZ STANDARD:	OKAY
CALIBRATION:	OKAY

PRINTER TEST

!"#\$%&'()*+,-./0123456789:;<=>?@ABCDEFGHI
HIJKLMNOPQRSTUVWXYZ[\]^_`abcdefg h i j k l m n o
p q r s t u v w x y z () * +

Operator Signature

[Handwritten Signature] 104

2208-02

Face This Side Down - This Edge In First

**BAC DataMaster
Evidence Ticket**

STATE OF MISSOURI
OWENSVILLE POLICE DEPARTMENT

BAC DATAMASTER SERIAL NUMBER 281292
88/49/14

TESTING OFFICER:
GRIFFITH, MATTHEW/SCOTT
OFFICER I.D.# 104
PERMIT NUMBER: 248138
EXPIRATION DATE: 04/03/16
MISCELLANEOUS DATA:

--- SUPERVISOR MODE ---

BLANK TEST	.000	00:38
INTERNAL STANDARD	VERIFIED	00:38
EXTERNAL STANDARD	.102	00:39
BLANK TEST	.000	00:39
EXTERNAL STANDARD	.102	00:40
BLANK TEST	.000	00:40
EXTERNAL STANDARD	.100	00:41
BLANK TEST	.000	00:41

N = 3
STL = .1
AVG. = .1023

Operator Signature

Matthew Griffith 104

Face This Side Down - This Edge In First

**BAC DataMaster
Evidence Ticket**

STATE OF MISSOURI
OWENSVILLE POLICE DEPARTMENT

BAC DATAMASTER SERIAL NUMBER 201292
06/09/14

ARREST TIME: 00:05
SUBJECT NAME:
RFI TEST
DOB: 09/09/89 SEX: M
STATE/D.L.L.: NA/NA
ARRESTING OFFICER:
NA
OFFICER I.D.: NA
TESTING OFFICER:
GRIFFITH/JONATHAN/SCOTT
OFFICER I.D.: 104
PERMIT NUMBER: 040100
EXPIRATION DATE: 04/03/16
MISCELLANEOUS DATA:
RFI TEST
RFI TEST

--- BREATH ANALYSIS ---

BLANK TEST .000 00:47
INTERNAL STANDARD VERIFIED 00:47
RADIO INTERFERENCE

Operator Signature

Jonathan Griffith 104

2208-02



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

2

PERMIT
TYPE II

JONATHAN S GRIFFITH

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

DATAMASTER

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 4/3/2014

NUMBER 240130

EXPIRES 4/3/2016

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES