



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
 STATE PUBLIC HEALTH LABORATORY
 BREATH ALCOHOL PROGRAM
DATAMASTER MAINTENANCE REPORT

RECEIVED 1/14/14 REPORT #6

Complete this report at the time of the regular monthly preventive maintenance check (not to exceed 35 days). Complete this report whenever the instrument is serviced or repaired and whenever it is placed into service. Retain the original and send a copy within 15 days to the Breath Alcohol Program, DHSS.

REVIEWED
 By Carol Day at 8:53 am, Mar 14, 2014

DATAMASTER SN 201294	NAME OF AGENCY MACON POLICE DEPARTMENT	DATE OF INSPECTION 1-2-14
LOCATION OF INSTRUMENT (STREET AND CITY) 301 E. BOURKE ST. MACON MO 63552		TIME OF INSPECTION 0759

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

<input checked="" type="checkbox"/> DIAGNOSTIC CHECK (PRINTOUT ATTACHED)	DATE AND TIME (from printout) 01-02-14 0759
<input checked="" type="checkbox"/> COMPUTER	<input checked="" type="checkbox"/> DETECTOR
<input checked="" type="checkbox"/> PROGRAM	<input checked="" type="checkbox"/> FILTERS
<input checked="" type="checkbox"/> HEATERS SAMPLE CHAMBER .49 °C	<input checked="" type="checkbox"/> QUARTZ STANDARD
<input checked="" type="checkbox"/> FLOW DETECTOR	<input checked="" type="checkbox"/> CALIBRATION
<input checked="" type="checkbox"/> PUMP HIGH SPEED	<input checked="" type="checkbox"/> PRINTER

<input checked="" type="checkbox"/> INDICATOR LIGHTS
<input checked="" type="checkbox"/> SIMULATOR SOLUTION SUPPLIER LUTH LAB LOT # 13010 EXP. DATE 01-9-13
<input checked="" type="checkbox"/> SIMULATOR TEMP (34°C ± 0.2°C) 34 °C SIMULATOR SN 2021009 EXP. DATE N/A

CALIBRATION CHECK - (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Mark the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

<input checked="" type="checkbox"/> 0.100% STANDARD - MUST READ BETWEEN 0.095% AND 0.105% INCLUSIVE
<input type="checkbox"/> 0.080% STANDARD - MUST READ BETWEEN 0.076% AND 0.084% INCLUSIVE
<input type="checkbox"/> 0.040% STANDARD - MUST READ BETWEEN 0.038% AND 0.042% INCLUSIVE

TEST 1 .101%	TEST 2 .101%	TEST 3 .102%
---------------------	---------------------	---------------------

PERFORM R.F.I. TEST (PRINTOUT ATTACHED)

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS	(0-.04) 1	(.05-.09) 0	(.10-.14) 1	(.15-.19) 1	OVER .19 0
----------	------------------	--------------------	--------------------	--------------------	-------------------

LIST ANY NEW PARTS AND DESCRIBE ANY ALTERATION OR MODIFICATION THAT WAS MADE TO RESTORE THE INSTRUMENT TO OPERATE SATISFACTORILY AND WITHIN ESTABLISHED LIMITS (USE OTHER SIDE IF NECESSARY).

MEETS DOH STANDARDS

INSPECTING OFFICER	
SIGNATURE Jeff Brown	PRINT FULL NAME JEFF BROWN
TYPE II PERMIT NUMBER/EXPIRATION DATE 22015 05-15-14	TELEPHONE NUMBER 660 385 2195

RETURN COMPLETED REPORT TO THE:
 Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
 2875 James Blvd.
 Poplar Bluff, MO 63901



GUTH LABORATORIES, INC.

890 NORTH 6TH STREET • HARRISBURG, PA 17111-4611 • TELEPHONE: 717-654-5470

CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number 13010 of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on January 14, 2013, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain 0.1218% (w/vol) ethyl alcohol. The expiration date for this lot number is January 9, 2015 at 11:59 PM.

When used in a calibrated Simulator, operating at 34°C +/- .2°C, this solution will give a breath alcohol analysis instrument reading of 0.100 g/210L +/- 3%.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President
GUTH LABORATORIES, INC.

NIST Traceability:

Testing was conducted using Cerilliant Reference Standard lot number FN122211-02 whose values are traceable to NIST.
All balances are calibrated annually by an outside agency using NIST traceable weights.
Calibration verification is done prior to each use utilizing NIST traceable weights.

State of Missouri
DEPARTMENT OF HEALTH

P E R M I T
TYPE II

JEFF BROWN



is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field repairs, and operate the following breath analyzer(s):
for the determination of the alcoholic content of blood from a sample of expired (alveolar) air, issued under the provisions of sections 577.020 through 577.047, RSMo 1986.

DATA MASTER

Date 05/15/2012

Number 220115

Expires 05/15/2014

MSD 640-0771 (D-40)

W. B. Brown
Director of State Public Health Laboratory

W. B. Brown
Director, Department of Health

Lab. # (RT-08)

FACE THIS SIDE DOWN - THIS EDGE IN FIRST

BAC DataMaster
Evidence Ticket

STATE OF MISSOURI
MASON POLICE DEPARTMENT

BAC DATA MASTER SERIAL NUMBER 281294
87-00273A

TESTING OFFICER:

BROWN, JEFF
OFFICER I.D. # 100
PERMIT NUMBER 281294
EXPIRATION DATE 05/15/94
MISCELLANEOUS DATA:
MONTHLY MAINTENANCE
N/A

SHIP YOUR TOOL

BLANK TEST	.000	08:03
INTERNAL STANDARD	VERTIFIED	08:03
EXTERNAL STANDARD	.101	08:03
BLANK TEST	.000	08:04
EXTERNAL STANDARD	.101	08:04
BLANK TEST	.000	08:05
EXTERNAL STANDARD	.102	08:05
BLANK TEST	.000	08:06

01-03
S.M. # 11
REV. # 11/13

OPERATOR SIGNATURE Jeff Brown

Card No.

REORDER ALL SUPPLIES FROM N.P.A.S.
P.O. BOX 1435, MANSFIELD, OH 44901

FACE THIS SIDE DOWN - THIS EDGE IN FIRST

BAC DataMaster
Evidence Ticket

STATE OF MISSOURI
MASON POLICE DEPARTMENT

BAC DATA MASTER SERIAL NUMBER 281294
87-00273A

TESTING OFFICER:

COMPILE NO. 08:07
 EXPIRATION DATE BY 2000-07 08:07
 REENTER:
 SAMPLE NUMBER 454
 FLOW BLANK TEST OKAY
 FLOW
 FILTER SPEEDS OKAY
 DETECTION OKAY
 FILTERS OKAY
 OUTSIDE STANDARD OKAY
 CALIBRATION OKAY

PERFORM TEST

REORDER ALL SUPPLIES FROM N.P.A.S.
P.O. BOX 1435, MANSFIELD, OH 44901

OPERATOR SIGNATURE Jeff Brown

Card Stock No.
60021

REORDER ALL SUPPLIES FROM N.P.A.S.
P.O. BOX 1435, MANSFIELD, OH 44901

FACE THIS SIDE DOWN - THIS EDGE IN FIRST

BAC DataMaster
Evidence Ticket

STATE OF MISSOURI
POLICE DEPARTMENT

BAC UNIT/TRANSFORMER SERIAL NUMBER 201794
01/02/14

HIREST TIME: 07:40

SUBJECT NAME:

BLOND-JOE

DOB: 12-04-88 SEX: M

PLATE: MO-848080078

ARRESTING OFFICER:

BLOND, JEFF

OFFICER I.D. #: 703

TRAINING OFFICER:

BLOND, JEFF

OFFICER I.D. #: 703

REPORT NUMBER: 20115

EXPIRATION DATE: 05-15-14

MISCELLANEOUS DATA:

RFT CHECK:

N/A

--- BREATH ANALYSIS ---

ALABAMA TEST	0.00	08:18
DIPERNAI STANDARD	VERIFIED	08:19
RADIO INTERFERENCE		

OPERATOR SIGNATURE



Card Stock No.
60021

REORDER ALL SUPPLIES FROM N.P.A.S.
P.O. BOX 1435, MANSFIELD, OH 44901