



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
 STATE PUBLIC HEALTH LABORATORY
 BREATH ALCOHOL PROGRAM
 DATAMASTER MAINTENANCE REPORT

REPORT #6
 received 5/9/14-cd

Complete this report at the time of the regular monthly preventive maintenance check (not to exceed 35 days).
 Complete this report whenever the instrument is serviced or repaired and whenever it is placed into service.
 Retain the original and send a copy within 15 days to the Breath Alcohol Program, DHSS.

REVIEWED
 By Carol Day at 9:28 am, May 12, 2014

DATAMASTER SN <u>201297</u>	NAME OF AGENCY <u>Webb City Police Department</u>	DATE OF INSPECTION <u>04-26-2014</u>
LOCATION OF INSTRUMENT (STREET AND CITY) <u>211 W. Broadway St. Webb City, MO 64870</u>		TIME OF INSPECTION <u>2116</u>

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

<input checked="" type="checkbox"/> DIAGNOSTIC CHECK (PRINTOUT ATTACHED)	DATE AND TIME (from printout) <u>04-26-2014 @ 2116</u>
<input checked="" type="checkbox"/> COMPUTER	<input checked="" type="checkbox"/> DETECTOR
<input type="checkbox"/> PROGRAM	<input checked="" type="checkbox"/> FILTERS
<input checked="" type="checkbox"/> HEATERS SAMPLE CHAMBER <u>49</u> °C	<input checked="" type="checkbox"/> QUARTZ STANDARD
<input checked="" type="checkbox"/> FLOW DETECTOR	<input checked="" type="checkbox"/> CALIBRATION
<input checked="" type="checkbox"/> PUMP HIGH SPEED	<input checked="" type="checkbox"/> PRINTER

INDICATOR LIGHTS

SIMULATOR SOLUTION SUPPLIER Guth Labs LOT # 13280 EXP. DATE 10-16-2015

SIMULATOR TEMP (34°C ± 0.2°C) 34.0 °C SIMULATOR SN SD2754 EXP. DATE 07-10-2014

CALIBRATION CHECK - (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Mark the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

0.100% STANDARD - MUST READ BETWEEN 0.095% AND 0.105% INCLUSIVE
 0.080% STANDARD - MUST READ BETWEEN 0.076% AND 0.084% INCLUSIVE
 0.040% STANDARD - MUST READ BETWEEN 0.038% AND 0.042% INCLUSIVE

TEST 1 <u>.099</u>	TEST 2 <u>.100</u>	TEST 3 <u>.100</u>
--------------------	--------------------	--------------------

PERFORM R.F.I. TEST (PRINTOUT ATTACHED)

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:
 (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS <u>0</u>	(0-.04) <u>0</u>	(.05-.09) <u>2</u>	(.10-.14) <u>3</u>	(.15-.19) <u>2</u>	OVER .19 <u>4</u>
-------------------	------------------	--------------------	--------------------	--------------------	-------------------

LIST ANY NEW PARTS AND DESCRIBE ANY ALTERATION OR MODIFICATION THAT WAS MADE TO RESTORE THE INSTRUMENT TO OPERATE SATISFACTORILY AND WITHIN ESTABLISHED LIMITS (USE OTHER SIDE IF NECESSARY).

INSPECTING OFFICER	
SIGNATURE <u>Jeff Cornett #413</u>	PRINT FULL NAME <u>Jeff Cornett</u>
TYPE II PERMIT NUMBER/EXPIRATION DATE <u>230322</u>	TELEPHONE NUMBER <u>417-673-1911</u>

RETURN COMPLETED REPORT TO THE:
 Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
 2875 James Blvd.
 Poplar Bluff, MO 63901



GUTH LABORATORIES, INC.

690 NORTH 67th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-684-6470

CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number 13280 of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on October 18, 2013, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain 0.1217% (w/vol) ethyl alcohol. The expiration date for this lot number is October 16, 2015 at 11:59 PM.

When used in a calibrated Simulator, operating at 34°C +/- .2°C, this solution will give a breath alcohol analysis instrument reading of 0.100 g/210L +/- 3%.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President
GUTH LABORATORIES, INC.

NIST Traceability:

Testing was conducted using Cerilliant Reference Standard lot number FN122211-02 whose values are traceable to NIST.

All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.

**BAC DataMaster
Evidence Ticket**

STATE OF MISSOURI
WEBB CITY POLICE DEPARTMENT
BAC DATAMASTER SERIAL NUMBER 201297
24/26/14

**BAC DataMaster
Evidence Ticket**

STATE OF MISSOURI
WEBB CITY POLICE DEPARTMENT
BAC DATAMASTER SERIAL NUMBER 201297
24/26/14

**BAC DataMaster
Evidence Ticket**

STATE OF MISSOURI
WEBB CITY POLICE DEPARTMENT
BAC DATAMASTER SERIAL NUMBER 201297
24/26/14

ARREST TIME: 20:02
SUBJECT NAME:
JOE/JOHN
DOB: 01/01/01 SEX: M
STATE/D.L.#: MO/123456789
ARRESTING OFFICER:
CORNETT/JEFF/A
OFFICER I.D.#: 443
TESTING OFFICER:
SAME
OFFICER I.D.#: 443
PERMIT NUMBER: 230322
EXPIRATION DATE: 12/23/15
MISCELLANEOUS DATA:
RFI TEST

TESTING OFFICER:
CORNETT/JEFF/A
OFFICER I.D.#: 443
PERMIT NUMBER: 230322
EXPIRATION DATE: 12/23/15
MISCELLANEOUS DATA:
---- SUPERVISOR MODE ----
BLANK TEST .000 21:19
INTERNAL STANDARD .099 21:19
EXTERNAL STANDARD .000 21:20
BLANK TEST .100 21:20
EXTERNAL STANDARD .000 21:21
INTERNAL STANDARD .100 21:22
EXTERNAL STANDARD .000 21:22

--- DIAGNOSTIC CHECK ---
COMPUTER: OKAY
PROGRAM (04-07-2009): OKAY
HEATERS: 49c
SAMPLE CHAMBER: OKAY
FLOW DETECTOR: OKAY
PUMP: OKAY
HIGH SPEED: OKAY
DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY
N = 3
SIM. = .1
AVG. = .0996

PRINTER TEST
!@#%&'()*+,-./:;<=>?@ABODEFG
HIJKLMNOPQRSTUVWXYZ[\]^_`abcdefhijklmnop
qrs tuvwxz{|}~

BLANK TEST .000 21:19
INTERNAL STANDARD .099 21:19
EXTERNAL STANDARD .000 21:20
RFI TEST

Operator Signature

Operator Signature

Operator Signature



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM



PERMIT
TYPE II

JEFFREY A CORNETT

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

DATAMASTER, INTOXILYZER 8000

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 12/23/2013

NUMBER 230322

EXPIRES 12/23/2015

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

acting director
DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MO 580-0771 (6-10)

LAB 4 (16-10)

STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator: CORNETT, JEFFREY
Permit No: 230322
Date issued: 12/23/2013 Date Expires: 12/23/2015