



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
 STATE PUBLIC HEALTH LABORATORY
 BREATH ALCOHOL PROGRAM
DATAMASTER MAINTENANCE REPORT

REPORT #6

RECEIVED
 By Carol Day at 12:18 pm, Oct 28, 2014

Complete this report at the time of the regular monthly preventive maintenance check (not to exceed 35 days). Complete this report whenever the instrument is serviced or repaired and whenever it is placed into service. Retain the original and send a copy within 15 days to the Breath Alcohol Program, DHSS.

DATAMASTER SN 201306	NAME OF AGENCY MOUNTAIN VIEW POLICE	DATE OF INSPECTION 10/14/2014
LOCATION OF INSTRUMENT (STREET AND CITY) 117 NORTH PINE STREET MOUNTAIN VIEW MO 65548		TIME OF INSPECTION 9:05 am

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

<input checked="" type="checkbox"/> DIAGNOSTIC CHECK (PRINTOUT ATTACHED)	DATE AND TIME (from printout) 10/14/14 9:30
<input checked="" type="checkbox"/> COMPUTER	<input checked="" type="checkbox"/> DETECTOR
<input checked="" type="checkbox"/> PROGRAM	<input checked="" type="checkbox"/> FILTERS
<input checked="" type="checkbox"/> HEATERS SAMPLE CHAMBER _____ 49 °C	<input checked="" type="checkbox"/> QUARTZ STANDARD
<input checked="" type="checkbox"/> FLOW DETECTOR	<input checked="" type="checkbox"/> CALIBRATION
<input checked="" type="checkbox"/> PUMP HIGH SPEED	<input checked="" type="checkbox"/> PRINTER

<input checked="" type="checkbox"/> INDICATOR LIGHTS
<input checked="" type="checkbox"/> SIMULATOR SOLUTION SUPPLIER GUTH LOT # 14110 EXP. DATE 05/01/2016
<input checked="" type="checkbox"/> SIMULATOR TEMP (34°C ± 0.2°C) 3402 °C SIMULATOR SN SD1425 EXP. DATE 07/29/2015

CALIBRATION CHECK - (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Mark the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

0.100% STANDARD - MUST READ BETWEEN 0.095% AND 0.105% INCLUSIVE
 0.080% STANDARD - MUST READ BETWEEN 0.076% AND 0.084% INCLUSIVE
 0.040% STANDARD - MUST READ BETWEEN 0.038% AND 0.042% INCLUSIVE

TEST 1 .099	TEST 2 .101	TEST 3 .102
--------------	--------------	--------------

PERFORM R.F.I. TEST (PRINTOUT ATTACHED)

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS	0	(0-.04)	5	(.05-.09)	0	(.10-.14)	2	(.15-.19)	0	OVER .19	0
----------	---	---------	---	-----------	---	-----------	---	-----------	---	----------	---

LIST ANY NEW PARTS AND DESCRIBE ANY ALTERATION OR MODIFICATION THAT WAS MADE TO RESTORE THE INSTRUMENT TO OPERATE SATISFACTORILY AND WITHIN ESTABLISHED LIMITS (USE OTHER SIDE IF NECESSARY).

MONTHLY MAINTANCE ALL OK

INSPECTING OFFICER	
SIGNATURE 	PRINT FULL NAME THOMAS . R. POINDEXTER
TYPE II PERMIT NUMBER/EXPIRATION DATE 240213 04/30/2016	TELEPHONE NUMBER (417) 934-2525

RETURN COMPLETED REPORT TO THE: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
 2875 James Blvd.
 Poplar Bluff, MO 63901



GUTH LABORATORIES, INC.

590 NORTH 67th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-564-5470

CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number **14110** of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on **May 5, 2014**, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain **0.1206%** (w/vol) ethyl alcohol. The expiration date for this lot number is **May 1, 2016** at 11:59 PM.

When used in a calibrated Simulator, operating at $34^{\circ}\text{C} \pm .2^{\circ}\text{C}$, this solution will give a breath alcohol analysis instrument reading of **0.100 g/210L \pm 3%**.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President
GUTH LABORATORIES, INC.

NIST Traceability:

Testing was conducted using Cerilliant Reference Standard lot number **FN122211-02** whose values are traceable to NIST.

All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.



Simulator Calibration Report

This calibration report is to certify the alcohol reference simulator listed below has been examined and tested using standards traceable to the National Institute of Standards and Technology (NIST) in accordance with the standards set by the Missouri Department of Health and Senior Services Rules and Regulations:
19 CSR 25-30.051 (4).

Checked: 07/29/2014 Expires: 07/29/2015
Digital Therm. SN:094948
MSC Tech:DRL Temp:34.00
Agency: Mountain View Police Dept
SD 1425



Technician Printed Name: _____

DAW LUCAS

Technician Signature: _____

[Handwritten Signature]

Date: _____

07/29/2014

Contact: Missouri Safety Center

Breath-Alcohol Instrument Training Program

660-543-4834



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

2

PERMIT
 TYPE II

THOMAS R POINDEXTER

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

DATAMASTER

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 4/30/2014

NUMBER 240213

EXPIRES 4/30/2016

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MO 580-0771 (6-10)

LAB-4 (R6-10)

STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD.

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator POINDEXTER, THOMAS
 Permit No 240213
 Date Issued 4/30/2014 Date Expires 4/30/2016

BAC DataMaster Evidence Ticket

STATE OF MISSOURI

BAC DATA MASTER SERIAL NUMBER 2013006
10/14/14

TESTING OFFICER:

POINDEXTER/TROMPETER

OFFICER I.D.# 122

PERMIT NUMBER: 240213

EXPIRATION DATE: 04/30/16

ADDITIONAL DATA:

----- SUPERVISOR MODE -----

BLANK TEST	.000	09:10
INTERNAL STANDARD	VERIFIED	09:10
EXTERNAL STANDARD	.099	09:10
BLANK TEST	.000	09:11
EXTERNAL STANDARD	.101	09:12
BLANK TEST	.000	09:12
EXTERNAL STANDARD	.100	09:13
BLANK TEST	.000	09:13

R = 3
S.M. = .1
R.M. = .1

Operator Signature

James R. Poinceter

BAC DataMaster Evidence Ticket

STATE OF MISSOURI

BAC DATA MASTER SERIAL NUMBER 2013006
10/14/14
09:05

----- DIAGNOSTIC CHECK -----

- COMPUTER: OKAY
- PROGRAM (04-07-2005): OKAY
- HEATERS: OKAY
- SAMPLE CHAMBER: 49c
- FLOW DETECTOR: OKAY
- PUMP: OKAY
- HIGH SPEED: OKAY
- DETECTOR: OKAY
- FILTERS: OKAY
- QUARTZ STANDARD: OKAY
- CALIBRATION: OKAY

PRINTER TEST:

!@#%&'()*+,-./0123456789:;<=>?@BCDEFG
HIJKLMNOPQRSTUVWXYZ[\]^_`abcdefg|hijklmnop
qrstuvwxyz{|}~*"

Operator Signature

James R. Poinceter

Face This Side Down - This Edge In First

BAC DataMaster Evidence Ticket

STATE OF MISSOURI
BAC DATAMASTER SERIAL NUMBER 1991366
18414/19

ARREST TIME: 08:05
SUBJECT NAME:
RADIO TEST
DOB: 01/02/34 SEX: M
STATE/D.L.: M2/05231
ARRESTING OFFICER:
FOUNDER/THOMAS/R
OFFICER I.D.: 122
TESTING OFFICER:
FOUNDER/THOMAS/R
OFFICER I.D.: 122
PERMIT NUMBER: 248212
EXPIRATION DATE: 04/30/16
MISCELLANEOUS DATA:

--- BREATH ANALYSIS ---

BLANK TEST	LAB	08:19
INTERNAL STANDARD	VERIFIED	08:19
WTO/O INTERFERENCE		

Operator Signature *Blaine R. Pinnick*