



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
 STATE PUBLIC HEALTH LABORATORY
 BREATH ALCOHOL PROGRAM
DATAMASTER MAINTENANCE REPORT

received 10/18/14-^{REPORT #6}cd

Complete this report at the time of the regular monthly preventive maintenance check (not to exceed 35 days) Complete this report whenever the instrument is serviced or repaired and whenever it is placed into service. Retain the original and send a copy within 15 days to the Breath Alcohol Program, DHSS.

REVIEWED
By Carol Day at 12:17 pm, Oct 28, 2014

DATAMASTER SN 202008	NAME OF AGENCY Missouri State Highway Patrol	DATE OF INSPECTION 10/12/2014
LOCATION OF INSTRUMENT (STREET AND CITY) Andrew County Sheriff's Department, Savannah		TIME OF INSPECTION 9:30 pm

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

<input checked="" type="checkbox"/> DIAGNOSTIC CHECK (PRINTOUT ATTACHED)	DATE AND TIME (from printout) 10/12/14 21:43
<input checked="" type="checkbox"/> COMPUTER	<input checked="" type="checkbox"/> DETECTOR
<input checked="" type="checkbox"/> PROGRAM	<input checked="" type="checkbox"/> FILTERS
<input checked="" type="checkbox"/> HEATERS SAMPLE CHAMBER 49 °C	<input checked="" type="checkbox"/> QUARTZ STANDARD
<input checked="" type="checkbox"/> FLOW DETECTOR	<input checked="" type="checkbox"/> CALIBRATION
<input checked="" type="checkbox"/> PUMP HIGH SPEED	<input checked="" type="checkbox"/> PRINTER

<input checked="" type="checkbox"/> INDICATOR LIGHTS
<input checked="" type="checkbox"/> SIMULATOR SOLUTION SUPPLIER Guth Laboratories, Inc LOT # 13280 EXP. DATE 10/16/2015
<input checked="" type="checkbox"/> SIMULATOR TEMP (34°C ± 0.2°C) 34.00 °C SIMULATOR SN 1101 EXP. DATE 11/18/2014

CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Mark the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

0.100% STANDARD - MUST READ BETWEEN 0.095% AND 0.105% INCLUSIVE
 0.080% STANDARD - MUST READ BETWEEN 0.076% AND 0.084% INCLUSIVE
 0.040% STANDARD - MUST READ BETWEEN 0.038% AND 0.042% INCLUSIVE

TEST 1 .098	TEST 2 .099	TEST 3 .099
-------------	-------------	-------------

PERFORM R.F.I. TEST (PRINTOUT ATTACHED)

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS	0	(0-.04)	1	(.05-.09)	4	(.10-.14)	1	(.15-.19)	2	OVER .19	2
----------	---	---------	---	-----------	---	-----------	---	-----------	---	----------	---

LIST ANY NEW PARTS AND DESCRIBE ANY ALTERATION OR MODIFICATION THAT WAS MADE TO RESTORE THE INSTRUMENT TO OPERATE SATISFACTORILY AND WITHIN ESTABLISHED LIMITS (USE OTHER SIDE IF NECESSARY).

INSPECTING OFFICER	
SIGNATURE 	PRINT FULL NAME Steven J. Force
TYPE II PERMIT NUMBER/EXPIRATION DATE 240090 03/11/2016	TELEPHONE NUMBER (816) 622-0800

RETURN COMPLETED REPORT TO THE: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
 2875 James Blvd.
 Poplar Bluff, MO 63901



GUTH LABORATORIES, INC.

590 NORTH 67th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-664-6470

CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number 13280 of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on October 18, 2013, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain 0.1217% (w/vol) ethyl alcohol. The expiration date for this lot number is October 16, 2015 at 11:59 PM.

When used in a calibrated Simulator, operating at 34°C +/- .2°C, this solution will give a breath alcohol analysis instrument reading of 0.100 g/210L +/- 3%.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President
GUTH LABORATORIES, INC.

NIST Traceability:

Testing was conducted using Cerilliant Reference Standard lot number FN122211-02 whose values are traceable to NIST.

All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

2

**PERMIT
TYPE II**

STEVEN J FORCE

hereby authorized to instruct and supervise operators, train instructors; inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

DATAMASTER, INTOX DMT

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

ISSUE DATE 3/11/2014

PERMIT NUMBER 240090

EXPIRES 3/11/2016

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

,acting director

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

LAB-4 (R6-10)



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.



Operator **FORCE, STEVEN**
Permit No **240090**
Date Issued **3/11/2014** Date Expires **3/11/2016**

Face This Side Down - This Edge In First

BAC DataMaster Evidence Ticket

DEPARTMENT OF TRANSPORTATION
NEW YORK STATE POLICE

DATE: 10/15/04
PAGE: 1

OFFICER'S ID NUMBER

OFFICER'S NAME

OFFICER'S RANK

OFFICER'S DIVISION

OFFICER'S STATION

OFFICER'S ASSIGNMENT

OFFICER'S PHONE

OFFICER'S ADDRESS


OFFICER'S CITY

OFFICER'S STATE

OFFICER'S SIGNATURE

Signature of Operator
Date of Collection: 10/15/04
Time of Collection: 10:00 AM

Operator Signature



Face This Side Down - This Edge In First

BAC DataMaster Evidence Ticket

DEPARTMENT OF TRANSPORTATION
NEW YORK STATE POLICE

DATE: 10/15/04
PAGE: 1

OFFICER'S ID NUMBER

OFFICER'S NAME

OFFICER'S RANK

OFFICER'S DIVISION

OFFICER'S STATION

OFFICER'S ASSIGNMENT

OFFICER'S PHONE

OFFICER'S ADDRESS

OFFICER'S CITY

OFFICER'S STATE

OFFICER'S SIGNATURE

DATE OF COLLECTION

TIME OF COLLECTION

OFFICER'S SIGNATURE

DATE OF COLLECTION


TIME OF COLLECTION

OFFICER'S SIGNATURE

DATE OF COLLECTION

TIME OF COLLECTION

Operator Signature



Face This Side Down - This Edge In First

BAC DataMaster Evidence Ticket

MISSOURI STATE LABORATORY
400 UNIVERSITY DRIVE, JEFFERSON CITY, MO 64101
(816) 425-5000

LABORATORY NUMBER

LABORATORY NAME

DATE

LABORATORY ADDRESS

CITY, STATE, ZIP CODE

LABORATORY CONTACT

PHONE

LABORATORY FAX

LABORATORY E-MAIL

LABORATORY WEBSITE

LABORATORY CONTACT NAME

LABORATORY CONTACT ADDRESS

LABORATORY CONTACT CITY, STATE, ZIP CODE

LABORATORY CONTACT PHONE

LABORATORY FAX

LABORATORY USE ONLY

Operator Signature

