



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
 STATE PUBLIC HEALTH LABORATORY
 BREATH ALCOHOL PROGRAM
DATAMASTER MAINTENANCE REPORT

REPORT #6

received 7/8/14-cd

Complete this report at the time of the regular monthly preventive maintenance check (not to exceed 35 days).
 Complete this report whenever the instrument is serviced or repaired and whenever it is placed into service.
 Retain the original and send a copy within 15 days to the Breath Alcohol Program, DHSS.

REVIEWED
 By Carol Day at 12:37 pm, Aug 14, 2014

DATAMASTER SN 202009	NAME OF AGENCY Missouri State Highway Patrol	DATE OF INSPECTION 06/28/2014
LOCATION OF INSTRUMENT (STREET AND CITY) Ray County Sheriff's Department, 200 W. 9th St., Henrietta, MO 64036		TIME OF INSPECTION 11:39 am

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

<input checked="" type="checkbox"/> DIAGNOSTIC CHECK (PRINTOUT ATTACHED)	DATE AND TIME (from printout) <u>06-28-2014 at 1139 hrs</u>
<input checked="" type="checkbox"/> COMPUTER	<input checked="" type="checkbox"/> DETECTOR
<input checked="" type="checkbox"/> PROGRAM	<input checked="" type="checkbox"/> FILTERS
<input checked="" type="checkbox"/> HEATERS SAMPLE CHAMBER <u>50</u> °C	<input checked="" type="checkbox"/> QUARTZ STANDARD
<input checked="" type="checkbox"/> FLOW DETECTOR	<input checked="" type="checkbox"/> CALIBRATION
<input checked="" type="checkbox"/> PUMP HIGH SPEED	<input checked="" type="checkbox"/> PRINTER

INDICATOR LIGHTS

SIMULATOR SOLUTION SUPPLIER Guth Laboratories, Inc. LOT # 13280 EXP. DATE 10/16/2015

SIMULATOR TEMP (34°C ± 0.2°C) 34.0 °C SIMULATOR SN MP2126 EXP. DATE 06/03/2015

CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Mark the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

0.100% STANDARD - MUST READ BETWEEN 0.095% AND 0.105% INCLUSIVE
 0.080% STANDARD - MUST READ BETWEEN 0.076% AND 0.084% INCLUSIVE
 0.040% STANDARD - MUST READ BETWEEN 0.038% AND 0.042% INCLUSIVE

TEST 1 .095 BAC	TEST 2 .096 BAC	TEST 3 .096 BAC
------------------	------------------	------------------

PERFORM R.F.I. TEST (PRINTOUT ATTACHED)

**INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:
 (DO NOT INCLUDE SELF-ADMINISTERED TESTS)**

REFUSALS	0	(0-.04)	0	(.05-.09)	0	(.10-.14)	0	(.15-.19)	1	OVER .19	0
----------	---	---------	---	-----------	---	-----------	---	-----------	---	----------	---

LIST ANY NEW PARTS AND DESCRIBE ANY ALTERATION OR MODIFICATION THAT WAS MADE TO RESTORE THE INSTRUMENT TO OPERATE SATISFACTORILY AND WITHIN ESTABLISHED LIMITS (USE OTHER SIDE IF NECESSARY).

Operating within MODHSS specifications
 Solution bottle # 1504, Guth 12V500 simulator

INSPECTING OFFICER	
SIGNATURE 	PRINT FULL NAME TPR C.D. Moeller #662
TYPE II PERMIT NUMBER/EXPIRATION DATE 240064 03/07/2016	TELEPHONE NUMBER (816) 622-0800

RETURN COMPLETED REPORT TO THE: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
 2875 James Blvd.
 Poplar Bluff, MO 63901



GUTH LABORATORIES, INC.

590 NORTH 67th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-564-5470

CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number 13280 of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on October 18, 2013, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain 0.1217% (w/vol) ethyl alcohol. The expiration date for this lot number is October 16, 2015 at 11:59 PM.

When used in a calibrated Simulator, operating at 34°C +/- .2°C, this solution will give a breath alcohol analysis instrument reading of 0.100 g/210L +/- 3%.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President
GUTH LABORATORIES, INC.

NIST Traceability:

Testing was conducted using Cerilliant Reference Standard lot number FN122211-02 whose values are traceable to NIST.

All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

2

PERMIT
TYPE II
CLAY D MOELLER

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

DATAMASTER, INTOX DMT

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 3/7/2014
NUMBER 240064
EXPIRES 3/7/2016


DIRECTOR OF STATE PUBLIC HEALTH LABORATORY


DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MO 580-0771 (6-10)

LAB-4 (R6-10)



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.



Operator MOELLER, CLAY
Permit No 240064
Date Issued 3/7/2014 Date Expires 3/7/2016

Face This Side Down - This Edge In First

BAC DataMaster
Evidence Ticket

MISSOURI STATE HIGHWAY PATROL
BAC DATAMASTER SERIAL NUMBER 202009
06/28/14
11:39

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY
PROGRAM (04-07-2009): OKAY
HEATERS
SAMPLE CHAMBER: 50c
FLOW DETECTOR: OKAY
PUMP
HIGH SPEED: OKAY
DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY

PRINTER TEST

!"#\$%&'()*+,-./0123456789:;<=>?@ABCDEFGH
IJKLMNOPQRSTUVWXYZ[\]^_`abcdefghijklmnop
qrstuvwxyz{|}~

Operator Signature

TPR CD Miller

Face This Side Down - This Edge In First

BAC DataMaster
Evidence Ticket

MISSOURI STATE HIGHWAY PATROL
BAC DATAMASTER SERIAL NUMBER 202009
06/28/14

TESTING OFFICER:
MOELLER/C/D
OFFICER I.D.: 0662
PERMIT NUMBER: 240064
EXPIRATION DATE: 03/07/16
MISCELLANEOUS DATA:
INV#F-25460

--- SUPERVISOR MODE ---

BLANK TEST	.000	11:41
INTERNAL STANDARD	VERIFIED	11:41
EXTERNAL STANDARD	.095	11:42
BLANK TEST	.000	11:42
EXTERNAL STANDARD	.096	11:43
BLANK TEST	.000	11:43
EXTERNAL STANDARD	.096	11:44
BLANK TEST	.000	11:44

N = 3
SIM. = .1
AVG. = .0956

Operator Signature

TPR CD Miller

Face This Side Down - This Edge In First

BAC DataMaster
Evidence Ticket

MISSOURI STATE HIGHWAY PATROL
BAC DATAMASTER SERIAL NUMBER 202009
06/28/14

ARREST TIME: 00:01
SUBJECT NAME:
RFI/TEST
DOB: 01/01/01 SEX: M
STATE/D.L.: MO/X
ARRESTING OFFICER:
MOELLER/C/D
OFFICER I.D.: 0662
TESTING OFFICER:
MOELLER/C/D
OFFICER I.D.: 0662
PERMIT NUMBER: 240064
EXPIRATION DATE: 03/07/16
MISCELLANEOUS DATA:
INV# F-25460

--- BREATH ANALYSIS ---

BLANK TEST	.000	11:46
INTERNAL STANDARD	VERIFIED	11:46
RADIO INTERFERENCE		

Face This Side Down - This Edge In First

BAC DataMaster
Evidence Ticket

MISSOURI STATE HIGHWAY PATROL
BAC DATAMASTER SERIAL NUMBER 202009
06/28/14

ARREST TIME: 00:01
SUBJECT NAME:
BLANK/TEST
DOB: 01/01/01 SEX: M
STATE/D.L.: MO/X
ARRESTING OFFICER:
MOELLER/C/D
OFFICER I.D.: 0662
TESTING OFFICER:
MOELLER/C/D
OFFICER I.D.: 0662
PERMIT NUMBER: 240064
EXPIRATION DATE: 03/07/16
MISCELLANEOUS DATA:
INV# F-25460

--- BREATH ANALYSIS ---

BLANK TEST	.000	11:49
INTERNAL STANDARD	VERIFIED	11:49
SUBJECT SAMPLE	.000	11:49
BLANK TEST	.000	11:50

Operator Signature

TPR *AMM* 7662

Operator Signature

TPR *AMM* 7662