



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
 STATE PUBLIC HEALTH LABORATORY
 BREATH ALCOHOL PROGRAM
DATAMASTER MAINTENANCE REPORT

received 1/28/14-cc REPORT #6

REVIEWED
 By Carol Day at 1:09 pm, Feb 04, 2014

Complete this report at the time of the regular monthly preventive maintenance check (not to exceed 3 months). Complete this report whenever the instrument is serviced or repaired and whenever it is placed into service. Retain the original and send a copy within 15 days to the Breath Alcohol Program, DHSS.

DATAMASTER SN: 202011 NAME OF AGENCY: MSHP I DATE OF INSPECTION: 1-13-14
 LOCATION OF INSTRUMENT (STREET AND CITY): DENT CO JAIL TIME OF INSPECTION: 1636

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

DIAGNOSTIC CHECK (PRINTOUT ATTACHED) DATE AND TIME (from printout) 1-13-14 1636

COMPUTER DETECTOR
 PROGRAM FILTERS
 HEATERS SAMPLE CHAMBER 49 °C QUARTZ STANDARD
 FLOW DETECTOR CALIBRATION
 PUMP HIGH SPEED PRINTER

INDICATOR LIGHTS

SIMULATOR SOLUTION SUPPLIER REPCO MARKETING LOT # 13001 EXP. DATE 3-7-15

SIMULATOR TEMP (34°C ± 0.2°C) 34.01 °C SIMULATOR SN 0919 EXP. DATE 6-7-14

CALIBRATION CHECK - (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Mark the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

0.100% STANDARD - MUST READ BETWEEN 0.095% AND 0.105% INCLUSIVE
 0.080% STANDARD - MUST READ BETWEEN 0.078% AND 0.084% INCLUSIVE
 0.040% STANDARD - MUST READ BETWEEN 0.038% AND 0.042% INCLUSIVE

TEST 1 .100% TEST 2 .101% TEST 3 .101%

PERFORM R.F.I. TEST (PRINTOUT ATTACHED)

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS 0 (0-.04) 0 (.05-.09) 0 (.10-.14) 0 (.15-.19) 0 OVER .19 0

LIST ANY NEW PARTS AND DESCRIBE ANY ALTERATION OR MODIFICATION THAT WAS MADE TO RESTORE THE INSTRUMENT TO OPERATE SATISFACTORILY AND WITHIN ESTABLISHED LIMITS (USE OTHER SIDE IF NECESSARY).

OPERATING WITHIN MODOHSS SPECIFICATIONS

INSPECTING OFFICER

SIGNATURE: K.D. Wilmont PRINT FULL NAME: K.D. WILMONT
 TYPE II PERMIT NUMBER/EXPIRATION DATE: 220051 2-9-14 TELEPHONE NUMBER: (573) 368-2345

RETURN COMPLETED REPORT TO THE: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
 2875 James Blvd.
 Poplar Bluff, MO 63901

CERTIFICATE OF ANALYSIS**MANUFACTURER AND SUPPLIER: RepCo Marketing, Inc.****LOT NUMBER: 13001****EXPIRATION DATE: March 7, 2015 at 11:59 p.m.**

RepCo Marketing, Inc. certifies the following:

RepCo Marketing, Inc. manufactured, tested and supplied Lot Number 13001 of Alcohol Certified Solution for simulators. Random samples of said lot number were analyzed by an independent laboratory utilizing a gas chromatograph and found to contain .1215 gms/dl +/- .003 gms/dl wt/vol ethanol (95% Confidence).

The alcohol and distilled water used in the solution were found to be free of any interfering substance.

This solution will produce a vapor alcohol value of .100 +/- 3% gms/210L Breath when heated to 34 Degrees Celsius +/- 0.2 Degrees Celsius in a simulator (95% Confidence).

The date of manufacture for this lot number is March 8, 2013
The expiration date for this lot number is March 7, 2015 at
11:59 p.m.

This document is a true representation of the original Certificate of Analysis.



Cecil B. Garner, President
RepCo Marketing, Inc.

State of Missouri
DEPARTMENT OF HEALTH



P E R M I T
TYPE II



KYLE D WILMONT

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field repairs, and operate the following breath analyzer(s):

DATAMASTER

for the determination of the alcoholic content of blood from a sample of expired (alveolar) air. Issued under the provisions of sections 577.041, RSMo 1986.

Date 02/09/2012

Number 220051

Expires 02/09/2014

MO 580-0771 (7-88)

Director of State Public Health Laboratory

Director, Department of Health

Lab. 4 (87-88)

Face This Side Down - This Edge In First

BAC DataMaster Evidence Ticket

INSTRUMENT IDENTIFICATION CONTROL
BAC DATAMASTER SERIAL NUMBER PREFIX
01-1-1-1-1

DATE THIS OPERATOR
RECEIVED (MM/DD)
OFFICER (LAST, FIRST)
POLICE NUMBER (XXXXXX)
EXPIRATION DATE (MM/DD/YY)
POLICE AGENCY (NAME)

DATE RECEIVED

BLANK YES?	.0000	16450
EXTERNAL STANDARD	VERIFIED	16450
BLANK YES?	.0000	16450
EXTERNAL STANDARD	.100	17400
BLANK YES?	.0000	17400
EXTERNAL STANDARD	.100	17400
BLANK YES?	.0000	17400

DATE
TIME
OFFICER

Operator Signature

K.D. Wilkins

Face This Side Down - This Edge In First

BAC DataMaster Evidence Ticket

INSTRUMENT IDENTIFICATION CONTROL
BAC DATAMASTER SERIAL NUMBER PREFIX
01-1-1-1-1
16450

DATE RECEIVED

DATE RECEIVED	MM/DD
OFFICER (LAST, FIRST)	MM/DD
POLICE NUMBER (XXXXXX)	MM/DD
EXPIRATION DATE (MM/DD/YY)	MM/DD
POLICE AGENCY (NAME)	MM/DD
DATE RECEIVED	MM/DD
OFFICER (LAST, FIRST)	MM/DD
POLICE NUMBER (XXXXXX)	MM/DD
EXPIRATION DATE (MM/DD/YY)	MM/DD
POLICE AGENCY (NAME)	MM/DD
DATE RECEIVED	MM/DD
OFFICER (LAST, FIRST)	MM/DD
POLICE NUMBER (XXXXXX)	MM/DD
EXPIRATION DATE (MM/DD/YY)	MM/DD
POLICE AGENCY (NAME)	MM/DD

DATE RECEIVED

INSTRUMENT IDENTIFICATION CONTROL
BAC DATAMASTER SERIAL NUMBER PREFIX
01-1-1-1-1
16450

Operator Signature

K.D. Wilkins

Face This Side Down - This Edge In First

BAC DataMaster Evidence Ticket

MISSOURI STATE HIGHWAY PATROL
BAC DATAMASTER SERIAL NUMBER 200001
01-17-14

DRIVER LICENSE # 104 891
SUBJECT # 00000
DOB # 0000
DOB # 000000000000000000000000
STATE # 000000000000000000000000
HIRE # 000000000000000000000000
MID # 0000
OFFICER # 000000000000000000000000
TESTING OFFICER #
STATE #
OFFICER # 000000000000000000000000
PRODUCT NUMBER # 000000000000000000000000
EXPIRES DATE 00/00/0000
RECALIBRATION DATE
YES

000000000000000000000000

000000000000000000000000 0000 0000
000000000000000000000000 00000000 000000
000000000000000000000000

Operator Signature

K.D. Wilmas