MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
STATE PUBLIC HEALTH LABORATORY
BREATH ALCOHOL PROGRAM

DATAMASTER MAINTENANCE REPORT

received 8/11/14 cd

REPORT #6

Complete this report at the time of the regular monthly preventive maintenance check (not to exceed 35 days). Complete this report whenever the instrument is serviced or repaired and whenever it is placed into service. Retain the original and send a copy within 15 days to the Breath Alcohol Program, DHSS.

REVIEWED
By Carol Day at 2:16 pm, Aug 27, 2014

DATAMASTER SN	NAME OF AGENCY	DATE OF INSPECTION
202013	Missouri State Highway Patrol	08/02/2014
LOCATION OF INSTRUMENT (STREET AND CITY)		TIME OF INSPECTION
4315 HOLLY TREE LANE, FARMINGTON (US 67 AT HOLLY TREE CKPT)		10:10 pm

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

- [x] DIAGNOSTIC CHECK (PRINTOUT ATTACHED) — DATE AND TIME (from printout) 08/02/2014 AT 2213 HOURS
- [x] COMPUTER — [x] DETECTOR
- [x] PROGRAM — [x] FILTERS
- [x] HEATERS SAMPLE CHAMBER 49 °C — [x] QUARTZ STANDARD
- [x] FLOW DETECTOR — [x] CALIBRATION
- [x] PUMP HIGH SPEED — [x] PRINTER
- [x] INDICATOR LIGHTS
- [x] SIMULATOR SOLUTION SUPPLIER GUTH LABORATORIES LOT # 13290 EXP. DATE 10/29/2015
- [x] SIMULATOR TEMP (34°C ± 0.2°C) 34 °C SIMULATOR SN G11095 EXP. DATE 05/16/2015
- [x] CALIBRATION CHECK – **(ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)**

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Mark the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

- [x] 0.100% STANDARD - MUST READ BETWEEN 0.095% AND 0.105% INCLUSIVE
- [ ] 0.080% STANDARD - MUST READ BETWEEN 0.076% AND 0.084% INCLUSIVE
- [ ] 0.040% STANDARD - MUST READ BETWEEN 0.038% AND 0.042% INCLUSIVE

TEST 1	TEST 2	TEST 3
.099	.099	.100

- [x] PERFORM R.F.I. TEST (PRINTOUT ATTACHED)

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS	(0-.04)	(.05-.09)	(.10-.14)	(.15-.19)	OVER .19
0	0	0	0	0	0

LIST ANY NEW PARTS AND DESCRIBE ANY ALTERATION OR MODIFICATION THAT WAS MADE TO RESTORE THE INSTRUMENT TO OPERATE SATISFACTORILY AND WITHIN ESTABLISHED LIMITS (USE OTHER SIDE IF NECESSARY).

This instrument is operating within established limits.

INSPECTING OFFICER

SIGNATURE	PRINT FULL NAME
Cpl. D.B. Reed	Corporal Dustin B. Reed
TYPE II PERMIT NUMBER/EXPIRATION DATE	TELEPHONE NUMBER
240068 03/07/2016	(573) 431-0166

RETURN COMPLETED REPORT TO THE: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
2875 James Blvd.
Poplar Bluff, MO 63901

Face This Side Down - This Edge In First

BAC DataMaster
Evidence Ticket

MISSOURI STATE HIGHWAY PATROL
BAC DATAMASTER SERIAL NUMBER 202013
08/02/14
22:13

--- DIAGNOSTIC CHECK ---

COMPUTER:	OKAY
PROGRAM (04-07-2009):	OKAY
HEATERS	
SAMPLE CHAMBER:	49c
FLOW DETECTOR:	OKAY
PUMP	
HIGH SPEED:	OKAY
DETECTOR:	OKAY
FILTERS:	OKAY
QUARTZ STANDARD:	OKAY
CALIBRATION:	OKAY

PRINTER TEST
!"#$%&'()*+,-./0123456789:;<=>?@ABCDEFG
HIJKLMNOPQRSTUVWXYZ[\]^_`abcdefghijklmno
pqrstuvwxyz{|}→¯

ator Signature Cpl. D.B. Reed

Face This Side Down - This Edge In First

BAC DataMaster
Evidence Ticket

MISSOURI STATE HIGHWAY PATROL
BAC DATAMASTER SERIAL NUMBER 202013
08/02/14

TESTING OFFICER:
REED/D/B
OFFICER I.D.: 988
PERMIT NUMBER: 240068
EXPIRATION DATE: 03/07/16
MISCELLANEOUS DATA:
CHECKPOINT 67 AT HOLLYTREE

--- SUPERVISOR MODE ---

BLANK TEST	.000	22:20
INTERNAL STANDARD	VERIFIED	22:20
EXTERNAL STANDARD	.099	22:20
BLANK TEST	.000	22:21
EXTERNAL STANDARD	.099	22:22
BLANK TEST	.000	22:22
EXTERNAL STANDARD	.100	22:23
BLANK TEST	.000	22:24

N = 3
SIM. = .1
AVG. = .0993

Operator Signature Cpl. D.B. Reed

Face This Side Down - This Edge In First

BAC DataMaster
Evidence Ticket

MISSOURI STATE HIGHWAY PATROL
BAC DATAMASTER SERIAL NUMBER 202013
08/02/14

ARREST TIME: 22:00
SUBJECT NAME:
TEST
DOB: 11/11/11 SEX: M
STATE/D.L.: MO/123456
ARRESTING OFFICER:
REED/D/B
OFFICER I.D.: 988
TESTING OFFICER:
REED/D/B
OFFICER I.D.: 988
PERMIT NUMBER: 240068
EXPIRATION DATE: 03/07/16
MISCELLANEOUS DATA:
RFI TEST

--- BREATH ANALYSIS ---

BLANK TEST	.000	22:26
INTERNAL STANDARD	VERIFIED	22:26
RADIO INTERFERENCE		

Operator Signature Cpl. D B Reed

GUTH LABORATORIES, INC.
590 NORTH 67th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-564-5470

CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number **13290** of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on **October 31, 2013,** using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain **0.1202%** (w/vol) ethyl alcohol. The expiration date for this lot number is **October 29, 2015** at 11:59 PM.

When used in a calibrated Simulator, operating at 34°C +/- .2°C, this solution will give a breath alcohol analysis instrument reading of **0.100 g/210L +/- 3%.**

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President
GUTH LABORATORIES, INC.

NIST Traceability:
*Testing was conducted using Cerilliant Reference Standard lot number **FN122211-02** whose values are traceable to NIST.*
All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.

STATE OF MISSOURI

DEPARTMENT OF HEALTH AND SENIOR SERVICES

BREATH ALCOHOL PROGRAM

2

PERMIT

TYPE II

DUSTIN B REED

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

DATAMASTER, INTOX DMT

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 3/7/2014

NUMBER 240068

EXPIRES 3/7/2016

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MO 580-0771 (6-10)

LAB-4 (06-10)

STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator REED, DUSTIN
Permit No 240068
Date Issued 3/7/2014 **Date Expires** 3/7/2016