



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 STATE PUBLIC HEALTH LABORATORY  
 BREATH ALCOHOL PROGRAM  
**DATAMASTER MAINTENANCE REPORT**

REPORT #6

Complete this report at the time of the regular monthly preventive maintenance check (not to exceed 35 days)  
 Complete this report whenever the instrument is serviced or repaired and whenever it is placed into service.  
 Retain the original and send a copy within 15 days to the Breath Alcohol Program, DHSS.

**RECEIVED**  
 By Carol Day at 9:00 am, Jan 16, 2014

DATAMASTER SN 203043	NAME OF AGENCY Odessa Police Department	DATE OF INSPECTION 01/12/2014
LOCATION OF INSTRUMENT (STREET AND CITY) 101 North Second Street Odessa		TIME OF INSPECTION 10:50 am

**CHECKLIST:** Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

<input checked="" type="checkbox"/> DIAGNOSTIC CHECK (PRINTOUT ATTACHED)	DATE AND TIME (from printout) 10:52am
<input checked="" type="checkbox"/> COMPUTER	<input checked="" type="checkbox"/> DETECTOR
<input checked="" type="checkbox"/> PROGRAM	<input checked="" type="checkbox"/> FILTERS
<input checked="" type="checkbox"/> HEATERS SAMPLE CHAMBER 50 °C	<input checked="" type="checkbox"/> QUARTZ STANDARD
<input checked="" type="checkbox"/> FLOW DETECTOR	<input checked="" type="checkbox"/> CALIBRATION
<input checked="" type="checkbox"/> PUMP HIGH SPEED	<input checked="" type="checkbox"/> PRINTER

INDICATOR LIGHTS

SIMULATOR SOLUTION SUPPLIER Repc Marketing Inc. LOT # 12002 EXP. DATE 08/29/2014

SIMULATOR TEMP (34°C ± 0.2°C) 34.0 °C SIMULATOR SN SD2765 EXP. DATE 12/10/2014

CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Mark the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

0.100% STANDARD - MUST READ BETWEEN 0.095% AND 0.105% INCLUSIVE  
 0.080% STANDARD - MUST READ BETWEEN 0.076% AND 0.084% INCLUSIVE  
 0.040% STANDARD - MUST READ BETWEEN 0.038% AND 0.042% INCLUSIVE

TEST 1  .100	TEST 2  .101	TEST 3  .102
--------------	--------------	--------------

PERFORM R.F.I. TEST (PRINTOUT ATTACHED)

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:  
 (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS	(0-.04) 1	(.05-.09) 0	(.10-.14) 0	(.15-.19) 0	OVER .19 1
----------	-----------	-------------	-------------	-------------	------------

LIST ANY NEW PARTS AND DESCRIBE ANY ALTERATION OR MODIFICATION THAT WAS MADE TO RESTORE THE INSTRUMENT TO OPERATE SATISFACTORILY AND WITHIN ESTABLISHED LIMITS (USE OTHER SIDE IF NECESSARY).

<b>INSPECTING OFFICER</b>	
SIGNATURE	PRINT FULL NAME Christopher Bodenhamer
TYPE II PERMIT NUMBER/EXPIRATION DATE 230300 12/11/2015	TELEPHONE NUMBER (816) 721-8236

RETURN COMPLETED REPORT TO THE: **Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office  
 2875 James Blvd.  
 Poplar Bluff, MO 63901**

REPCO MARKETING, INC.

3101-188 STONYBROOK DRIVE  
RALEIGH, N.C. 27604  
919-876-3480

CERTIFICATE OF ANALYSIS

MANUFACTURER AND SUPPLIER: RepCo Marketing, Inc.

LOT NUMBER: 12002

EXPIRATION DATE: August 29, 2014 at 11:59 p.m.

RepCo Marketing, Inc. certifies the following:

RepCo Marketing, Inc. manufactured, tested and supplied Lot Number 12002 of Alcohol Certified Solution for simulators. Random samples of said lot number were analyzed by an independent laboratory utilizing a gas chromatograph and found to contain .1209 gms/dl +/- .003 gms/dl wt/vol ethanol (95% Confidence).

The alcohol and distilled water used in the solution were found to be free of any interfering substance.

This solution will produce a vapor alcohol value of .100 +/- 3% gms/210L Breath when heated to 34 Degrees Celsius +/- 0.2 Degrees Celsius in a simulator (95% Confidence).

The date of manufacture for this lot number is August 30, 2012  
The expiration date for this lot number is August 29, 2014 at 11:59 p.m.

This document is a true representation of the original Certificate of Analysis.



Cecil B. Garner, President  
RepCo Marketing, Inc.

Face This Side Down - This Edge In First

# BAC DataMaster Evidence Ticket

STATE OF MISSOURI  
ODESSA POLICE DEPARTMENT

BAC DATAMASTER SERIAL NUMBER 200043  
01/12/14  
16:52

--- DIAGNOSTIC LOG ---

COMPUTER: OKAY

PROGRAM (04-07-2009): OKAY

HEATERS

SAMPLE CHAMBER: 50c

FLOW DETECTOR: OKAY

PUMP

HIGH SPEED: OKAY

DETECTOR: OKAY

FILTERS: OKAY

QUARTZ STANDARD: OKAY

CALIBRATION: OKAY

### PRINTER TEST

!"#\$%&'()\*+,-./0123456789:;<=>?@ABCDEFGHI  
HIJKLMNOPQRSTUVWXYZ[\]^\_`abcdefghijklmnop  
qrstuvwxyz{|}~"

Face This Side Down - This Edge In First

# BAC DataMaster Evidence Ticket

STATE OF MISSOURI  
ODESSA POLICE DEPARTMENT

BAC DATAMASTER SERIAL NUMBER 200043  
01/12/14

TESTING OFFICER:  
SODENHAGER/5041  
OFFICER I.D.# 505  
PERMIT NUMBER 221300  
EXPIRATION DATE 12/31/15  
MISCELLANEOUS DATA

--- BUREAU TEST LOG ---

BLANK TEST	1.00	1.00
INTERNAL STANDARD	1.00	1.00
EXTERNAL STANDARD	1.00	1.00
BLANK TEST	1.00	1.00
EXTERNAL STANDARD	1.00	1.00
BLANK TEST	1.00	1.00
EXTERNAL STANDARD	1.00	1.00
BLANK TEST	1.00	1.00

N = 3  
S.D. = .11  
AVG. = 1.00

Face This Side Down - This Edge In First

# BAC DataMaster Evidence Ticket

STATE OF MISSOURI  
ODESSA POLICE DEPARTMENT

BAC DATAMASTER SERIAL NUMBER 200043  
01/12/14

ARREST TIME: 09:10

SUBJECT NAME:

RF/TEST

DOB: 01/01/12

STATE/LOCAL NO: 123456

ARRESTING OFFICER:

SODENHAGER/5041

OFFICER I.D.# 505

TESTING OFFICER:

SODENHAGER/5041

OFFICER I.D.# 505

PERMIT NUMBER 221300

EXPIRATION DATE 12/31/15

MISCELLANEOUS DATA

N N

--- BUREAU TEST LOG ---

BLANK TEST	1.00	1.00
INTERNAL STANDARD	1.00	1.00
RADIO INTERNAL STANDARD	1.00	1.00

for Signature





STATE OF MISSOURI  
DEPARTMENT OF HEALTH AND SENIOR SERVICES  
BREATH ALCOHOL PROGRAM

2

PERMIT  
TYPE II

CHRISTOPHER BODENHAMER

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

**DATAMASTER**

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.


DATE 12/11/2013

NUMBER 230300

EXPIRES 12/11/2015

MO 560-0771 (8-10)

  
DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

  
acting director  
DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

LAB-4 (8-10)

 STATE OF MISSOURI  
DEPARTMENT OF HEALTH AND SENIOR SERVICES  
BREATH ALCOHOL PROGRAM

**INSTRUMENT OPERATOR CARD**

*The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.*



Operator BODENHAMER, CHRISTOPHER  
Permit No 230300  
Date Issued 12/11/2013 Date Expires 12/11/2015