



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 STATE PUBLIC HEALTH LABORATORY  
 BREATH ALCOHOL PROGRAM  
**DATAMASTER MAINTENANCE REPORT**

RECEIVED 2/19/14 REPORT #6

Complete this report at the time of the regular monthly preventive maintenance check (not to exceed 30 days). Complete this report whenever the instrument is serviced or repaired and whenever it is placed in service. Retain the original and send a copy within 15 days to the Breath Alcohol Program, DHSS.

**REVIEWED**  
 By Carol Day at 10:01 am, Mar 14, 2014

DATAMASTER SN 203043	NAME OF AGENCY Odessa Police Department	DATE OF INSPECTION 02/14/2014
LOCATION OF INSTRUMENT (STREET AND CITY) 101 North Second Street Odessa		TIME OF INSPECTION 11:20 am

**CHECKLIST:** Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

<input checked="" type="checkbox"/> DIAGNOSTIC CHECK (PRINTOUT ATTACHED)	DATE AND TIME (from printout) 11:21am
<input checked="" type="checkbox"/> COMPUTER	<input checked="" type="checkbox"/> DETECTOR
<input checked="" type="checkbox"/> PROGRAM	<input checked="" type="checkbox"/> FILTERS
<input checked="" type="checkbox"/> HEATERS SAMPLE CHAMBER _____ 50 °C	<input checked="" type="checkbox"/> QUARTZ STANDARD
<input checked="" type="checkbox"/> FLOW DETECTOR	<input checked="" type="checkbox"/> CALIBRATION
<input checked="" type="checkbox"/> PUMP HIGH SPEED	<input checked="" type="checkbox"/> PRINTER

INDICATOR LIGHTS

SIMULATOR SOLUTION SUPPLIER Repeco Marketing Inc. LOT # 13001 EXP. DATE 03/07/2015

SIMULATOR TEMP (34°C ± 0.2°C) 34.0 °C SIMULATOR SN SD2765 EXP. DATE 12/10/2014

CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Mark the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

0.100% STANDARD - MUST READ BETWEEN 0.095% AND 0.105% INCLUSIVE  
 0.080% STANDARD - MUST READ BETWEEN 0.076% AND 0.084% INCLUSIVE  
 0.040% STANDARD - MUST READ BETWEEN 0.038% AND 0.042% INCLUSIVE

TEST 1  .101	TEST 2  .101	TEST 3  .102
--------------	--------------	--------------

PERFORM R.F.I. TEST (PRINTOUT ATTACHED)

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS	(0-.04)	1	(.05-.09)	0	(.10-.14)	1	(.15-.19)	0	OVER .19	0
----------	---------	---	-----------	---	-----------	---	-----------	---	----------	---

LIST ANY NEW PARTS AND DESCRIBE ANY ALTERATION OR MODIFICATION THAT WAS MADE TO RESTORE THE INSTRUMENT TO OPERATE SATISFACTORILY AND WITHIN ESTABLISHED LIMITS (USE OTHER SIDE IF NECESSARY).

<b>INSPECTING OFFICER</b>	
SIGNATURE 	PRINT FULL NAME Christopher Bodenhamer
TYPE & PERMIT NUMBER/EXPIRATION DATE 230300 12/11/2015	TELEPHONE NUMBER (816) 721-8236

RETURN COMPLETED REPORT TO THE: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office  
 2875 James Blvd.  
 Poplar Bluff, MO 63901

CERTIFICATE OF ANALYSIS

MANUFACTURER AND SUPPLIER: RepCo Marketing, Inc.

LOT NUMBER: 13001

EXPIRATION DATE: March 7, 2015 at 11:59 p.m.

RepCo Marketing, Inc. certifies the following:

RepCo Marketing, Inc. manufactured, tested and supplied Lot Number 13001 of Alcohol Certified Solution for simulators. Random samples of said lot number were analyzed by an independent laboratory utilizing a gas chromatograph and found to contain .1215 gms/dl +/-0.003 gms/dl wt/vol ethanol (95% Confidence).

The alcohol and distilled water used in the solution were found to be free of any interfering substance.

This solution will produce a vapor alcohol value of .100 +/-3% gms/210L Breath when heated to 34 Degrees Celsius +/-0.2 Degrees Celsius in a simulator (95% Confidence).

The date of manufacture for this lot number is March 8, 2013  
The expiration date for this lot number is March 7, 2015 at  
11:59 p.m.

This document is a true representation of the original Certificate of Analysis.



Cecil B. Garner, President  
RepCo Marketing, Inc.

# BAC DataMaster Evidence Ticket

STATE OF MISSOURI  
ODESSA POLICE DEPARTMENT

BAC DATA MASTER SERIAL NUMBER 883043  
02/14/14  
10121

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY

PROGRAM (04-27-2003)T: OKAY

HEATERS: OKAY

SAMPLE CHAMBER: 500

LOW DETECTOR: OKAY

PUMP: OKAY

HIGH SPEED: OKAY

DETECTOR: OKAY

FILTERS: OKAY

QUARTZ STANDARD: OKAY

CALIBRATION: OKAY

### PRINTER TEST

!"#\$%&'()\*+,-./0123456789:;@<=>?A B C D E F G  
H I J K L M N O P Q R S T U V W X Y Z [ \ ] ^ \_ ` a b c d e f g h i j k l m n o  
p q r s t u v w x y z

Operator Signature

BAC DATA MASTER  
INTERNAL STANDARD  
PERMITS  
PERMITS

TESTING OFFICER:  
BODETH-ROCK CHAMBER  
OFFICER J.D. 589  
PERMIT NUMBER: 25000  
EXPIRATION DATE: 12-31-12  
MISCELLANEOUS DATA:  
N  
N

# BAC DataMaster Evidence Ticket

STATE OF MISSOURI  
ODESSA POLICE DEPARTMENT

BAC DATA MASTER SERIAL NUMBER 883043  
02/14/14

TESTING OFFICER:  
BODETH-ROCK CHAMBER  
OFFICER J.D. 589  
PERMIT NUMBER: 25000  
EXPIRATION DATE: 12-31-12  
MISCELLANEOUS DATA:  
N  
N

--- PERFORM TEST ---

BLANK TEST	1.000	1.183
INTERNAL STANDARD	0.000	1.183
EXTERNAL STANDARD	1.000	1.183
BLANK TEST	1.000	1.183
EXTERNAL STANDARD	1.000	1.183
BLANK TEST	1.000	1.183
EXTERNAL STANDARD	1.000	1.183
BLANK TEST	1.000	1.183
EXTERNAL STANDARD	1.000	1.183

N = 3  
SUM = 3.549  
AVG. = 1.183

# BAC DataMaster Evidence Ticket

STATE OF MISSOURI  
ODESSA POLICE DEPARTMENT

BAC DATA MASTER SERIAL NUMBER 883043  
02/14/14

TESTING OFFICER:  
BODETH-ROCK CHAMBER  
OFFICER J.D. 589  
PERMIT NUMBER: 25000  
EXPIRATION DATE: 12-31-12  
MISCELLANEOUS DATA:  
N  
N

TESTING OFFICER:  
BODETH-ROCK CHAMBER  
OFFICER J.D. 589  
PERMIT NUMBER: 25000  
EXPIRATION DATE: 12-31-12  
MISCELLANEOUS DATA:  
N  
N

3

STATE OF MISSOURI  
DEPARTMENT OF HEALTH AND SENIOR SERVICES  
BREATH ALCOHOL PROGRAM



2

STATE OF MISSOURI  
DEPARTMENT OF HEALTH AND SENIOR SERVICES  
BREATH ALCOHOL PROGRAM



PERMIT  
TYPE II

CHRISTOPHER BODENHAMER

visions of

NY

SERVICES  
LAB-6 (R6-10)

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

DATAMASTER

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 12/11/2013

NUMBER 230300

EXPIRES 12/11/2015

*W. V. ...*

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

*Paul Veeberly*

acting director  
DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES