



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES  
STATE PUBLIC HEALTH LABORATORY  
BREATH ALCOHOL PROGRAM

**DATAMASTER MAINTENANCE REPORT**

**RECEIVED**

REPORT #6

By Carol Day at 10:11 am, Apr 11, 2014

Complete this report at the time of the regular monthly preventive maintenance check (not to exceed 90 days).  
Complete this report whenever the instrument is serviced or repaired and whenever it is placed into service.  
Retain the original and send a copy within 15 days to the Breath Alcohol Program, DHSS.

DATAMASTER SN 203043	NAME OF AGENCY Odessa Police Department	DATE OF INSPECTION 04/06/2014
LOCATION OF INSTRUMENT (STREET AND CITY) 101 North Second Street Odessa		TIME OF INSPECTION 11:10 am

**CHECKLIST:** Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

<input checked="" type="checkbox"/> DIAGNOSTIC CHECK (PRINTOUT ATTACHED)	DATE AND TIME (from printout) 11:11am
<input checked="" type="checkbox"/> COMPUTER	<input checked="" type="checkbox"/> DETECTOR
<input checked="" type="checkbox"/> PROGRAM	<input checked="" type="checkbox"/> FILTERS
<input checked="" type="checkbox"/> HEATERS SAMPLE CHAMBER 49 °C	<input checked="" type="checkbox"/> QUARTZ STANDARD
<input checked="" type="checkbox"/> FLOW DETECTOR	<input checked="" type="checkbox"/> CALIBRATION
<input checked="" type="checkbox"/> PUMP HIGH SPEED	<input checked="" type="checkbox"/> PRINTER

INDICATOR LIGHTS

SIMULATOR SOLUTION SUPPLIER Repc Marketing Inc. LOT # 13002 EXP. DATE 06/19/2015

SIMULATOR TEMP (34°C ± 0.2°C) 34.0 °C SIMULATOR SN SD2765 EXP. DATE 12/10/2014

CALIBRATION CHECK - (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Mark the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

0.100% STANDARD - MUST READ BETWEEN 0.095% AND 0.105% INCLUSIVE

0.080% STANDARD - MUST READ BETWEEN 0.076% AND 0.084% INCLUSIVE

0.040% STANDARD - MUST READ BETWEEN 0.038% AND 0.042% INCLUSIVE

TEST 1  .100	TEST 2  .100	TEST 3  .100
--------------	--------------	--------------

PERFORM R.F.I. TEST (PRINTOUT ATTACHED)

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:  
(DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS	(0-.04) 1	(.05-.09) 0	(.10-.14) 1	(.15-.19) 0	OVER .19 0
----------	-----------	-------------	-------------	-------------	------------

LIST ANY NEW PARTS AND DESCRIBE ANY ALTERATION OR MODIFICATION THAT WAS MADE TO RESTORE THE INSTRUMENT TO OPERATE SATISFACTORILY AND WITHIN ESTABLISHED LIMITS (USE OTHER SIDE IF NECESSARY).

INSPECTING OFFICER	
SIGNATURE 	PRINT FULL NAME Christopher Bodenhamer
TYPE II PERMIT NUMBER/EXPIRATION DATE 230300 12/11/2015	TELEPHONE NUMBER (816) 721-8236

RETURN COMPLETED REPORT TO THE: **Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office  
2875 James Blvd.  
Poplar Bluff, MO 63901**

**CERTIFICATE OF ANALYSIS**

**MANUFACTURER AND SUPPLIER:** RepCo Marketing, Inc.

**LOT NUMBER:** 13002

**EXPIRATION DATE:** June 19, 2015 at 11:59 p.m.

RepCo Marketing, Inc. certifies the following:


RepCo Marketing, Inc. manufactured, tested and supplied Lot Number 13002 of Alcohol Certified Solution for simulators. Random samples of said lot number were analyzed by an independent laboratory utilizing a gas chromatograph and found to contain .1217 gms/dl +/- .003 gms/dl wt/vol ethanol (95% Confidence).

The alcohol and distilled water used in the solution were found to be free of any interfering substance.

This solution will produce a vapor alcohol value of .100 +/- 3% gms/210L Breath when heated to 34 Degrees Celsius +/- 0.2 Degrees Celsius in a simulator (95% Confidence).

The date of manufacture for this lot number is June 20, 2013  
The expiration date for this lot number is June 19, 2015 at 11:59 p.m.

This document is a true representation of the original Certificate of Analysis.



Cecil B. Garner, President  
RepCo Marketing, Inc.

Face This Side Down - This Edge In First

# BAC DataMaster Evidence Ticket

STATE OF MISSOURI  
COOSA POLICE DEPARTMENT

BAC DATAMASTER SERIAL NUMBER 209943  
04-06-10  
11-11

--- OPERATOR NAME ---

OPERATOR:	---
PROGRAM (EA-ET-ET-ET):	DRY
OFFERS:	
SAMPLE CHAMBER:	4.0
FLOW SELECTOR:	DRY
PUMP:	
RICK SPEED:	DRY
DETECTOR:	DRY
FILTERS:	---
QUARTZ TUBE (WT):	---
CALIBRATION:	DRY

Face This Side Down - This Edge In First

# BAC DataMaster Evidence Ticket

STATE OF MISSOURI  
COOSA POLICE DEPARTMENT

BAC DATAMASTER SERIAL NUMBER 209943  
04-06-10

TESTING OFFICER:  
BODENHARPER/CHRIE  
OFFICER I.D.# 589  
PERMIT NUMBER: 209943  
EXPIRATION DATE: 12/11/15  
MISCELLANEOUS DATA:  
N  
N

--- SUPERVISOR NAME ---

BLANK TEST	1.000	11.17
INTERNAL STANDARD	VERIFIED	11.1
EXTERNAL STANDARD	1.000	11.23
BLANK TEST	1.000	11.16
EXTERNAL STANDARD	1.000	11.19
BLANK TEST	1.000	11.17
EXTERNAL STANDARD	1.000	11.21
BLANK TEST	1.000	11.20

N = 9  
STDEV. = .1  
AVG. = 1.1

Face This Side Down - This Edge In First

# BAC DataMaster Evidence Ticket

STATE OF MISSOURI  
COOSA POLICE DEPARTMENT  
BAC DATAMASTER SERIAL NUMBER 209943  
04-06-10

ARREST TIME: 11:30  
SUBJECT NAME:  
RFI/TEST  
DOB: 01/01/1958 SEX: M  
STATE ID: MO-209943  
ARRESTING OFFICER:  
BODENHARPER/CHRIE  
OFFICER I.D.# 589  
TESTING OFFICER:  
BODENHARPER/CHRIE  
OFFICER I.D.# 589  
PERMIT NUMBER: 209943  
EXPIRATION DATE: 12/11/15  
MISCELLANEOUS DATA:

N  
N

--- OPERATOR NAME ---

BLANK TEST 1.000 11.21  
INTERNAL STANDARD VERIFIED 11.21  
RADIO INTERFERENCE

Operator Signature



STATE OF MISSOURI  
DEPARTMENT OF HEALTH AND SENIOR SERVICES  
BREATH ALCOHOL PROGRAM

2

PERMIT  
TYPE II

CHRISTOPHER BODENHAMER

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

DATAMASTER


for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 806.111 through 806.119 RSMo.

DATE 12/11/2013

NUMBER 230300

EXPIRES 12/11/2015

MO 680-0771 (6-10)

  
\_\_\_\_\_  
DIRECTOR OF STATE PUBLIC HEALTH LABORATORY  
  
*Gail Vesterly*  
\_\_\_\_\_  
DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES, acting director

LAB-4 (R6-10)

STATE OF MISSOURI  
DEPARTMENT OF HEALTH AND SENIOR SERVICES  
BREATH ALCOHOL PROGRAM

**INSTRUMENT OPERATOR CARD**

*The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.*



Operator BODENHAMER, CHRISTOPHER  
Permit No. 230300  
Date Issued 12/11/2013 Date Expires 12/11/2015