



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
 STATE PUBLIC HEALTH LABORATORY
 BREATH ALCOHOL PROGRAM
DATAMASTER MAINTENANCE REPORT

RECEIVED
 By Carol Day at 12:20 pm, Apr 07, 2014 REPORT #6

Complete this report at the time of the regular monthly preventive maintenance check (not to exceed 35 days).
 Complete this report whenever the instrument is serviced or repaired and whenever it is placed into service.
 Retain the original and send a copy within 15 days to the Breath Alcohol Program, DHSS.

DATAMASTER SN 203047	NAME OF AGENCY Concordia Police Department	DATE OF INSPECTION 04/07/2014
LOCATION OF INSTRUMENT (STREET AND CITY) Concordia PD, 618 S. Main Street, Concordia Mo		TIME OF INSPECTION 0:07 am

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

<input checked="" type="checkbox"/> DIAGNOSTIC CHECK (PRINTOUT ATTACHED)	DATE AND TIME (from printout) 04/07/14 00:07
<input checked="" type="checkbox"/> COMPUTER	<input checked="" type="checkbox"/> DETECTOR
<input checked="" type="checkbox"/> PROGRAM	<input checked="" type="checkbox"/> FILTERS
<input checked="" type="checkbox"/> HEATERS SAMPLE CHAMBER _____ 49 °C	<input checked="" type="checkbox"/> QUARTZ STANDARD
<input checked="" type="checkbox"/> FLOW DETECTOR	<input checked="" type="checkbox"/> CALIBRATION
<input checked="" type="checkbox"/> PUMP HIGH SPEED	<input checked="" type="checkbox"/> PRINTER

<input checked="" type="checkbox"/> INDICATOR LIGHTS
<input checked="" type="checkbox"/> SIMULATOR SOLUTION SUPPLIER <u>Guth Laboratories</u> LOT # <u>14030</u> EXP. DATE <u>01/20/2016</u>
<input checked="" type="checkbox"/> SIMULATOR TEMP (34°C ± 0.2°C) <u>34</u> °C SIMULATOR SN <u>SD2316</u> EXP. DATE <u>02/07/2015</u>

CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Mark the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

0.100% STANDARD - MUST READ BETWEEN 0.095% AND 0.105% INCLUSIVE
 0.080% STANDARD - MUST READ BETWEEN 0.076% AND 0.084% INCLUSIVE
 0.040% STANDARD - MUST READ BETWEEN 0.038% AND 0.042% INCLUSIVE

TEST 1 <input checked="" type="checkbox"/> .100%	TEST 2 <input checked="" type="checkbox"/> .100%	TEST 3 <input checked="" type="checkbox"/> .100%
--	--	--

PERFORM R.F.I. TEST (PRINTOUT ATTACHED)

**INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:
 (DO NOT INCLUDE SELF-ADMINISTERED TESTS)**

REFUSALS	1	(0-.04)	0	(.05-.09)	1	(.10-.14)	0	(.15-.19)	0	OVER .19	0
----------	---	---------	---	-----------	---	-----------	---	-----------	---	----------	---

LIST ANY NEW PARTS AND DESCRIBE ANY ALTERATION OR MODIFICATION THAT WAS MADE TO RESTORE THE INSTRUMENT TO OPERATE SATISFACTORILY AND WITHIN ESTABLISHED LIMITS (USE OTHER SIDE IF NECESSARY).

Instrument operating with established DOH limits.

INSPECTING OFFICER	
SIGNATURE <i>Anthony L. Helms</i>	PRINT FULL NAME Anthony L. Helms
TYPE II PERMIT NUMBER/EXPIRATION DATE 220204 08/21/2014	TELEPHONE NUMBER (660) 463-2140

RETURN COMPLETED REPORT TO THE: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
 2875 James Blvd.
 Poplar Bluff, MO 63901



GUTH LABORATORIES, INC.

590 NORTH 67th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-564-5470

CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number 14030 of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on January 22, 2014, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain 0.1215% (w/vol) ethyl alcohol. The expiration date for this lot number is January 20, 2016 at 11:59 PM.

When used in a calibrated Simulator, operating at 34°C +/- .2°C, this solution will give a breath alcohol analysis instrument reading of 0.100 g/210L +/- 3%.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President
GUTH LABORATORIES, INC.

NIST Traceability:

Testing was conducted using Cerilliant Reference Standard lot number FN122211-02 whose values are traceable to NIST.

All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.

BAC DataMaster
Evidence Ticket

STATE OF MISSISSIPPI
CONCORDIA POLICE DEPARTMENT
BAC DATA MASTER SERIAL NUMBER 200447
ARREST TIME: 08:00 AM
SUBJECT NAME: [REDACTED]
CONCORDIA POLICE
DOB: 11/12/13 M SEX: M
STATE/D.L. #: MO/123456789
ARRESTING OFFICER: [REDACTED]
OFFICER I.D. #: 200
TESTING OFFICER: [REDACTED]
HELMS
OFFICER I.D. #: 200
PERMIT NUMBER: 220204
EXPIRATION DATE: 06/30/18
MISCELLANEOUS DATA:
ALCOHOL TEST: 0.1
MUNCI TEST: 0.1
INTERNAL STORAGE: [REDACTED]
EXTERNAL STORAGE: [REDACTED]
MISCELLANEOUS DATA: [REDACTED]

OPERATOR SIGNATURE *[Signature]* 708

Card Stock No.
60021

BAC DataMaster
Evidence Ticket

STATE OF MISSISSIPPI
CONCORDIA POLICE DEPARTMENT
BAC DATA MASTER SERIAL NUMBER 200447
ARREST TIME: 08:00 AM
SUBJECT NAME: [REDACTED]
CONCORDIA POLICE
DOB: 11/12/13 M SEX: M
STATE/D.L. #: MO/123456789
ARRESTING OFFICER: [REDACTED]
OFFICER I.D. #: 200
TESTING OFFICER: [REDACTED]
HELMS
OFFICER I.D. #: 200
PERMIT NUMBER: 220204
EXPIRATION DATE: 06/30/18
MISCELLANEOUS DATA:
ALCOHOL TEST: 0.1
MUNCI TEST: 0.1
INTERNAL STORAGE: [REDACTED]
EXTERNAL STORAGE: [REDACTED]
MISCELLANEOUS DATA: [REDACTED]

OPERATOR SIGNATURE *[Signature]*

Card Stock No.
60021

State of Missouri
DEPARTMENT OF HEALTH



P E R M I T
TYPE II



ANTHONY L HELMS

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field repairs, and operate the following breath analyzer(s):

DATAMASTER

for the determination of the alcoholic content of blood from a sample of expired (alveolar) air. Issued under the provisions of sections 577.020 through 577.041, RSMo 1986.

Date 08/21/2012

Number 220204

Expires 08/21/2014

Director of State Public Health Laboratory

Director, Department of Health