



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
 STATE PUBLIC HEALTH LABORATORY
 BREATH ALCOHOL PROGRAM
DATAMASTER MAINTENANCE REPORT

RECEIVED
 By Carol Day at 11:34 am, Feb 05, 2014 #6

Complete this report at the time of the regular monthly preventive maintenance check (not to exceed 35 days).
 Complete this report whenever the instrument is serviced or repaired and whenever it is placed into service.
 Retain the original and send a copy within 15 days to the Breath Alcohol Program, DHSS.

DATAMASTER SN 203053	NAME OF AGENCY Hallsville Police Department	DATE OF INSPECTION 02/04/2014
LOCATION OF INSTRUMENT (STREET AND CITY) 202 Highway 124 East, Hallsville, Missouri 65255		TIME OF INSPECTION 1:02 pm

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

<input checked="" type="checkbox"/> DIAGNOSTIC CHECK (PRINTOUT ATTACHED)	DATE AND TIME (from printout) 02/04/14 13:02
<input checked="" type="checkbox"/> COMPUTER	<input checked="" type="checkbox"/> DETECTOR
<input checked="" type="checkbox"/> PROGRAM	<input checked="" type="checkbox"/> FILTERS
<input checked="" type="checkbox"/> HEATERS SAMPLE CHAMBER _____ 49 °C	<input checked="" type="checkbox"/> QUARTZ STANDARD
<input checked="" type="checkbox"/> FLOW DETECTOR	<input checked="" type="checkbox"/> CALIBRATION
<input checked="" type="checkbox"/> PUMP HIGH SPEED	<input checked="" type="checkbox"/> PRINTER

INDICATOR LIGHTS

SIMULATOR SOLUTION SUPPLIER RepCo Marking, Inc LOT # 13001 EXP. DATE 03/07/2015

SIMULATOR TEMP (34°C ± 0.2°C) 34 °C SIMULATOR SN 093752 EXP. DATE 07/23/2014

CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Mark the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

0.100% STANDARD - MUST READ BETWEEN 0.095% AND 0.105% INCLUSIVE
 0.080% STANDARD - MUST READ BETWEEN 0.076% AND 0.084% INCLUSIVE
 0.040% STANDARD - MUST READ BETWEEN 0.038% AND 0.042% INCLUSIVE

TEST 1 .097	TEST 2 .099	TEST 3 .100
--------------	--------------	--------------

PERFORM R.F.I. TEST (PRINTOUT ATTACHED)

**INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:
 (DO NOT INCLUDE SELF-ADMINISTERED TESTS)**

REFUSALS	1	(0-.04)	0	(.05-.09)	0	(.10-.14)	1	(.15-.19)	0	OVER .19	0
----------	---	---------	---	-----------	---	-----------	---	-----------	---	----------	---

LIST ANY NEW PARTS AND DESCRIBE ANY ALTERATION OR MODIFICATION THAT WAS MADE TO RESTORE THE INSTRUMENT TO OPERATE SATISFACTORILY AND WITHIN ESTABLISHED LIMITS (USE OTHER SIDE IF NECESSARY).

Instrument is operating satisfactory.

INSPECTING OFFICER	
SIGNATURE 	PRINT FULL NAME Edward (Tony) Fields
TYPE II PERMIT NUMBER/EXPIRATION DATE 220036 02/09/2014	TELEPHONE NUMBER (573) 696-3838

RETURN COMPLETED REPORT TO THE: **Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
 2875 James Blvd.
 Poplar Bluff, MO 63901**

BAC DataMaster Evidence Ticket

STATE OF MISSOURI
HILLSVILLE POLICE DEPARTMENT

BAC DATA MASTER SERIAL NUMBER 203053

02/04/14
13:02

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY

PROGRAM (04-07-2009): OKAY

HEATERS
SAMPLE CHAMBER: 49C

FLOW DETECTOR: OKAY

PUMP
HIGH SPEED: OKAY

DETECTOR: OKAY

FILTERS: OKAY

QUARTZ STANDARD: OKAY

CALIBRATION: OKAY

PRINTER TEST

!@#%&'()*+,-./0123456789:;<=>?@ABCDEF G
HIJKLMNOPQRSTUVWXYZ[\]^_`abcdefg hijklmno
pqrs tuvwxyz{|}~

Operator Signature  2208-02

BAC DataMaster Evidence Ticket

STATE OF MISSOURI
HILLSVILLE POLICE DEPARTMENT

BAC DATA MASTER SERIAL NUMBER 203053

02/04/14

TESTING OFFICER:
FIELDS/EDWARD

OFFICER I.D.: 9014

PERMIT NUMBER: 220036

EXPIRATION DATE: 02/09/14


MISCELLANEOUS DATA:

--- SUPERVISOR MODE ---

BLANK TEST
INTERNAL STANDARD
EXTERNAL STANDARD
BLANK TEST
EXTERNAL STANDARD
INTERNAL STANDARD
BLANK TEST

.000 VERIFIED 13:07
.097 13:08
.000 13:08
.099 13:08
.000 13:10
.100 13:10
.000 13:11

N = 3
SIM. = .1
RW9. = .0986

Operator Signature  2208-02

BAC DataMaster Evidence Ticket

STATE OF MISSOURI
HILLSVILLE POLICE DEPARTMENT

BAC DATA MASTER SERIAL NUMBER 203053

02/04/14

ARREST TIME: 12:50

SUBJECT NAME:
TEST/TEST/T

DOB: 06/06/66 SEX: F

STATE/D.L.: MO/12345

ARRESTING OFFICER:
FIELDS/EDWARD

OFFICER I.D.: 9014

TESTING OFFICER:
FIELDS/EDWARD

OFFICER I.D.: 9014

PERMIT NUMBER: 220036

EXPIRATION DATE: 02/09/14

MISCELLANEOUS DATA:

--- BREATH ANALYSIS ---

BLANK TEST
INTERNAL STANDARD
RADIO INTERFERENCE

.000 VERIFIED 13:13

Operator Signature  2208-02

CERTIFICATE OF ANALYSIS

MANUFACTURER AND SUPPLIER: RepCo Marketing, Inc.
LOT NUMBER: 13001
EXPIRATION DATE: March 7, 2015 at 11:59 p.m.

RepCo Marketing, Inc. certifies the following:

RepCo Marketing, Inc. manufactured, tested and supplied Lot Number 13001 of Alcohol Certified Solution for simulators. Random samples of said lot number were analyzed by an independent laboratory utilizing a gas chromatograph and found to contain .1215 gms/dl +/- .003 gms/dl wt/vol ethanol (95% Confidence).

The alcohol and distilled water used in the solution were found to be free of any interfering substance.

This solution will produce a vapor alcohol value of .100 +/- 3% gms/210L Breath when heated to 34 Degree Celsius +/- 0.2 Degree Celsius in a simulator (95% Confidence).

The date of manufacture for this lot number is March 8, 2013
The expiration date for this lot number is March 7, 2015 at
11:59 p.m.

This document is a true representation of the original Certificate of Analysis.



Cecil B. Garner, President
RepCo Marketing, Inc.



P E R M I T
TYPE II



EDWARD FIELDS

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field repairs, and operate the following breath analyzer(s):

DATANASTER

for the determination of the alcoholic content of blood from a sample of expired (alveolar) air, issued under the provisions of sections 577.020 through 577.041, RSMo 1993.

Date 02/09/2012

Number 220036

Expires 02/09/2014

MO 690-0771 (2-93)

Director of State Public Health Laboratory

Director, Department of Health

Lab. 4 (17-55)