



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
 STATE PUBLIC HEALTH LABORATORY
 BREATH ALCOHOL PROGRAM
DATAMASTER MAINTENANCE REPORT

RECEIVED REPORT #6
 By Carol Day at 9:12 am, Jun 04, 2014

Complete this report at the time of the regular monthly preventive maintenance check (not to be completed more than once every 30 days).
 Complete this report whenever the instrument is serviced or repaired and whenever it is placed into service.
 Retain the original and send a copy within 15 days to the Breath Alcohol Program, DHSS.

DATAMASTER SN 204010	NAME OF AGENCY Lincoln County Sheriff's Office	DATE OF INSPECTION 06/01/2014
LOCATION OF INSTRUMENT (STREET AND CITY) 65 Business Park Drive, Troy, Missouri 63379		TIME OF INSPECTION 6:05 am

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

<input checked="" type="checkbox"/> DIAGNOSTIC CHECK (PRINTOUT ATTACHED)	DATE AND TIME (from printout) 06/01/2014 06:05 AM
<input checked="" type="checkbox"/> COMPUTER	<input checked="" type="checkbox"/> DETECTOR
<input checked="" type="checkbox"/> PROGRAM	<input checked="" type="checkbox"/> FILTERS
<input checked="" type="checkbox"/> HEATERS SAMPLE CHAMBER _____ 49 °C	<input checked="" type="checkbox"/> QUARTZ STANDARD
<input checked="" type="checkbox"/> FLOW DETECTOR	<input checked="" type="checkbox"/> CALIBRATION
<input checked="" type="checkbox"/> PUMP HIGH SPEED	<input checked="" type="checkbox"/> PRINTER

INDICATOR LIGHTS

SIMULATOR SOLUTION SUPPLIER Guth Labs (Bottle #805) LOT # 14030 EXP. DATE 01/20/2016

SIMULATOR TEMP (34°C ± 0.2°C) 34 °C SIMULATOR SN SD2508 EXP. DATE 03/18/2015

CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Mark the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

0.100% STANDARD - MUST READ BETWEEN 0.095% AND 0.105% INCLUSIVE
 0.080% STANDARD - MUST READ BETWEEN 0.076% AND 0.084% INCLUSIVE
 0.040% STANDARD - MUST READ BETWEEN 0.038% AND 0.042% INCLUSIVE

TEST 1 ➔ .098	TEST 2 ➔ .099	TEST 3 ➔ .100
---------------	---------------	---------------

PERFORM R.F.I. TEST (PRINTOUT ATTACHED)

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS	0	(0-.04)	0	(.05-.09)	0	(.10-.14)	0	(.15-.19)	0	OVER .19	0
----------	---	---------	---	-----------	---	-----------	---	-----------	---	----------	---

LIST ANY NEW PARTS AND DESCRIBE ANY ALTERATION OR MODIFICATION THAT WAS MADE TO RESTORE THE INSTRUMENT TO OPERATE SATISFACTORILY AND WITHIN ESTABLISHED LIMITS (USE OTHER SIDE IF NECESSARY).

N/A

INSPECTING OFFICER	
SIGNATURE 	PRINT FULL NAME Sgt. Michael K. Pirtle
TYPE II PERMIT NUMBER/EXPIRATION DATE 230232 10/17/2015	TELEPHONE NUMBER (636) 528-8546

RETURN COMPLETED REPORT TO THE: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
 2875 James Blvd.
 Poplar Bluff, MO 63901

Face This Side Down - This Edge In First

**BAC DataMaster
Evidence Ticket**

STATE OF MISSOURI
LINCOLN COUNTY SHERIFF DEPARTMENT

BAC DATAMASTER SERIAL NUMBER 204010
06/01/14
06:05


--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY
PROGRAM (04-07-2009): OKAY
HEATERS
SAMPLE CHAMBER: 49c
FLOW DETECTOR: OKAY
PUMP
HIGH SPEED: OKAY
DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY

PRINTER TEST

!"#\$%&'()*+,-./0123456789:;<=>?@ABCDEFGH
IJKLMNOPQRSTUVWXYZ[\]^_`abcdefghijklmnop
qrstuvwxyz{|}~

Operator Signature



Face This Side Down - This Edge In First

**BAC DataMaster
Evidence Ticket**

STATE OF MISSOURI
LINCOLN COUNTY SHERIFF DEPARTMENT

BAC DATAMASTER SERIAL NUMBER 204010
06/01/14

TESTING OFFICER:

PIRTLE/M/K
OFFICER I.D.: 490
PERMIT NUMBER: 230232
EXPIRATION DATE: 10/17/15
MISCELLANEOUS DATA:

--- SUPERVISOR MODE ---

BLANK TEST	.000	06:07
INTERNAL STANDARD	VERIFIED	06:07
EXTERNAL STANDARD	.098	06:08
BLANK TEST	.000	06:09
EXTERNAL STANDARD	.099	06:09
BLANK TEST	.000	06:10
EXTERNAL STANDARD	.100	06:10
BLANK TEST	.000	06:11

N = 3
SIM. = .1
AVG. = .099

Operator Signature



Face This Side Down - This Edge In First

**BAC DataMaster
Evidence Ticket**

STATE OF MISSOURI
LINCOLN COUNTY SHERIFF DEPARTMENT

BAC DATAMASTER SERIAL NUMBER 204010
06/01/14

ARREST TIME: 05:50
SUBJECT NAME:
N/A
DOB: 01/01/81 SEX: M
STATE/D.L.: MO/1234
ARRESTING OFFICER:
N/A
OFFICER I.D.:
TESTING OFFICER:
PIRTLE/M/K
OFFICER I.D.: 490
PERMIT NUMBER: 230232
EXPIRATION DATE: 10/17/15
MISCELLANEOUS DATA:
BLANK TEST

--- BREATH ANALYSIS ---

BLANK TEST	.000	06:18
INTERNAL STANDARD	VERIFIED	06:18
SUBJECT SAMPLE	.000	06:18
BLANK TEST	.000	06:19

Operator Signature



Face This Side Down - This Edge In First

**BAC DataMaster
Evidence Ticket**

STATE OF MISSOURI
LINCOLN COUNTY SHERIFF DEPARTMENT


BAC DATAMASTER SERIAL NUMBER 204010
06/01/14

ARREST TIME: 05:50
SUBJECT NAME:
N/A
DOB: 01/01/81 SEX: M
STATE/D.L.: MO/1234
ARRESTING OFFICER:
N/A
OFFICER I.D.:
TESTING OFFICER:
PIRTLE/M/K
OFFICER I.D.: 490
PERMIT NUMBER: 230232
EXPIRATION DATE: 10/17/15
MISCELLANEOUS DATA:
RFI TEST

--- BREATH ANALYSIS ---

BLANK TEST	.000	06:16
INTERNAL STANDARD	VERIFIED	06:16
RADIO INTERFERENCE		

Operator Signature





STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

2

PERMIT
TYPE II
MICHAEL K PIRTLE

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

DATAMASTER

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 10/17/2013

NUMBER 230232

EXPIRES 10/17/2015

MO 580-0771 (6-10)

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

,acting director

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

LAB-4 (R6-10)



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.



Operator PIRTLE, MICHAEL
Permit No 230232
Date Issued 10/17/2013 Date Expires 10/17/2015