



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 STATE PUBLIC HEALTH LABORATORY  
 BREATH ALCOHOL PROGRAM  
 DATAMASTER MAINTENANCE REPORT

**RECEIVED** REPORT #6  
 By Carol Day at 1:30 pm, Jul 08, 2014

Complete this report at the time of the regular monthly preventive maintenance check (not to exceed 35 days).  
 Complete this report whenever the instrument is serviced or repaired and whenever it is placed into service.  
 Retain the original and send a copy within 15 days to the Breath Alcohol Program, DHSS.

|  |   |                                  |
|--|---|----------------------------------|
| DATAMASTER SN<br>204061  | NAME OF AGENCY<br>Missouri State Highway Patrol | DATE OF INSPECTION<br>07/07/2014 |
| LOCATION OF INSTRUMENT (STREET AND CITY)<br>710 South Kingshighway, Perryville, MO 63775 |   | TIME OF INSPECTION<br>12:50 pm   |

**CHECKLIST:** Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

|   |   |
|---|---|
| <input checked="" type="checkbox"/> DIAGNOSTIC CHECK (PRINTOUT ATTACHED)  | DATE AND TIME (from printout) 07/07/2014 12:50 p.m. |
| <input checked="" type="checkbox"/> COMPUTER  | <input checked="" type="checkbox"/> DETECTOR        |
| <input checked="" type="checkbox"/> PROGRAM   | <input checked="" type="checkbox"/> FILTERS         |
| <input checked="" type="checkbox"/> HEATERS SAMPLE CHAMBER 50°C   | <input checked="" type="checkbox"/> QUARTZ STANDARD |
| <input checked="" type="checkbox"/> FLOW DETECTOR   | <input checked="" type="checkbox"/> CALIBRATION     |
| <input checked="" type="checkbox"/> PUMP HIGH SPEED   | <input checked="" type="checkbox"/> PRINTER         |
| <input checked="" type="checkbox"/> INDICATOR LIGHTS  |   |
| <input checked="" type="checkbox"/> SIMULATOR SOLUTION SUPPLIER RepCo Marketing Inc. LOT # 13001 EXP. DATE 03/07/2015 |   |
| <input checked="" type="checkbox"/> SIMULATOR TEMP (34°C ± 0.2°C) 34.2 °C SIMULATOR SN G6828 EXP. DATE 07/09/2014     |   |

CALIBRATION CHECK - (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Mark the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

0.100% STANDARD - MUST READ BETWEEN 0.095% AND 0.105% INCLUSIVE  
 0.080% STANDARD - MUST READ BETWEEN 0.076% AND 0.084% INCLUSIVE  
 0.040% STANDARD - MUST READ BETWEEN 0.038% AND 0.042% INCLUSIVE

|             |             |             |
|-------------|-------------|-------------|
| TEST 1 .103 | TEST 2 .104 | TEST 3 .104 |
|-------------|-------------|-------------|

PERFORM R.F.I. TEST (PRINTOUT ATTACHED)

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:  
 (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

|            |           |             |             |             |            |
|------------|-----------|-------------|-------------|-------------|------------|
| REFUSALS 0 | (0-.04) 0 | (.05-.09) 1 | (.10-.14) 0 | (.15-.19) 3 | OVER .19 1 |
|------------|-----------|-------------|-------------|-------------|------------|

LIST ANY NEW PARTS AND DESCRIBE ANY ALTERATION OR MODIFICATION THAT WAS MADE TO RESTORE THE INSTRUMENT TO OPERATE SATISFACTORILY AND WITHIN ESTABLISHED LIMITS (USE OTHER SIDE IF NECESSARY).

Instrument operating within MODHSS specifications. Solution Bottle #0461 Datamaster Inventory #127191

**INSPECTING OFFICER**

|  |  |
|--|--|
| SIGNATURE<br>  | PRINT FULL NAME<br>Tpr. James R. Sauer   |
| TYPE II PERMIT NUMBER EXPIRATION DATE<br>240070 03/04/2016 | TELEPHONE NUMBER<br>(636) 300-2800   |
| RETURN COMPLETED REPORT TO THE:                            | Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office<br>2875 James Blvd.<br>Poplar Bluff, MO 63901 |

CERTIFICATE OF ANALYSIS

MANUFACTURER AND SUPPLIER: RepCo Marketing, Inc.  
LOT NUMBER: 13001  
EXPIRATION DATE: March 7, 2015 at 11:59 p.m.

RepCo Marketing, Inc. certifies the following:


RepCo Marketing, Inc. manufactured, tested and supplied Lot Number 13001 of Alcohol Certified Solution for simulators. Random samples of said lot number were analyzed by an independent laboratory utilizing a gas chromatograph and found to contain .1215 gms/dl +/- .003 gms/dl wt/vol ethanol (95% Confidence).

The alcohol and distilled water used in the solution were found to be free of any interfering substance.

This solution will produce a vapor alcohol value of .100 +/- 3% gms/210L Breath when heated to 34 Degrees Celsius +/- 0.2 Degrees Celsius in a simulator (95% Confidence).

The date of manufacture for this lot number is March 8, 2013  
The expiration date for this lot number is March 7, 2015 at 11:59 p.m.

This document is a true representation of the original Certificate of Analysis.



Cecil B. Garner, President  
RepCo Marketing, Inc.

Face This Side Down - This Edge In First

**BAC DataMaster**  
Evidence Ticket

MISSOURI STATE HIGHWAY PATROL  
BAC DATA MASTER SERIAL NUMBER 204061  
07/07/14

TESTING OFFICER:  
SNUER/JAMES/R  
OFFICER I.D.: 745  
PERMIT NUMBER: 240078  
EXPIRATION DATE: 03/07/16  
MISCELLANEOUS DATA:

--- SUPERVISOR MODE ---

|                   |      |          |       |
|-------------------|------|----------|-------|
| BLANK TEST        | .000 | VERIFIED | 13:15 |
| INTERNAL STANDARD | .183 | VERIFIED | 13:15 |
| EXTERNAL STANDARD | .800 | VERIFIED | 13:16 |
| BLANK TEST        | .194 | VERIFIED | 13:16 |
| EXTERNAL STANDARD | .000 | VERIFIED | 13:17 |
| INTERNAL STANDARD | .104 | VERIFIED | 13:18 |
| BLANK TEST        | .000 | VERIFIED | 13:18 |

N = 3  
STN. = .1  
AVG. = .1036

Operator Signature   
220640

Face This Side Down - This Edge In First

**BAC DataMaster**  
Evidence Ticket

MISSOURI STATE HIGHWAY PATROL  
BAC DATA MASTER SERIAL NUMBER 204061  
07/07/14

ARREST TIME: 11:11  
SUBJECT NAME:  
DOE/JON  
DOB: 11/11/11 SEX: M  
STATE/D.L.: MO/12345  
ARRESTING OFFICER:  
SNUER/JAMES/R  
OFFICER I.D.: 745  
TESTING OFFICER:  
SAME

OFFICER I.D.: 745  
PERMIT NUMBER: 240078  
EXPIRATION DATE: 03/07/16  
MISCELLANEOUS DATA:

--- BREATH ANALYSIS ---

|                    |      |          |       |
|--------------------|------|----------|-------|
| BLANK TEST         | .000 | VERIFIED | 13:22 |
| INTERNAL STANDARD  |      |          |       |
| RADIO INTERFERENCE |      |          |       |

Operator Signature   
220642

Face This Side Down - This Edge In First

**BAC DataMaster**  
Evidence Ticket

MISSOURI STATE HIGHWAY PATROL  
BAC DATA MASTER SERIAL NUMBER 204061  
07/07/14  
12:50

--- DIAGNOSTIC CHECK ---

|                       |      |
|-----------------------|------|
| COMPUTER:             | OKAY |
| PROGRAM (04-07-2009): | OKAY |
| HEATERS               | OKAY |
| SAMPLE CHAMBER:       | OKAY |
| FLOW DETECTOR:        | OKAY |
| PUMP                  | OKAY |
| HIGH SPEED:           | OKAY |
| DETECTOR1             | OKAY |
| FILTERS:              | OKAY |
| QUARTZ STANDARD:      | OKAY |
| CALIBRATION:          | OKAY |

PRINTER TEST  
!#%&'()\*+,-./0123456789:;<=>?@ABCDEFGHI  
HI,KL,MNOPQRSTUVWXYZ[\]^\_`abodefgghijklnmo  
pqrs,tuvwxyz{|}~"

Operator Signature   
220642



STATE OF MISSOURI  
DEPARTMENT OF HEALTH AND SENIOR SERVICES  
BREATH ALCOHOL PROGRAM

2

**PERMIT**  
**TYPE II**  
**JAMES R SAUER**

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

**DATAMASTER, INTOX DMT**

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 3/7/2014

NUMBER 240070

EXPIRES 3/7/2016

MO 680-0771 (5-10)

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

LAB-4 (96-10)

 STATE OF MISSOURI  
DEPARTMENT OF HEALTH AND SENIOR SERVICES  
BREATH ALCOHOL PROGRAM

**INSTRUMENT OPERATOR CARD**

*The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.*



Operator SAUER, JAMES  
Permit No 240070  
Date Issued 3/7/2014 Date Expires 3/7/2016