



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
 STATE PUBLIC HEALTH LABORATORY
 BREATH ALCOHOL PROGRAM
 DATAMASTER MAINTENANCE REPORT

RECEIVED REPORT #6
 By Carol Day at 12:32 pm, Apr 08, 2014

Complete this report at the time of the regular monthly preventive maintenance check (not to exceed 35 days).
 Complete this report whenever the instrument is serviced or repaired and whenever it is placed into service.
 Retain the original and send a copy within 15 days to the Breath Alcohol Program, DHSS.

DATAMASTER SN 204062	NAME OF AGENCY MSP	DATE OF INSPECTION 4-1-14
LOCATION OF INSTRUMENT (STREET AND CITY) CRAWFORD CO SHERIFF'S DEPT		TIME OF INSPECTION 1546

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

<input checked="" type="checkbox"/> DIAGNOSTIC CHECK (PRINTOUT ATTACHED)	DATE AND TIME (from printout) 4-1-14 1546
<input checked="" type="checkbox"/> COMPUTER	<input checked="" type="checkbox"/> DETECTOR
<input checked="" type="checkbox"/> PROGRAM	<input checked="" type="checkbox"/> FILTERS
<input checked="" type="checkbox"/> HEATERS SAMPLE CHAMBER 50 °C	<input checked="" type="checkbox"/> QUARTZ STANDARD
<input checked="" type="checkbox"/> FLOW DETECTOR	<input checked="" type="checkbox"/> CALIBRATION
<input checked="" type="checkbox"/> PUMP HIGH SPEED	<input checked="" type="checkbox"/> PRINTER

<input checked="" type="checkbox"/> INDICATOR LIGHTS
<input checked="" type="checkbox"/> SIMULATOR SOLUTION SUPPLIER PEPCO MARKETING LOT # 13001 EXP. DATE 3-7-15
<input checked="" type="checkbox"/> SIMULATOR TEMP (34°C ± 0.2°C) 34 °C SIMULATOR SN 66820 EXP. DATE 02-14-15

CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Mark the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

<input checked="" type="checkbox"/> 0.100% STANDARD - MUST READ BETWEEN 0.095% AND 0.105% INCLUSIVE
<input type="checkbox"/> 0.080% STANDARD - MUST READ BETWEEN 0.076% AND 0.084% INCLUSIVE
<input type="checkbox"/> 0.040% STANDARD - MUST READ BETWEEN 0.038% AND 0.042% INCLUSIVE

TEST 1 • .101	TEST 2 • .103	TEST 3 • .104
----------------------	----------------------	----------------------

PERFORM R.F.I. TEST (PRINTOUT ATTACHED)

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:
 (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS 0	(0-.04) 0	(.05-.09) 0	(.10-.14) 0	(.15-.19) 1	OVER .19 0
-------------------	------------------	--------------------	--------------------	--------------------	-------------------

LIST ANY NEW PARTS AND DESCRIBE ANY ALTERATION OR MODIFICATION THAT WAS MADE TO RESTORE THE INSTRUMENT TO OPERATE SATISFACTORILY AND WITHIN ESTABLISHED LIMITS (USE OTHER SIDE IF NECESSARY).

INSPECTING OFFICER	
SIGNATURE <i>[Signature]</i>	PRINT FULL NAME COL. D. DAVIS
TYPE II PERMIT NUMBER/EXPIRATION DATE 220375 10-31-14	TELEPHONE NUMBER 573-368-2395

RETURN COMPLETED REPORT TO THE: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
 2875 James Blvd.
 Poplar Bluff, MO 63901

CERTIFICATE OF ANALYSIS

MANUFACTURER AND SUPPLIER: RepCo Marketing, Inc.

LOT NUMBER: 13001

EXPIRATION DATE: March 7, 2015 at 11:59 p.m.

RepCo Marketing, Inc. certifies the following:

RepCo Marketing, Inc. manufactured, tested and supplied Lot Number 13001 of Alcohol Certified Solution for simulators. Random samples of said lot number were analyzed by an independent laboratory utilizing a gas chromatograph and found to contain .1215 gms/dl +/-0.003 gms/dl wt/vol ethanol (95% Confidence).

The alcohol and distilled water used in the solution were found to be free of any interferring substance.

This solution will produce a vapor alcohol value of .100 +/-3% gms/210L Breath when heated to 34 Degrees Celsius +/-0.2 Degrees Celsius in a simulator (95% Confidence).

The date of manufacture for this lot number is March 8, 2013
The expiration date for this lot number is March 7, 2015 at
11:59 p.m.

This document is a true representation of the original Certificate of Analysis.



Cecil B. Garner, President
RepCo Marketing, Inc.

STATE OF MISSOURI
DEPARTMENT OF HEALTH



P E R M I T
TYPE II



DANIAL E DICUS

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field repairs, and operate the following breath analyzer(s):

DATAMASTER

for the determination of the alcohol content of blood from a sample of expired (alveolar) air. Issued under the provisions of sections 577.020 through 577.041, RSMo 1986.

Date 10/31/2012

Number 220375

Expires 10/31/2014

Director of State Public Health Laboratory

Director, Department of Health

Face This Side Down - This Edge In First

BAC DataMaster Evidence Ticket


MISSOURI STATE HIGHWAY PATROL
BAC DATAMASTER SERIAL NUMBER 204062
04/01/14

ARREST TIME: 13:00
SUBJECT NAME:
DOE/JON
DOB: 02/13/56 SEX: M
STATE-D.L.L.# MO/8876776
ARRESTING OFFICER:
DICUS/DANIAL/E
OFFICER I.D.# 441
TESTING OFFICER:
SAME
OFFICER I.D.# 441
PERMIT NUMBER: 220375
EXPIRATION DATE: 10/31/14
MISCELLANEOUS DATA:
RADIO TEST

--- BREATH ANALYSIS ---

RADIO INTERFERENCE

Operator Signature



Face This Side Down - This Edge In First

BAC DataMaster Evidence Ticket

MISSOURI STATE HIGHWAY PATROL
BAC DATAMASTER SERIAL NUMBER 204062
04/01/14

TESTING OFFICER:
DICUS/DANIAL/E
OFFICER I.D.# 441
PERMIT NUMBER: 220375
EXPIRATION DATE: 10/31/14
MISCELLANEOUS DATA:
TEST 10 SOLUTION

--- SUPERVISOR MODE ---

BLANK TEST	.000	16:10
INTERNAL STANDARD	VERIFIED	16:10
EXTERNAL STANDARD	.101	16:11
BLANK TEST	.000	16:12
EXTERNAL STANDARD	.100	16:12
BLANK TEST	.000	16:13
EXTERNAL STANDARD	.104	16:13
BLANK TEST	.000	16:14

1 = 0
SPL = .1
AVG. = .1026

Operator Signature

