



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
 STATE PUBLIC HEALTH LABORATORY
 BREATH ALCOHOL PROGRAM
DATAMASTER MAINTENANCE REPORT

RECEIVED
 By Carol Day at 10:19 am, Jan 23, 2014

REPORT #6

Complete this report at the time of the regular monthly preventive maintenance check (not to exceed 35 days).
 Complete this report whenever the instrument is serviced or repaired and whenever it is placed into service.
 Retain the original and send a copy within 15 days to the Breath Alcohol Program, DHSS.

DATAMASTER SN 204064	NAME OF AGENCY Texas Co. SO, Houston	DATE OF INSPECTION 01/17/2014
LOCATION OF INSTRUMENT (STREET AND CITY) 2920 North Shamrock Road, Jefferson City		TIME OF INSPECTION 10:45

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

<input checked="" type="checkbox"/> DIAGNOSTIC CHECK (PRINTOUT ATTACHED)	DATE AND TIME (from printout) <u>01/17/2014 @ 10:45</u>
<input checked="" type="checkbox"/> COMPUTER	<input checked="" type="checkbox"/> DETECTOR
<input checked="" type="checkbox"/> PROGRAM	<input checked="" type="checkbox"/> FILTERS
<input checked="" type="checkbox"/> HEATERS SAMPLE CHAMBER <u>+49°C</u>	<input checked="" type="checkbox"/> QUARTZ STANDARD
<input checked="" type="checkbox"/> FLOW DETECTOR	<input checked="" type="checkbox"/> CALIBRATION
<input checked="" type="checkbox"/> PUMP HIGH SPEED	<input checked="" type="checkbox"/> PRINTER
<input checked="" type="checkbox"/> INDICATOR LIGHTS	
<input checked="" type="checkbox"/> SIMULATOR SOLUTION SUPPLIER <u>RepCo Marketing</u>	LOT # <u>13001</u> EXP. DATE <u>03/07/2015</u>
<input checked="" type="checkbox"/> SIMULATOR TEMP (34°C ± 0.2°C) <u>+34.00°C</u>	SIMULATOR SN <u>DR5375</u> EXP. DATE <u>04/15/2014</u>

CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Mark the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

0.100% STANDARD - MUST READ BETWEEN 0.095% AND 0.105% INCLUSIVE
 0.080% STANDARD - MUST READ BETWEEN 0.076% AND 0.084% INCLUSIVE
 0.040% STANDARD - MUST READ BETWEEN 0.038% AND 0.042% INCLUSIVE

TEST 1 ← <u>.100</u>	TEST 2 ← <u>.100</u>	TEST 3 ← <u>.100</u>
----------------------	----------------------	----------------------

PERFORM R.F.I. TEST (PRINTOUT ATTACHED)

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS	0	(0-.04)	0	(.05-.09)	0	(.10-.14)	0	(.15-.19)	0	OVER .19	0
----------	---	---------	---	-----------	---	-----------	---	-----------	---	----------	---

LIST ANY NEW PARTS AND DESCRIBE ANY ALTERATION OR MODIFICATION THAT WAS MADE TO RESTORE THE INSTRUMENT TO OPERATE SATISFACTORILY AND WITHIN ESTABLISHED LIMITS (USE OTHER SIDE IF NECESSARY).
 Replaced and adjusted printer head for optimal evidence ticket impact.

INSPECTING OFFICER	
SIGNATURE <i>Dewayne D. Carver</i>	PRINT FULL NAME Dewayne D. Carver
TYPE II PERMIT NUMBER EXPIRATION DATE 230101 05/28/2015	TELEPHONE NUMBER (573) 751-4722

RETURN COMPLETED REPORT TO THE: **Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
 2875 James Blvd.
 Poplar Bluff, MO 63901**

CERTIFICATE OF ANALYSIS

MANUFACTURER AND SUPPLIER: RepCo Marketing, Inc.

LOT NUMBER: 13001

EXPIRATION DATE: March 7, 2015 at 11:59 p.m.

RepCo Marketing, Inc. certifies the following:

RepCo Marketing, Inc. manufactured, tested and supplied Lot Number 13001 of Alcohol Certified Solution for simulators. Random samples of said lot number were analyzed by an independent laboratory utilizing a gas chromatograph and found to contain .1215 gms/dl +/-0.003 gms/dl wt/vol ethanol (95% Confidence).

The alcohol and distilled water used in the solution were found to be free of any interfering substance.

This solution will produce a vapor alcohol value of .100 +/-3% gms/210L Breath when heated to 34 Degrees Celsius +/-0.2 Degrees Celsius in a simulator (95% Confidence).

The date of manufacture for this lot number is March 8, 2013
The expiration date for this lot number is March 7, 2015 at
11:59 p.m.

This document is a true representation of the original Certificate of Analysis.



Cecil B. Garner, President
RepCo Marketing, Inc.

FACE THIS SIDE DOWN - THIS EDGE IN FIRST

**BAC DataMaster
Evidence Ticket**

MISSOURI STATE HIGHWAY PATROL
BAC DATAMASTER SERIAL NUMBER 204064
01/17/14
10:45

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY
PROGRAM (04-07-2009): OKAY
HEATERS
SAMPLE CHAMBER: 49c
FLOW DETECTOR: OKAY
PUMP
HIGH SPEED: OKAY
DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY

PRINTER TEST
!"##&'()*+,-./0123456789:;<=>?@ABCDEF0
HIJKLMNPQRSTUWXYZ[\]^_`abcdefgijklmno
pqrstuvwxyz{|}~

OPERATOR SIGNATURE

Card Stock No.
60021

REORDER ALL SUPPLIES FROM N.P.A.S.
P.O. BOX 1436 MANSFIELD, OH 44901

FACE THIS SIDE DOWN - THIS EDGE IN FIRST

**BAC DataMaster
Evidence Ticket**

MISSOURI STATE HIGHWAY PATROL
BAC DATAMASTER SERIAL NUMBER 204064
01/17/14

TESTING OFFICER:
CARVER/DEWAYNE/D
OFFICER I.D.: C41
PERMIT NUMBER: 230101
EXPIRATION DATE: 05/28/15
MISCELLANEOUS DATA:
.100 VAPOR ACCURACY CHECK
REPCO MKTG LOT 13001 EXP 3/7/2015

--- SUPERVISOR MODE ---

BLANK TEST	.000	10:47
INTERNAL STANDARD	VERIFIED	10:47
EXTERNAL STANDARD	.100	10:48
BLANK TEST	.000	10:48
EXTERNAL STANDARD	.100	10:49
BLANK TEST	.000	10:49
EXTERNAL STANDARD	.100	10:50
BLANK TEST	.000	10:50

N = 3
SIM. = .1
AVG. = .1

FACE THIS SIDE DOWN - THIS EDGE IN FIRST

**BAC DataMaster
Evidence Ticket**

MISSOURI STATE HIGHWAY PATROL
BAC DATAMASTER SERIAL NUMBER 204064
01/17/14


ARREST TIME: 00:00
SUBJECT NAME:
X
DOB: 01/01/01 SEX: M
STATE/D.L.: XX/X
ARRESTING OFFICER:
X
OFFICER I.D.: X
TESTING OFFICER:
CARVER/DEWAYNE/D
OFFICER I.D.: C41
PERMIT NUMBER: 230101
EXPIRATION DATE: 05/28/15
MISCELLANEOUS DATA:
RFI TEST

--- BREATH ANALYSIS ---

BLANK TEST	.000	10:55
INTERNAL STANDARD	VERIFIED	10:55
SUBJECT SAMPLE	.000	10:56
RADIO INTERFERENCE		



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM



PERMIT
TYPE II

DEWAYNE D CARVER

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV W/PRINTER, DATAMASTER, INTOX DMT

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE **05/28/2013**

NUMBER **230101**

EXPIRES **05/28/2015**

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

David Vanhook
 Acting Director

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

LAB-4 (F6-10)

MO 580-0771 (6-10)