



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
 STATE PUBLIC HEALTH LABORATORY
 BREATH ALCOHOL PROGRAM
DATAMASTER MAINTENANCE REPORT

RECEIVED 2/2/14-CD

REVIEWED
 By Carol Day at 9:50 am, Feb 10, 2014

Complete this report at the time of the regular monthly preventive maintenance check (not to exceed 35 days).
 Complete this report whenever the instrument is serviced or repaired and whenever it is placed into service.
 Retain the original and send a copy within 15 days to the Breath Alcohol Program, DHSS.

DATAMASTER SN 204064	NAME OF AGENCY Missouri State Highway Patrol	DATE OF INSPECTION 01/25/2014
LOCATION OF INSTRUMENT (STREET AND CITY) Texas County Sheriff's Department, Houston, Missouri 65483		TIME OF INSPECTION 11:35 am

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

<input checked="" type="checkbox"/> DIAGNOSTIC CHECK (PRINTOUT ATTACHED)	DATE AND TIME (from printout) 01/25/2013 11:35am
<input checked="" type="checkbox"/> COMPUTER	<input checked="" type="checkbox"/> DETECTOR
<input checked="" type="checkbox"/> PROGRAM	<input checked="" type="checkbox"/> FILTERS
<input checked="" type="checkbox"/> HEATERS SAMPLE CHAMBER _____ 49 °C	<input checked="" type="checkbox"/> QUARTZ STANDARD
<input checked="" type="checkbox"/> FLOW DETECTOR	<input checked="" type="checkbox"/> CALIBRATION
<input checked="" type="checkbox"/> PUMP HIGH SPEED	<input checked="" type="checkbox"/> PRINTER

INDICATOR LIGHTS

SIMULATOR SOLUTION SUPPLIER Guth Laboratories, Inc. LOT # 13001 EXP. DATE 03/07/2015

SIMULATOR TEMP (34°C ± 0.2°C) 33.99 °C SIMULATOR SN 1324 EXP. DATE 01/07/2015

CALIBRATION CHECK - (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Mark the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

- 0.100% STANDARD - MUST READ BETWEEN 0.095% AND 0.105% INCLUSIVE
- 0.080% STANDARD - MUST READ BETWEEN 0.076% AND 0.084% INCLUSIVE
- 0.040% STANDARD - MUST READ BETWEEN 0.038% AND 0.042% INCLUSIVE

TEST 1 <input checked="" type="checkbox"/> .095%	TEST 2 <input checked="" type="checkbox"/> .095%	TEST 3 <input checked="" type="checkbox"/> .097%
--	--	--

PERFORM R.F.I. TEST (PRINTOUT ATTACHED)

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:
 (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS	1	(0-.04)	0	(.05-.09)	2	(.10-.14)	0	(.15-.19)	0	OVER .19	1
----------	---	---------	---	-----------	---	-----------	---	-----------	---	----------	---

LIST ANY NEW PARTS AND DESCRIBE ANY ALTERATION OR MODIFICATION THAT WAS MADE TO RESTORE THE INSTRUMENT TO OPERATE SATISFACTORILY AND WITHIN ESTABLISHED LIMITS (USE OTHER SIDE IF NECESSARY).

This instrument is operating satisfactorily within the guidelines set by the Missouri Department of Health, Guth Laboratories Inc., 100% solution, Lot #13001, Bottle # 0410, expiration date 03/07/2015. This instrument was placed into service after being repaired for printer issues.

INSPECTING OFFICER

SIGNATURE <i>Cpl D. B. Pounds #984</i>	PRINT FULL NAME Corporal Dale B. Pounds
TYPE II PERMIT NUMBER/EXPIRATION DATE 220325 09/28/2014	TELEPHONE NUMBER (417) 469-3121

RETURN COMPLETED REPORT TO THE:
 Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
 2875 James Blvd.
 Poplar Bluff, MO 63901

CERTIFICATE OF ANALYSIS

MANUFACTURER AND SUPPLIER: RepCo Marketing, Inc.
LOT NUMBER: 13001
EXPIRATION DATE: March 7, 2015 at 11:59 p.m.

RepCo Marketing, Inc. certifies the following:


RepCo Marketing, Inc. manufactured, tested and supplied Lot Number 13001 of Alcohol Certified Solution for simulators. Random samples of said lot number were analyzed by an independent laboratory utilizing a gas chromatograph and found to contain .1215 gms/dl +/- .003 gms/dl wt/vol ethanol (95% Confidence).

The alcohol and distilled water used in the solution were found to be free of any interfering substance.

This solution will produce a vapor alcohol value of .100 +/- 3% gms/210L Breath when heated to 34 Degrees Celsius +/- 0.2 Degrees Celsius in a simulator (95% Confidence).

The date of manufacture for this lot number is March 8, 2013
The expiration date for this lot number is March 7, 2015 at 11:59 p.m.

This document is a true representation of the original Certificate of Analysis.



Cecil B. Garner, President
RepCo Marketing, Inc.

State of Missouri
DEPARTMENT OF HEALTH



P E R M I T
TYPE II



DALE B POUNDS

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field repairs, and operate the following breath analyzer(s):

DATAMASTER

for the determination of the alcoholic content of blood from a sample of expired (alveolar) air. Issued under the provisions of sections 577.020 through 577.041, RSMo 1986.

Date 09/28/2012

Number 220325

Expires 09/28/2014

MO 580-0771 (7-88)

Director of State Public Health Laboratory

Director, Department of Health

Lab. 4 (R7-88)

BAC DataMaster
Evidence Ticket

MISSOURI STATE HIGHWAY PATROL
BAC DATAMASTER SERIAL NUMBER 204064
01/25/14
11:36

DIAGNOSTIC CHECK

COMPUTER: OKAY
OGRAM (04-07-2009): OKAY
METERS: 49C
AMPLE CHAMBER: OKAY
OIL DETECTOR: OKAY
IMP: OKAY
OIL SPEED: OKAY
OIL TECTOR: OKAY
LITERS: OKAY
MARTZ STANDARD: OKAY
CALIBRATION: OKAY

PRINTER TEST

!##%&'()*+,-./0123456789:;<=>?@ABCDEFGHI
[JKLMNOPQRSTUVWXYZ\]^_`abcdefgijklm
noprstuvwxyz{|}~

Signature Cpl D. B. Pounds

Operator Signature Cpl D. B. Pounds

BAC DataMaster
Evidence Ticket

MISSOURI STATE HIGHWAY PATROL
BAC DATAMASTER SERIAL NUMBER 204064
01/25/14

ARREST TIME: 11:00

SUBJECT NAME:

NELSON/TERRY/R

DOB: 09/09/62

SEX: M

STATE/D.L.: MO/123456789

ARRESTING OFFICER:

POUNDS/D/B

OFFICER I.D.: 984

TESTING OFFICER:

SAME

OFFICER I.D.: SAME

PERMIT NUMBER: 220325

EXPIRATION DATE: 09/28/14

MISCELLANEOUS DATA:

BREATH ANALYSIS

BLANK TEST
INTERNAL STANDARD
RADIO INTERFERENCE

.000
VERIFIED

11:4
11:4

BAC DataMaster
Evidence Ticket

MISSOURI STATE HIGHWAY PATROL
BAC DATAMASTER SERIAL NUMBER 204064
01/25/14

TESTING OFFICER:

POUNDS/D/B

OFFICER I.D.: 984

PERMIT NUMBER: 220325

EXPIRATION DATE: 09/28/14

MISCELLANEOUS DATA:

PRINTER REPAIRED

SUPERVISOR MODE

BLANK TEST
INTERNAL STANDARD
EXTERNAL STANDARD
BLANK TEST
EXTERNAL STANDARD
BLANK TEST
INTERNAL STANDARD
BLANK TEST

N = 3
SIM. = .1
RWG. = .0956

.000
VERIFIED
.095
.000
.095
.000
.097
.000

11:39
11:39
11:39
11:40
11:40
11:41
11:41
11:42

Operator Signature Cpl D. B. Pounds