



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
 STATE PUBLIC HEALTH LABORATORY
 BREATH ALCOHOL PROGRAM
DATAMASTER MAINTENANCE REPORT

RECEIVED REPORT #6
 By Carol Day at 1:40 pm, Jan 16, 2014

Complete this report at the time of the regular monthly preventive maintenance check (not to exceed 35 days).
 Complete this report whenever the instrument is serviced or repaired and whenever it is placed into service.
 Retain the original and send a copy within 15 days to the Breath Alcohol Program, DHSS.

DATAMASTER SN 204067	NAME OF AGENCY MSHP - I	DATE OF INSPECTION 01/07/2014
LOCATION OF INSTRUMENT (STREET AND CITY) Maries County Sheriff's Department, Vienna, Missouri		TIME OF INSPECTION 11:56 am

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

<input checked="" type="checkbox"/> DIAGNOSTIC CHECK (PRINTOUT ATTACHED)	DATE AND TIME (from printout) 01/07/14 / 1156 hours
<input checked="" type="checkbox"/> COMPUTER	<input checked="" type="checkbox"/> DETECTOR
<input checked="" type="checkbox"/> PROGRAM	<input checked="" type="checkbox"/> FILTERS
<input checked="" type="checkbox"/> HEATERS SAMPLE CHAMBER 50 °C	<input checked="" type="checkbox"/> QUARTZ STANDARD
<input checked="" type="checkbox"/> FLOW DETECTOR	<input checked="" type="checkbox"/> CALIBRATION
<input checked="" type="checkbox"/> PUMP HIGH SPEED	<input checked="" type="checkbox"/> PRINTER

INDICATOR LIGHTS

SIMULATOR SOLUTION SUPPLIER Repc Marketing Inc LOT # 13001 EXP. DATE 03/07/2015

SIMULATOR TEMP (34°C ± 0.2°C) 34.0 °C SIMULATOR SN G6796 EXP. DATE 04/03/2014

CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Mark the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

- 0.100% STANDARD - MUST READ BETWEEN 0.095% AND 0.105% INCLUSIVE
- 0.080% STANDARD - MUST READ BETWEEN 0.076% AND 0.084% INCLUSIVE
- 0.040% STANDARD - MUST READ BETWEEN 0.038% AND 0.042% INCLUSIVE

TEST 1 .101	TEST 2 .102	TEST 3 .103
--------------	--------------	--------------

PERFORM R.F.I. TEST (PRINTOUT ATTACHED)

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:
 (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS	0	(0-.04)	0	(.05-.09)	0	(.10-.14)	0	(.15-.19)	0	OVER .19	0
----------	---	---------	---	-----------	---	-----------	---	-----------	---	----------	---

LIST ANY NEW PARTS AND DESCRIBE ANY ALTERATION OR MODIFICATION THAT WAS MADE TO RESTORE THE INSTRUMENT TO OPERATE SATISFACTORILY AND WITHIN ESTABLISHED LIMITS (USE OTHER SIDE IF NECESSARY).

INSPECTING OFFICER

SIGNATURE 	PRINT FULL NAME Corporal Matthew A. Goodson
TYPE II PERMIT NUMBER/EXPIRATION DATE 230021 02/08/2015	TELEPHONE NUMBER (573) 368-2345

RETURN COMPLETED REPORT TO THE: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
 2875 James Blvd.
 Poplar Bluff, MO 63901



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM



PERMIT
TYPE II

MATTHEW A GOODSON

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

DATAMASTER

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119, RSMo.

DATE 02/08/2013

NUMBER 230021

EXPIRES 02/08/2015

W. A. Goodson

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

Shad Vorkel

Acting Director

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

CERTIFICATE OF ANALYSIS

MANUFACTURER AND SUPPLIER: RepCo Marketing, Inc.

LOT NUMBER: 13001

EXPIRATION DATE: March 7, 2015 at 11:59 p.m.

RepCo Marketing, Inc. certifies the following:

RepCo Marketing, Inc. manufactured, tested and supplied Lot Number 13001 of Alcohol Certified Solution for simulators. Random samples of said lot number were analyzed by an independent laboratory utilizing a gas chromatograph and found to contain .1215 gms/dl +/-0.003 gms/dl wt/vol ethanol (95% Confidence).

The alcohol and distilled water used in the solution were found to be free of any interfering substance.

This solution will produce a vapor alcohol value of .100 +/-3% gms/210L Breath when heated to 34 Degrees Celsius +/-0.2 Degrees Celsius in a simulator (95% Confidence).

The date of manufacture for this lot number is March 8, 2013
The expiration date for this lot number is March 7, 2015 at
11:59 p.m.

This document is a true representation of the original Certificate of Analysis.



Cecil B. Garner, President
RepCo Marketing, Inc.

This Side Down - This Edge In First

BAC DataMaster Evidence Ticket

MISSOURI STATE HIGHWAY PATROL
BAC DATAMASTER SERIAL NUMBER 204067
01/07/14

OFFICER:
GOODSON M/A
I.D.: 570
PERMIT NUMBER: 230021
EXPIRATION DATE: 02/08/15
MISCELLANEOUS DATA:
ANNUARY MAINT CHECK

--- SUPERVISOR MODE ---

TEST	.000	12:04
INTERNAL STANDARD	VERIFIED	12:04
INTERNAL STANDARD	.101	12:04
TEST	.000	12:05
INTERNAL STANDARD	.102	12:06
TEST	.000	12:06
INTERNAL STANDARD	.103	12:07
TEST	.000	12:07

.102

CPL M. Goodson

Face This Side Down - This Edge In First

BAC DataMaster Evidence Ticket

MISSOURI STATE HIGHWAY PATROL
BAC DATAMASTER SERIAL NUMBER 204067
01/07/14

ARREST TIME: 11:30
SUBJECT NAME:
ANNUARY MAINT CHECK
DOB: 01/07/14 SEX: M
STATE/D.L.: MO/123456789
ARRESTING OFFICER:
GOODSON M/A
OFFICER I.D.: 570
TESTING OFFICER:
GOODSON M/A
OFFICER I.D.: 570
PERMIT NUMBER: 230021
EXPIRATION DATE: 02/08/15
MISCELLANEOUS DATA:
ANNUARY MAINT CHECK

--- BREATH ANALYSIS ---

BLANK TEST	.000	12:12
INTERNAL STANDARD	VERIFIED	12:12
RADIO INTERFERENCE		

Operator Signature *CPL M. Goodson*

Face This Side Down - This Edge In First

BAC DataMaster Evidence Ticket

MISSOURI STATE HIGHWAY PATROL
BAC DATA MASTER SERIAL NUMBER 204067
01/07/14
11:56

DIAGNOSTIC CHECK: ---

OKAY

24-07-2009) OKAY

AMBER: 50c

CTOR: OKAY

DI: OKAY

OKAY

OKAY

ANDARD: OKAY

OM: OKAY

PRINTER TEST

0)++,-./@123456789:;<=>?@ABCDEFGHIJK
LMNOPQRSTUVWXYZ[\]^_`abcdefghijklmnopq
rstuvwxyz{|}~*+,-./@123456789:;<=>?@

CPL M.A. GARDSON

Face This Side Down - This Edge In First

BAC DataMaster Evidence Ticket

MISSOURI STATE HIGHWAY PATROL
BAC DATA MASTER SERIAL NUMBER 204067
01/07/14

TESTING OFFICER:

GOODSON, M.A.

OFFICER I.D.# 970

PERMIT NUMBER: 230021

EXPIRATION DATE: 02/02/15

MISCELLANEOUS DATA:

JANUARY MAINT CHECK

--- SUPERVISOR MODE ---

BLANK TEST	.000	12:04
INTERNAL STANDARD	VERIFIED	12:04
EXTERNAL STANDARD	.101	12:04
BLANK TEST	.000	12:05
EXTERNAL STANDARD	.102	12:05
BLANK TEST	.000	12:05
EXTERNAL STANDARD	.103	12:07
BLANK TEST	.000	12:07

AVG. = .102

Operator Signature

CPL M.A. GARDSON