

1227



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
STATE PUBLIC HEALTH LABORATORY
BREATH ALCOHOL PROGRAM
DATAMASTER MAINTENANCE REPORT

RECEIVED
By Carol Day at 11:54 am, Sep 10, 2014

Complete this report at the time of the regular monthly preventive maintenance check (not to exceed 35 days).
Complete this report whenever the instrument is serviced or repaired and whenever it is placed into service.
Retain the original and send a copy within 15 days to the Breath Alcohol Program, DHSS.

DATAMASTER SN <u>204073</u>	NAME OF AGENCY <u>MSHP</u>	DATE OF INSPECTION <u>8/29/14</u>
LOCATION OF INSTRUMENT (STREET AND CITY) <u>DALLAS CO. JAIL, 204 S. POPLAR, BUFFALO MO. 65622</u>		TIME OF INSPECTION <u>1939</u>

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observ where determined.) Unmarked items must be corrected before using instrument.

<input checked="" type="checkbox"/> DIAGNOSTIC CHECK (PRINTOUT ATTACHED)	DATE AND TIME (from printout) <u>8/29/14 1939</u>
<input checked="" type="checkbox"/> COMPUTER	<input checked="" type="checkbox"/> DETECTOR
<input checked="" type="checkbox"/> PROGRAM	<input checked="" type="checkbox"/> FILTERS
<input checked="" type="checkbox"/> HEATERS SAMPLE CHAMBER <u>49</u> °C	<input checked="" type="checkbox"/> QUARTZ STANDARD
<input checked="" type="checkbox"/> FLOW DETECTOR	<input checked="" type="checkbox"/> CALIBRATION
<input checked="" type="checkbox"/> PUMP HIGH SPEED	<input checked="" type="checkbox"/> PRINTER

<input checked="" type="checkbox"/> INDICATOR LIGHTS
<input checked="" type="checkbox"/> SIMULATOR SOLUTION SUPPLIER <u>GUTH LABS</u> LOT# <u>13280</u> EXP. DATE <u>10/16/15</u>
<input checked="" type="checkbox"/> SIMULATOR TEMP (34°C ± 0.2°C) <u>34.020</u> °C SIMULATOR SN <u>G11687</u> EXP. DATE <u>2/12/14</u>

CALIBRATION CHECK - (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of less. Mark the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

<input checked="" type="checkbox"/> 0.100% STANDARD - MUST READ BETWEEN 0.095% AND 0.105% INCLUSIVE
<input type="checkbox"/> 0.080% STANDARD - MUST READ BETWEEN 0.076% AND 0.084% INCLUSIVE
<input type="checkbox"/> 0.040% STANDARD - MUST READ BETWEEN 0.038% AND 0.042% INCLUSIVE

TEST 1 <u>.100</u>	TEST 2 <u>.101</u>	TEST 3 <u>.101</u>
--------------------	--------------------	--------------------

PERFORM R.F.I. TEST (PRINTOUT ATTACHED)

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS <u>0</u>	(0-.04) <u>1</u>	(.05-.09) <u>1</u>	(.10-.14) <u>0</u>	(.15-.19) <u>1</u>	OVER .19 <u>1</u>
-------------------	------------------	--------------------	--------------------	--------------------	-------------------

LIST ANY NEW PARTS AND DESCRIBE ANY ALTERATION OR MODIFICATION THAT WAS MADE TO RESTORE THE INSTRUMENT TO OPERATE SATISFACTORILY AND WITHIN ESTABLISH (USE OTHER SIDE IF NECESSARY).
N/A

SIGNATURE <u>[Signature]</u>	PRINT FULL NAME <u>MICHAEL D. MURRILL</u>
TYPE II PERMIT NUMBER/EXPIRATION DATE <u>240065 3/7/14</u>	TELEPHONE NUMBER <u>417-895-6868</u>

RETURN COMPLETED REPORT TO THE: **Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District
2875 James Blvd.
Poplar Bluff, MO 63901**



GUTH LABORATORIES, INC.

830 NORTH 6TH STREET • HARRISBURG, PA 17111-4611 • TELEPHONE: 717-664-5470

CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number **13280** of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on October 18, 2013, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain **0.1217%** (w/vol) ethyl alcohol. The expiration date for this lot number is **October 16, 2015** at 11:59 PM.

When used in a calibrated Simulator, operating at **34°C +/- .2°C**, this solution will give a breath alcohol analysis instrument reading of **0.100 g/210L +/- 3%**.

The alcohol and water used in this solution were free of test interfering substances.

**Ted L. Pauley, President
GUTH LABORATORIES, INC.**

NIST Traceability:

Testing was conducted using Cerilliant Reference Standard lot number FN122211-02 whose values are traceable to NIST.

All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

2

**PERMIT
TYPE II**

MICHAEL D MURRILL

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

DATAMASTER, INTOX DMT

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 3/7/2014

NUMBER 240065

EXPIRES 3/7/2016

MO 580-0771 (6-10)

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

LAB-4 (R6-10)



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.



Operator MURRILL, MICHAEL
Permit No 240065
Date Issued 3/7/2014 Date Expires 3/7/2016

Face This Side Down - This Edge In First

**BAC DataMaster
Evidence Ticket**

MISSOURI STATE HIGHWAY PATROL
BAC DATAMASTER SERIAL NUMBER 204073
08-29-14

ARREST TIME: 19180
SUBJECT NAME:
M/D/M
DOB: 01/01/71 SEX: M
STATE/IDLT# MO: 03456789
ARRESTING OFFICER:
M/D/M
OFFICER I.D.# 649
TESTING OFFICER:
M/D/M
OFFICER I.D.# 649
PERMIT NUMBER: 240065
EXPIRATION DATE: 03/07/16
MISCELLANEOUS DATA:

--- BREATH ANALYSIS ---

BLANK TEST	.000	19149
INTERNAL STANDARD	VERIFIED	19149
RADIO INTERFERENTIAL		

Operator Signature TRAVIS I.D. #649

Face This Side Down - This Edge In First

**BAC DataMaster
Evidence Ticket**

MISSOURI STATE HIGHWAY PATROL
BAC DATAMASTER SERIAL NUMBER 204073
08-29-14

ARREST TIME: 19180
SUBJECT NAME:
M/D/M
DOB: 01/01/71 SEX: M
STATE/IDLT# MO: 03456789
ARRESTING OFFICER:
M/D/M
OFFICER I.D.# 649
TESTING OFFICER:
M/D/M
OFFICER I.D.# 649
PERMIT NUMBER: 240065
EXPIRATION DATE: 03/07/16
MISCELLANEOUS DATA:

--- BREATH ANALYSIS ---

BLANK TEST	.000	19150
INTERNAL STANDARD	VERIFIED	19151
SUBJECT SAMPLE	.000	19152
BLANK TEST	.000	19153

Operator Signature TRAVIS I.D. #649