



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 STATE PUBLIC HEALTH LABORATORY  
 BREATH ALCOHOL PROGRAM  
**DATAMASTER MAINTENANCE REPORT**

received 1/10/14 **REPORT #6**

Complete this report at the time of the regular monthly preventive maintenance check (not to exceed 35 days).  
 Complete this report whenever the instrument is serviced or repaired and whenever it is placed into service.  
 Retain the original and send a copy within 15 days to the Breath Alcohol Program, DHSS.

**REVIEWED**  
 By Carol Day at 1:36 pm, Jan 16, 2014

DATAMASTER SN 204078	NAME OF AGENCY Missouri State Highway Patrol, Troop B, Z-6	DATE OF INSPECTION 01/01/2014
-------------------------	---	----------------------------------

LOCATION OF INSTRUMENT (STREET AND CITY) 518 North Lincoln Street, Kahoka	TIME OF INSPECTION 1:37 pm
--	-------------------------------

**CHECKLIST:** Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

- DIAGNOSTIC CHECK (PRINTOUT ATTACHED)      DATE AND TIME (from printout) 01/01/14 13:37
- COMPUTER       DETECTOR
- PROGRAM       FILTERS
- HEATERS SAMPLE CHAMBER      49 °C       QUARTZ STANDARD
- FLOW DETECTOR       CALIBRATION
- PUMP HIGH SPEED       PRINTER

INDICATOR LIGHTS

SIMULATOR SOLUTION SUPPLIER RepCo Marketing Inc.      LOT # 13001      EXP. DATE 03/07/2015

SIMULATOR TEMP ( $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$ )      34.1      °C      SIMULATOR SN      G11055      EXP. DATE 06/18/2014

CALIBRATION CHECK -- (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within  $\pm 5\%$  of the standard value and must have a spread of .005 or less. Mark the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

- 0.100% STANDARD - MUST READ BETWEEN 0.095% AND 0.105% INCLUSIVE
- 0.080% STANDARD - MUST READ BETWEEN 0.076% AND 0.084% INCLUSIVE
- 0.040% STANDARD - MUST READ BETWEEN 0.038% AND 0.042% INCLUSIVE

TEST 1  .099	TEST 2  .100	TEST 3  .100
--------------	--------------	--------------

PERFORM R.F.I. TEST (PRINTOUT ATTACHED)

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:  
 (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS	0	(0-.04)	0	(.05-.09)	2	(.10-.14)	1	(.15-.19)	1	OVER .19	0
----------	---	---------	---	-----------	---	-----------	---	-----------	---	----------	---

LIST ANY NEW PARTS AND DESCRIBE ANY ALTERATION OR MODIFICATION THAT WAS MADE TO RESTORE THE INSTRUMENT TO OPERATE SATISFACTORILY AND WITHIN ESTABLISHED LIMITS (USE OTHER SIDE IF NECESSARY).

**INSPECTING OFFICER**

SIGNATURE 	PRINT FULL NAME Joseph D. Davidson
---------------	---------------------------------------

TYPE II PERMIT NUMBER/EXPIRATION DATE 220015      02/03/2014	TELEPHONE NUMBER (660) 385-2132
---	------------------------------------

RETURN COMPLETED REPORT TO THE:      Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office  
 2875 James Blvd.  
 Poplar Bluff, MO 63901

**CERTIFICATE OF ANALYSIS**

**MANUFACTURER AND SUPPLIER: RepCo Marketing, Inc.**

**LOT NUMBER: 13001**

**EXPIRATION DATE: March 7, 2015 at 11:59 p.m.**

RepCo Marketing, Inc. certifies the following:


RepCo Marketing, Inc. manufactured, tested and supplied Lot Number 13001 of Alcohol Certified Solution for simulators. Random samples of said lot number were analyzed by an independent laboratory utilizing a gas chromatograph and found to contain .1215 gms/dl +/-0.003 gms/dl wt/vol ethanol (95% Confidence).

The alcohol and distilled water used in the solution were found to be free of any interfering substance.

This solution will produce a vapor alcohol value of .100 +/-3% gms/210L Breath when heated to 34 Degrees Celsius +/-0.2 Degrees Celsius in a simulator (95% Confidence).

The date of manufacture for this lot number is March 8, 2013  
The expiration date for this lot number is March 7, 2015 at  
11:59 p.m.

This document is a true representation of the original Certificate of Analysis.



Cecil B. Garner, President  
RepCo Marketing, Inc.

Face This Side Down - This Edge In First

**BAC DataMaster  
Evidence Ticket**

MISSOURI STATE HIGHWAY PATROL  
BAC DATAMASTER SERIAL NUMBER 204078  
01/01/14  
13:37

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY  
PROGRAM (04-07-2009): OKAY  
HEATERS  
SAMPLE CHAMBER: 49c  
FLOW DETECTOR: OKAY  
PUMP  
HIGH SPEED: OKAY  
DETECTOR: OKAY  
FILTERS: OKAY  
QUARTZ STANDARD: OKAY  
CALIBRATION: OKAY

PRINTER TEST

!"#\$%&'()\*+,-./0123456789:;<=>?@ABCDEFGHI  
HIJKLMNOPQRSTUVWXYZ[\]^\_`abcdefghijklmnopq  
rstuvwxyz{|}~

Operator Signature

*Tpa JD Davison*

Face This Side Down - This Edge In First

**BAC DataMaster  
Evidence Ticket**

MISSOURI STATE HIGHWAY PATROL  
BAC DATAMASTER SERIAL NUMBER 284078  
01/01/14

TESTING OFFICER:  
DAVIDSON/JOSEPH/D  
OFFICER I.D.: 191  
PERMIT NUMBER: 220015  
EXPIRATION DATE: 02/03/14  
MISCELLANEOUS DATA:  
.10 MAINT.

--- SUPERVISOR MODE ---

BLANK TEST	.000	13:41
INTERNAL STANDARD	VERIFIED	13:41
EXTERNAL STANDARD	.099	13:41
BLANK TEST	.000	13:42
EXTERNAL STANDARD	.100	13:42
BLANK TEST	.000	13:43
EXTERNAL STANDARD	.100	13:43
BLANK TEST	.000	13:44

N = 3  
SIM. = .1  
AVG. = .0996

Operator Signature

*TPR JD Davidson*

Face This Side Down - This Edge In First

**BAC DataMaster  
Evidence Ticket**

MISSOURI STATE HIGHWAY PATROL  
BAC DATAMASTER SERIAL NUMBER 204078  
01/01/14

ARREST TIME: 13:25  
SUBJECT NAME:  
TEST/RFI  
DOB: 01/01/90      SEX: M  
STATE/D.L.: NA/NA  
ARRESTING OFFICER:  
DAVIDSON/JOSEPH/D  
OFFICER I.D.: 191  
TESTING OFFICER:  
DAVIDSON/JOSEPH/D  
OFFICER I.D.: 191  
PERMIT NUMBER: 220015  
EXPIRATION DATE: 02/03/14  
MISCELLANEOUS DATA:  
RFI TEST

--- BREATH ANALYSIS ---

BLANK TEST	.000	13:47
INTERNAL STANDARD	VERIFIED	13:47
RADIO INTERFERENCE		

Operator Signature \_\_\_\_\_

*Tea JD Daniels*

State of Missouri  
DEPARTMENT OF HEALTH



P E R M I T  
TYPE II



JOSEPH D DAVIDSON

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field repairs, and operate the following breath analyzer(s):

**DATAMASTER**

for the determination of the alcoholic content of blood from a sample of expired (alveolar) air. Issued under the provisions of sections 577.020 through 577.041, RSMo 1986.

Date 02/03/2012

Number **220015**

Expires 02/03/2014

MO 580-0771 (7-88)

Director of State Public Health Laboratory

Director, Department of Health

Lab. 4 (R7-88)