MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
STATE PUBLIC HEALTH LABORATORY
BREATH ALCOHOL PROGRAM

DATAMASTER MAINTENANCE REPORT

REPORT #6

Complete this report at the time of the regular monthly preventive maintenance check (not to exceed 35 days).
Complete this report whenever the instrument is serviced or repaired and whenever it is placed into service.
Retain the original and send a copy within 15 days to the Breath Alcohol Program, DHSS.

RECEIVED
By Carol Day at 11:11 am, Jan 13, 2014

DATAMASTER SN	NAME OF AGENCY	DATE OF INSPECTION
204091 - Inv# 127213	Missouri State Highway Patrol	01/06/2014

LOCATION OF INSTRUMENT (STREET AND CITY)	TIME OF INSPECTION
Wayne County Sheriff's Department, Greenville, Missouri, 63944	10:43 pm

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

- ☑ DIAGNOSTIC CHECK (PRINTOUT ATTACHED) — DATE AND TIME (from printout) 01/06/2014 - 22:43
- ☑ COMPUTER — ☑ DETECTOR
- ☑ PROGRAM — ☑ FILTERS
- ☑ HEATERS SAMPLE CHAMBER 49 °C — ☑ QUARTZ STANDARD
- ☑ FLOW DETECTOR — ☑ CALIBRATION
- ☑ PUMP HIGH SPEED — ☑ PRINTER
- ☑ INDICATOR LIGHTS
- ☑ SIMULATOR SOLUTION SUPPLIER Guth Laboratories Inc LOT # 13280 EXP. DATE 10/16/2015
- ☑ SIMULATOR TEMP (34°C ± 0.2°C) 34.0 °C SIMULATOR SN G6797 EXP. DATE 03/04/2014
- ☑ CALIBRATION CHECK – **(ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)**

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Mark the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

- ☑ 0.100% STANDARD - MUST READ BETWEEN 0.095% AND 0.105% INCLUSIVE
- ☐ 0.080% STANDARD - MUST READ BETWEEN 0.076% AND 0.084% INCLUSIVE
- ☐ 0.040% STANDARD - MUST READ BETWEEN 0.038% AND 0.042% INCLUSIVE

TEST 1	TEST 2	TEST 3
.098%	.098%	.098%

- ☑ PERFORM R.F.I. TEST (PRINTOUT ATTACHED)

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS	(0-.04)	(.05-.09)	(.10-.14)	(.15-.19)	OVER .19
1	0	1	1	0	1

LIST ANY NEW PARTS AND DESCRIBE ANY ALTERATION OR MODIFICATION THAT WAS MADE TO RESTORE THE INSTRUMENT TO OPERATE SATISFACTORILY AND WITHIN ESTABLISHED LIMITS (USE OTHER SIDE IF NECESSARY).

INSPECTING OFFICER

SIGNATURE	PRINT FULL NAME
▸ John J. West	John J. West

TYPE II PERMIT NUMBER/EXPIRATION DATE	TELEPHONE NUMBER
220284 - 09/15/2014	(573) 840-9500

RETURN COMPLETED REPORT TO THE: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
2875 James Blvd.
Poplar Bluff, MO 63901

AN EQUAL OPPORTUNITY/AFFIRMATIVE ACTION EMPLOYER
services provided on a nondiscrimatory basis

GUTH LABORATORIES, INC.
590 NORTH 67th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-564-5470

CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number **13280** of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on **October 18, 2013,** using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain **0.1217%** (w/vol) ethyl alcohol. The expiration date for this lot number is **October 16, 2015** at 11:59 PM.

When used in a calibrated Simulator, operating at 34°C +/- .2°C, this solution will give a breath alcohol analysis instrument reading of **0.100 g/210L +/- 3%.**

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President
GUTH LABORATORIES, INC.

NIST Traceability:
Testing was conducted using Cerilliant Reference Standard lot number FN122211-02 whose values are traceable to NIST.
All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.

Face This Side Down - This Edge In First

BAC DataMaster
Evidence Ticket

MISSOURI STATE HIGHWAY PATROL
BAC DATAMASTER SERIAL NUMBER 204091
01/06/14

TESTING OFFICER:
WEST/JOHN/J
OFFICER I.D.: 1145
PERMIT NUMBER: 220284
EXPIRATION DATE: 09/15/14
MISCELLANEOUS DATA:

--- SUPERVISOR MODE ---

BLANK TEST	.000	22:57
INTERNAL STANDARD	VERIFIED	22:58
EXTERNAL STANDARD	.098	22:58
BLANK TEST	.000	22:59
EXTERNAL STANDARD	.098	22:59
BLANK TEST	.000	23:00
EXTERNAL STANDARD	.098	23:00
BLANK TEST	.000	23:01

N = 3
SIM. = .1
AVG. = .098

Operator Signature _____

Face This Side Down - This Edge In First

BAC DataMaster
Evidence Ticket

MISSOURI STATE HIGHWAY PATROL
BAC DATAMASTER SERIAL NUMBER 204091
01/06/14
22:43

--- DIAGNOSTIC CHECK ---

COMPUTER:	OKAY
PROGRAM (04-07-2009):	OKAY
HEATERS	
SAMPLE CHAMBER:	49c
FLOW DETECTOR:	OKAY
PUMP	
HIGH SPEED:	OKAY
DETECTOR:	OKAY
FILTERS:	OKAY
QUARTZ STANDARD:	OKAY
CALIBRATION:	OKAY

PRINTER TEST
!"#$%&'()*+,-./0123456789:;<=>?@ABCDEFG
HIJKLMNOPQRSTUVWXYZ[\]^_`abcdefghijklmno
pqrstuvwxyz{|}~

Operator Signature _____

Face This Side Down - This Edge In First

BAC DataMaster
Evidence Ticket

MISSOURI STATE HIGHWAY PATROL
BAC DATAMASTER SERIAL NUMBER 204091
01/06/14

ARREST TIME: 22:00
SUBJECT NAME:
TEST
DOB: 10/10/90 SEX: M
STATE/D.L.: MO/1234567890
ARRESTING OFFICER:
WEST/JOHN/J
OFFICER I.D.: 1145
TESTING OFFICER:
WEST/JOHN/J
OFFICER I.D.: 1145
PERMIT NUMBER: 220284
EXPIRATION DATE: 09/15/14
MISCELLANEOUS DATA:

--- BREATH ANALYSIS ---

BLANK TEST .000 23:02
RADIO INTERFERENCE

Operator Signature