



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
 STATE PUBLIC HEALTH LABORATORY
 BREATH ALCOHOL PROGRAM
DATAMASTER MAINTENANCE REPORT

received 01/08/14-cd **REPORT #6**

Complete this report at the time of the regular monthly preventive maintenance check (not to exceed 35 days).
 Complete this report whenever the instrument is serviced or repaired and whenever it is placed into service.
 Retain the original and send a copy within 15 days to the Breath Alcohol Program, DHSS.

REVIEWED
 By Carol Day at 1:30 pm, Jan 16, 2014

DATAMASTER SN 204094 Inv#127220	NAME OF AGENCY Missouri State Highway Patrol	DATE OF INSPECTION 01/01/2014
LOCATION OF INSTRUMENT (STREET AND CITY) Douglas County Sheriff's Office, 209 SE 2nd Ave., Ava, Missouri 65608		TIME OF INSPECTION 6:47 pm

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

<input checked="" type="checkbox"/> DIAGNOSTIC CHECK (PRINTOUT ATTACHED)	DATE AND TIME (from printout) 01/01/2014 1847hrs
<input checked="" type="checkbox"/> COMPUTER	<input checked="" type="checkbox"/> DETECTOR
<input checked="" type="checkbox"/> PROGRAM	<input checked="" type="checkbox"/> FILTERS
<input checked="" type="checkbox"/> HEATERS SAMPLE CHAMBER 50 °C	<input checked="" type="checkbox"/> QUARTZ STANDARD
<input checked="" type="checkbox"/> FLOW DETECTOR	<input checked="" type="checkbox"/> CALIBRATION
<input checked="" type="checkbox"/> PUMP HIGH SPEED	<input checked="" type="checkbox"/> PRINTER
<input checked="" type="checkbox"/> INDICATOR LIGHTS	
<input checked="" type="checkbox"/> SIMULATOR SOLUTION SUPPLIER RepCo Marketing Inc.	LOT # 13001bot#0487 EXP. DATE 03/07/2013
<input checked="" type="checkbox"/> SIMULATOR TEMP (34°C ± 0.2°C) 34.1C °C	SIMULATOR SN G11072 EXP. DATE 11/19/2014
<input checked="" type="checkbox"/> CALIBRATION CHECK - (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)	
Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Mark the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)	
<input checked="" type="checkbox"/> 0.100% STANDARD - MUST READ BETWEEN 0.095% AND 0.105% INCLUSIVE	
<input type="checkbox"/> 0.080% STANDARD - MUST READ BETWEEN 0.076% AND 0.084% INCLUSIVE	
<input type="checkbox"/> 0.040% STANDARD - MUST READ BETWEEN 0.038% AND 0.042% INCLUSIVE	

TEST 1 ♦ .102%	TEST 2 ♦ .100%	TEST 3 ♦ .101%
----------------	----------------	----------------

PERFORM R.F.I. TEST (PRINTOUT ATTACHED)

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:
 (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS 0	(0-.04) 1	(.05-.09) 0	(.10-.14) 2	(.15-.19) 0	OVER .19 1
------------	-----------	-------------	-------------	-------------	------------

LIST ANY NEW PARTS AND DESCRIBE ANY ALTERATION OR MODIFICATION THAT WAS MADE TO RESTORE THE INSTRUMENT TO OPERATE SATISFACTORILY AND WITHIN ESTABLISHED LIMITS (USE OTHER SIDE IF NECESSARY).

This instrument is operating within the guidelines set by the Missouri Department of Health.

SIGNATURE <i>Trooper D.J. Johnson #439</i>	PRINT FULL NAME Trooper D.J. Johnson #439
TYPE II PERMIT NUMBER/EXPIRATION DATE 220092 04/10/2014	TELEPHONE NUMBER (417) 469-3121

RETURN COMPLETED REPORT TO THE:
 Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
 2875 James Blvd.
 Poplar Bluff, MO 63901

CERTIFICATE OF ANALYSIS

MANUFACTURER AND SUPPLIER: RepCo Marketing, Inc.

LOT NUMBER: 13001

EXPIRATION DATE: March 7, 2015 at 11:59 p.m.

RepCo Marketing, Inc. certifies the following:

RepCo Marketing, Inc. manufactured, tested and supplied Lot Number 13001 of Alcohol Certified Solution for simulators. Random samples of said lot number were analyzed by an independent laboratory utilizing a gas chromatograph and found to contain .1215 gms/dl +/- .003 gms/dl wt/vol ethanol (95% Confidence).

The alcohol and distilled water used in the solution were found to be free of any interfering substance.

This solution will produce a vapor alcohol value of .100 +/- 3% gms/210L Breath when heated to 34 Degrees Celsius +/- 0.2 Degrees Celsius in a simulator (95% Confidence).

The date of manufacture for this lot number is March 8, 2013
The expiration date for this lot number is March 7, 2015 at 11:59 p.m.

This document is a true representation of the original Certificate of Analysis.



Cecil B. Garner, President
RepCo Marketing, Inc.

State of Missouri
DEPARTMENT OF HEALTH



P E R M I T
TYPE II



DANIEL J JOHNSON

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field repairs, and operate the following breath analyzer(s):

DATAMASTER

for the determination of the alcoholic content of blood from a sample of expired (alveolar) air. Issued under the provisions of sections 577.020 through 577.041, RSMo 1986.

Date 04/10/2012

Number 220092

Expires 04/10/2014

MO 580-0771 (7-88)

Director of State Public Health Laboratory

Director, Department of Health

Lab. 4 (R7-88)

Face This Side Down - This Edge In First

BAC DataMaster Evidence Ticket

MISSOURI STATE POLICE
BAC DATA MASTER SERIAL NUMBER 2648824
REV. 01-13

DATE OF TEST: 10/13/13
SUBJECT NAME:
ADDRESS:
TIME OF TEST: 11:15 AM
TEST TYPE: 100% BAC
OFFICER: [REDACTED]
OFFICER ID: [REDACTED]
PERMIT NUMBER: 2648824
EXPIRATION DATE: 01/13/14
MOTOR VEHICLE ID:
MOTOR TYPE:
MOTOR MAKE:

DRIVER LICENSE:

DRIVER LICENSE	10000	10000
EXPIRATION DATE	01/13/14	01/13/14
ISSUE DATE	01/13/13	01/13/13

Face This Side Down - This Edge In First

BAC DataMaster Evidence Ticket

MISSOURI STATE POLICE
BAC DATA MASTER SERIAL NUMBER 2648824
REV. 01-13

DATE OF TEST: 10/13/13
SUBJECT NAME:
ADDRESS:
TIME OF TEST: 11:15 AM
TEST TYPE: 100% BAC
OFFICER: [REDACTED]
OFFICER ID: [REDACTED]
PERMIT NUMBER: 2648824
EXPIRATION DATE: 01/13/14
MOTOR VEHICLE ID:
MOTOR TYPE:
MOTOR MAKE:

DRIVER LICENSE:

DRIVER LICENSE	10000	10000
EXPIRATION DATE	01/13/14	01/13/14
ISSUE DATE	01/13/13	01/13/13

Operator Signature

Trooper D.J. Florman #1439

Operator Signature

Trooper D.J. Florman #1439

