



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
 STATE PUBLIC HEALTH LABORATORY
 BREATH ALCOHOL PROGRAM
DATAMASTER MAINTENANCE REPORT

748

RECEIVED
 By Carol Day at 9:27 am, Feb 10, 2014

Complete this report at the time of the regular monthly preventive maintenance check (not to exceed 35 days).
 Complete this report whenever the instrument is serviced or repaired and whenever it is placed into service.
 Retain the original and send a copy within 15 days to the Breath Alcohol Program, DHSS.

DATAMASTER SN 204096	NAME OF AGENCY Missouri State Highway Patrol	DATE OF INSPECTION 01/28/2014
LOCATION OF INSTRUMENT (STREET AND CITY) Marion County Sheriff's Dept., 1703 Marion City Road, Palmyra, MO		TIME OF INSPECTION 5:30 pm

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

<input checked="" type="checkbox"/> DIAGNOSTIC CHECK (PRINTOUT ATTACHED)	DATE AND TIME (from printout) <u>01/28/2014 1730</u>
<input checked="" type="checkbox"/> COMPUTER	<input checked="" type="checkbox"/> DETECTOR
<input checked="" type="checkbox"/> PROGRAM	<input checked="" type="checkbox"/> FILTERS
<input checked="" type="checkbox"/> HEATERS SAMPLE CHAMBER <u>49</u> °C	<input checked="" type="checkbox"/> QUARTZ STANDARD
<input checked="" type="checkbox"/> FLOW DETECTOR	<input checked="" type="checkbox"/> CALIBRATION
<input checked="" type="checkbox"/> PUMP HIGH SPEED	<input checked="" type="checkbox"/> PRINTER

<input checked="" type="checkbox"/> INDICATOR LIGHTS
<input checked="" type="checkbox"/> SIMULATOR SOLUTION SUPPLIER <u>RepCo Marketing Inc.</u> LOT # <u>13001</u> EXP. DATE <u>03/07/2015</u>
<input checked="" type="checkbox"/> SIMULATOR TEMP (34°C ± 0.2°C) <u>34.0</u> °C SIMULATOR SN <u>G6817</u> EXP. DATE <u>05/02/2014</u>

CALIBRATION CHECK - (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Mark the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

0.100% STANDARD - MUST READ BETWEEN 0.095% AND 0.105% INCLUSIVE
 0.080% STANDARD - MUST READ BETWEEN 0.076% AND 0.084% INCLUSIVE
 0.040% STANDARD - MUST READ BETWEEN 0.038% AND 0.042% INCLUSIVE

TEST 1 .098%	TEST 2 .099%	TEST 3 .100%
---------------	---------------	---------------

PERFORM R.F.I. TEST (PRINTOUT ATTACHED)

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:
 (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS	1	(.0-.04)	0	(.05-.09)	0	(.10-.14)	2	(.15-.19)	0	OVER .19	0
----------	---	----------	---	-----------	---	-----------	---	-----------	---	----------	---

LIST ANY NEW PARTS AND DESCRIBE ANY ALTERATION OR MODIFICATION THAT WAS MADE TO RESTORE THE INSTRUMENT TO OPERATE SATISFACTORILY AND WITHIN ESTABLISHED LIMITS (USE OTHER SIDE IF NECESSARY).

INSPECTING OFFICER	
SIGNATURE 	PRINT FULL NAME James P. Johnson, Trooper
TYPE II PERMIT NO./SER/EXPIRATION DATE 220250 / 09/11/2014	TELEPHONE NUMBER (660) 385-2132

RETURN COMPLETED REPORT TO THE:
 Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
 2875 James Blvd.
 Poplar Bluff, MO 63901

CERTIFICATE OF ANALYSIS

MANUFACTURER AND SUPPLIER: RepCo Marketing, Inc.

LOT NUMBER: 13001

EXPIRATION DATE: March 7, 2015 at 11:59 p.m.

RepCo Marketing, Inc. certifies the following:

RepCo Marketing, Inc. manufactured, tested and supplied Lot Number 13001 of Alcohol Certified Solution for simulators. Random samples of said lot number were analyzed by an independent laboratory utilizing a gas chromatograph and found to contain .1215 gms/dl +/- .003 gms/dl wt/vol ethanol (95% Confidence).

The alcohol and distilled water used in the solution were found to be free of any interfering substance.

This solution will produce a vapor alcohol value of .100 +/- 3% gms/210L Breath when heated to 34 Degrees Celsius +/- 0.2 Degrees Celsius in a simulator (95% Confidence).

The date of manufacture for this lot number is March 8, 2013
The expiration date for this lot number is March 7, 2015 at
11:59 p.m.

This document is a true representation of the original Certificate of Analysis.



Cecil B. Garner, President
RepCo Marketing, Inc.

State of Missouri
DEPARTMENT OF HEALTH



P E R M I T
TYPE II



JAMES P JOHNSON

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field repairs, and operate the following breath analyzer(s):

DATAMASTER

for the determination of the alcoholic content of blood from a sample of expired (alveolar) air. Issued under the provisions of sections 577.020 through 577.041, RSMo 1986.

Date 09/11/2012

Number 220250

Expires 09/11/2014

MO 580-0771 (7-88)

Director of State Public Health Laboratory

Director, Department of Health

Lab. 4 (R7-88)

Face This Side Down - This Edge In First

BAC DataMaster Evidence Ticket

MISSOURI STATE HIGHWAY PATROL
BAC DATAMASTER SERIAL NUMBER 284896
01-28-14


TESTING OFFICER:
JOHNSON/JAMES/P
OFFICER I.D.#: 487
PERMIT NUMBER: 220256
EXPIRATION DATE: 06/30/14
MISCELLANEOUS DATA:

--- SUPERVISOR MODE ---

BLANK TEST	.000	17:32
INTERNAL STANDARD	VERIFIED	17:32
EXTERNAL STANDARD	.098	17:32
BLANK TEST	.000	17:33
EXTERNAL STANDARD	.099	17:33
BLANK TEST	.000	17:34
EXTERNAL STANDARD	.100	17:34
BLANK TEST	.000	17:35

R = 3
S.D. = .03
M.S. = .000

Operator Signature



Face This Side Down - This Edge In First

BAC DataMaster Evidence Ticket


MISSOURI STATE HIGHWAY PATROL
BAC DATAMASTER SERIAL NUMBER 284896
01-28-14

ARRIVAL TIME: 17:30
SUBJECT NAME:
RE-TEST
DOB: 11/11-11 SEX: M
STATUS: 1: 10/10
ARRESTING OFFICER:
NO
OFFICER I.D.#: 487
TESTING OFFICER:
JOHNSON/JAMES/P
OFFICER I.D.#: 487
PERMIT NUMBER: 220256
EXPIRATION DATE: 06/30/14
MISCELLANEOUS DATA:
BAC0 PREVIOUSLY UNLAWFUL 11/11

--- SUBJECT IDENTIFY ---

BLANK TEST	.000	17:34
INTERNAL STANDARD	VERIFIED	17:34
EXTERNAL STANDARD	.100	17:34

Operator Signature



Face This Side Down - This Edge In First

BAC DataMaster Evidence Ticket

MISSOURI STATE HIGHWAY PATROL
BAC DATAMASTER SERIAL NUMBER: 204096
01/28/14
1:30

--- DIAGNOSTIC CHECK ---

COMPUTER:	OKAY
PROGRAM (BA-07-2009):	OKAY
HEATERS:	
SAMPLE CHAMBER:	45c
FLOW DETECTOR:	OKAY
PUMP:	
HIGH SPEED:	OKAY
DETECTORS:	OKAY
FILTERS:	OKAY
WEIGHT STANDARDS:	OKAY
PERFORMANCE:	OKAY

PRINTER TEST

MISSOURI STATE HIGHWAY PATROL
LABORATORY

Operator Signature

