



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
STATE PUBLIC HEALTH LABORATORY
BREATH ALCOHOL PROGRAM
DATAMASTER MAINTENANCE REPORT

RECEIVED 2/5/14-CD

REVIEWED
By Carol Day at 10:05 am, Mar 14, 2014

Complete this report at the time of the regular monthly preventive maintenance check (not to exceed 35 days).
Complete this report whenever the instrument is serviced or repaired and whenever it is placed into service.
Retain the original and send a copy within 15 days to the Breath Alcohol Program, DHSS.

DATAMASTER SN 204100	NAME OF AGENCY Missouri State Highway Patrol	DATE OF INSPECTION 01/28/2014
LOCATION OF INSTRUMENT (STREET AND CITY) Chariton County Sheriff's Department, 307 S. Cherry, Keytesville, MO		TIME OF INSPECTION 9:01 am

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

<input checked="" type="checkbox"/> DIAGNOSTIC CHECK (PRINTOUT ATTACHED)	DATE AND TIME (from printout) 01/28/14 0901
<input checked="" type="checkbox"/> COMPUTER	<input checked="" type="checkbox"/> DETECTOR
<input checked="" type="checkbox"/> PROGRAM	<input checked="" type="checkbox"/> FILTERS
<input checked="" type="checkbox"/> HEATERS SAMPLE CHAMBER 49°C	<input checked="" type="checkbox"/> QUARTZ STANDARD
<input checked="" type="checkbox"/> FLOW DETECTOR	<input checked="" type="checkbox"/> CALIBRATION
<input checked="" type="checkbox"/> PUMP HIGH SPEED	<input checked="" type="checkbox"/> PRINTER

INDICATOR LIGHTS

SIMULATOR SOLUTION SUPPLIER RepCo Marketing Inc. LOT # 12002 EXP. DATE 08/29/2014

SIMULATOR TEMP (34°C ± 0.2°C) 34 °C SIMULATOR SN G 6815 EXP. DATE 04/03/2014

CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Mark the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

0.100% STANDARD - MUST READ BETWEEN 0.095% AND 0.105% INCLUSIVE
 0.080% STANDARD - MUST READ BETWEEN 0.076% AND 0.084% INCLUSIVE
 0.040% STANDARD - MUST READ BETWEEN 0.038% AND 0.042% INCLUSIVE

TEST 1 \blacklozenge .100	TEST 2 \blacklozenge .100	TEST 3 \blacklozenge .101
-----------------------------	-----------------------------	-----------------------------

PERFORM R.F.I. TEST (PRINTOUT ATTACHED)

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:
(DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS	0	(0-.04)	3	(.05-.09)	0	(.10-.14)	1	(.15-.19)	0	OVER .19	1
----------	---	---------	---	-----------	---	-----------	---	-----------	---	----------	---

LIST ANY NEW PARTS AND DESCRIBE ANY ALTERATION OR MODIFICATION THAT WAS MADE TO RESTORE THE INSTRUMENT TO OPERATE SATISFACTORILY AND WITHIN ESTABLISHED LIMITS
(USE OTHER SIDE IF NECESSARY).

Meets Department of Health Standards

INSPECTING OFFICER

SIGNATURE <i>Lane P. Quinn</i>	PRINT FULL NAME Lane P. Quinn
TYPE II PERMIT NUMBER/EXPIRATION DATE 220023 / 02-03-14	TELEPHONE NUMBER (660) 385-2132

RETURN COMPLETED REPORT TO THE: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
2875 James Blvd.
Poplar Bluff, MO 63901

CERTIFICATE OF ANALYSIS

MANUFACTURER AND SUPPLIER: RepCo Marketing, Inc.
LOT NUMBER: 12002
EXPIRATION DATE: August 29, 2014 at 11:59 p.m.

RepCo Marketing, Inc. certifies the following:


RepCo Marketing, Inc. manufactured, tested and supplied Lot Number 12002 of Alcohol Certified Solution for simulators. Random samples of said lot number were analyzed by an independent laboratory utilizing a gas chromatograph and found to contain .1209 gms/dl +/-0.003 gms/dl wt/vol ethanol (95% Confidence).

The alcohol and distilled water used in the solution were found to be free of any interfering substance.

This solution will produce a vapor alcohol value of .100 +/-3% gms/210L Breath when heated to 34 Degrees Celsius +/-0.2 Degrees Celsius in a simulator (95% Confidence).

The date of manufacture for this lot number is August 30, 2012
The expiration date for this lot number is August 29, 2014 at 11:59 p.m.

This document is a true representation of the original Certificate of Analysis.



Cecil B. Garner, President
RepCo Marketing, Inc.

State of Missouri
DEPARTMENT OF HEALTH



P E R M I T
TYPE II



LANE P QUINN

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field repairs, and operate the following breath analyzer(s):

DATAMASTER

for the determination of the alcoholic content of blood from a sample of expired (alveolar) air. Issued under the provisions of sections 577.020 through 577.041, RSMo 1986.

Date 02/03/2012

Number 220023

Expires 02/03/2014

Director of State Public Health Laboratory

Director, Department of Health

Face This Side Down - This Edge In First

BAC DataMaster Evidence Ticket

MISSOURI STATE HIGHWAY PATROL
BAC DATAMASTER SERIAL NUMBER 204100
01/28/14

ARREST TIME: 08:30
SUBJECT NAME:
RFI/TEST
DOB: 01/01/81 SEX: M
STATE/D.L.: MO/1111111111
ARRESTING OFFICER:
QUINN/L/P
OFFICER I.D.#: 937
TESTING OFFICER:
QUINN/L/P
OFFICER I.D.#: 937
PERMIT NUMBER: 220825
EXPIRATION DATE: 02/03/14
MISCELLANEOUS DATA:

--- BREATH ANALYSIS ---

BLANK TEST .000 09:13
INTERNAL STANDARD VERIFIED 09:14
RADIO INTERFERENCE

Operator Signature

2208-02

Face This Side Down - This Edge In First

BAC DataMaster Evidence Ticket

MISSOURI STATE HIGHWAY PATROL
BAC DATAMASTER SERIAL NUMBER 204100
01/28/14
09:01

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY
PROGRAM (04-07-2009): OKAY
HEATERS
SAMPLE CHAMBER: 49c
FLOW DETECTOR: OKAY
PUMP
HIGH SPEED: OKAY
DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY

PRINTER TEST

!"#\$%&'()*+,-./0123456789:;<=>?@ABCDEFGHI
HIJKLANDPQRSTUvwxyz{|}~^_`abc,defghijklmnop
qrs tuvwxyz () * ^

Operator Signature

2208-02

Face This Side Down - This Edge In First

Bals

BAC DataMaster Evidence Ticket

MISSOURI STATE HIGHWAY PATROL
BAC DATAMASTER SERIAL NUMBER 2204150
01/28/14

TESTING OFFICER:
QUINN/L/P
OFFICER I.D.# 937
PERMIT NUMBER: 220023
EXPIRATION DATE: 02/03/14
MISCELLANEOUS DATA:

--- SUPERVISOR MODE ---

BLANK TEST	.000	09:03
INTERNAL STANDARD	VERIFIED	09:03
EXTERNAL STANDARD	.100	09:04
BLANK TEST	.000	09:04
EXTERNAL STANDARD	.100	09:05
BLANK TEST	.000	09:05
EXTERNAL STANDARD	.101	09:06
BLANK TEST	.000	09:06

N = 3
Stn. = .1
AVG. = .1003

Operator Signature

