MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
STATE PUBLIC HEALTH LABORATORY
BREATH ALCOHOL PROGRAM

received 1/28/14-cd

DATAMASTER MAINTENANCE REPORT

REVIEWED REPORT #6
By Carol Day at 1:14 pm, Feb 04, 2014

Complete this report at the time of the regular monthly preventive maintenance check (not to exceed 35 days). Complete this report whenever the instrument is serviced or repaired and whenever it is placed into service. Retain the original and send a copy within 15 days to the Breath Alcohol Program, DHSS.

DATAMASTER SN	NAME OF AGENCY	DATE OF INSPECTION
204101	MISSOURI STATE HIGHWAY PATROL	01/20/2014
LOCATION OF INSTRUMENT (STREET AND CITY)		TIME OF INSPECTION
PUTNAM COUNTY SHERIFF'S OFFICE, 1601 MAIN, UNIONVILLE, MO		8:30 pm

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

- ☑ DIAGNOSTIC CHECK (PRINTOUT ATTACHED) — DATE AND TIME (from printout) 01/20/2014 2030 HOURS
  - ☑ COMPUTER — ☑ DETECTOR
  - ☑ PROGRAM — ☑ FILTERS
  - ☑ HEATERS SAMPLE CHAMBER 49 °C — ☑ QUARTZ STANDARD
  - ☑ FLOW DETECTOR — ☑ CALIBRATION
  - ☑ PUMP HIGH SPEED — ☑ PRINTER
- ☑ INDICATOR LIGHTS
- ☑ SIMULATOR SOLUTION SUPPLIER REPCO MARKETING, INC LOT # 13001 EXP. DATE 03/07/2015
- ☑ SIMULATOR TEMP (34°C ± 0.2°C) 34 °C SIMULATOR SN G11077 EXP. DATE 10/28/2014
- ☑ CALIBRATION CHECK – **(ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)**

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Mark the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

- ☑ 0.100% STANDARD - MUST READ BETWEEN 0.095% AND 0.105% INCLUSIVE
- ☐ 0.080% STANDARD - MUST READ BETWEEN 0.076% AND 0.084% INCLUSIVE
- ☐ 0.040% STANDARD - MUST READ BETWEEN 0.038% AND 0.042% INCLUSIVE

TEST 1 ☛ .095	TEST 2 ☛ .097	TEST 3 ☛ .098

- ☑ PERFORM R.F.I. TEST (PRINTOUT ATTACHED)

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS	(0-.04)	(.05-.09)	(.10-.14)	(.15-.19)	OVER .19
2	0	1	0	0	0

LIST ANY NEW PARTS AND DESCRIBE ANY ALTERATION OR MODIFICATION THAT WAS MADE TO RESTORE THE INSTRUMENT TO OPERATE SATISFACTORILY AND WITHIN ESTABLISHED LIMITS (USE OTHER SIDE IF NECESSARY).

MEETS DEPARTMENT OF HEALTH AND SENIOR SERVICES STANDARDS

INSPECTING OFFICER

SIGNATURE	PRINT FULL NAME
▸ Reid T. Fessler	REID T. FESSLER
TYPE II PERMIT NUMBER/EXPIRATION DATE	TELEPHONE NUMBER
220016 02/03/2014	(660) 385-2132

RETURN COMPLETED REPORT TO THE: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
2875 James Blvd.
Poplar Bluff, MO 63901

REPCO MARKETING INC.

3101-188 STONYBROOK DRIVE
RALEIGH, N.C. 27604
919-876-5480

CERTIFICATE OF ANALYSIS

MANUFACTURER AND SUPPLIER: RepCo Marketing, Inc.
LOT NUMBER: 13001
EXPIRATION DATE: March 7, 2015 at 11:59 p.m.

RepCo Marketing, Inc. certifies the following:

RepCo Marketing, Inc. manufactured, tested and supplied Lot Number **13001** of Alcohol Certified Solution for simulators. Random samples of said lot number were analyzed by an independent laboratory utilizing a gas chromatograph and found to contain **.1215** gms/dl +/-.003 gms/dl wt/vol ethanol (95% Confidence).

The alcohol and distilled water used in the solution were found to be free of any interferring substance.

This solution will produce a vapor alcohol value of **.100** +/-3% gms/210L Breath when heated to 34 Degrees Celsius +/-0.2 Degrees Celsius in a simulator (95% Confidence).

The date of manufacture for this lot number is **March 8, 2013** The expiration date for this lot number is **March 7, 2015** at 11:59 p.m.

This document is a true representation of the original Certificate of Analysis.

Cecil B. Garner, President
RepCo Marketing, Inc.

Form RM 02

State of Missouri
DEPARTMENT OF HEALTH

PERMIT
TYPE II

REID T FESSLER

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field repairs, and operate the following breath analyzer(s):

DATAMASTER

for the determination of the alcoholic content of blood from a sample of expired (alveolar) air. Issued under the provisions of sections 577.020 through 577.041, RSMo 1986.

Date 02/03/2012

Number **220016**

Expires 02/03/2014

Director of State Public Health Laboratory

Director, Department of Health

MO 580-0771 (7-88)

Lab. 4 (R7-88)

Face This Side Down - This Edge In First

BAC DataMaster
Evidence Ticket

gnature

Face This Side Down - This Edge In First

BAC DataMaster
Evidence Ticket

Operator Signature

Face This Side Down - This Edge In First

BAC DataMaster
Evidence Ticket

Operator Signature