



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
STATE PUBLIC HEALTH LABORATORY
BREATH ALCOHOL PROGRAM

DATAMASTER MAINTENANCE REPORT

REPORT #6

RECEIVED
By Carol Day at 9:29 am, May 12, 2014

Complete this report at the time of the regular monthly preventive maintenance check (not to exceed 35 days).
Complete this report whenever the instrument is serviced or repaired and whenever it is placed into service.
Retain the original and send a copy within 15 days to the Breath Alcohol Program, DHSS.

DATAMASTER SN 204101	NAME OF AGENCY MISSOURI STATE HIGHWAY PATROL	DATE OF INSPECTION 04/28/2014
LOCATION OF INSTRUMENT (STREET AND CITY) PUTNAM COUNTY SHERIFF'S OFFICE, 1601 MAIN, UNIONVILLE, MO		TIME OF INSPECTION 10:00 am

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

<input checked="" type="checkbox"/> DIAGNOSTIC CHECK (PRINTOUT ATTACHED)	DATE AND TIME (from printout) <u>04/28/2014 10000 HOURS</u>
<input checked="" type="checkbox"/> COMPUTER	<input checked="" type="checkbox"/> DETECTOR
<input checked="" type="checkbox"/> PROGRAM	<input checked="" type="checkbox"/> FILTERS
<input checked="" type="checkbox"/> HEATERS SAMPLE CHAMBER <u>50</u> °C	<input checked="" type="checkbox"/> QUARTZ STANDARD
<input checked="" type="checkbox"/> FLOW DETECTOR	<input checked="" type="checkbox"/> CALIBRATION
<input checked="" type="checkbox"/> PUMP HIGH SPEED	<input checked="" type="checkbox"/> PRINTER

INDICATOR LIGHTS

SIMULATOR SOLUTION SUPPLIER GUTH LABORATORIES, INC LOT # 13280 EXP. DATE 10/16/2015

SIMULATOR TEMP (34°C ± 0.2°C) 34 °C SIMULATOR SN G11077 EXP. DATE 10/28/2014

CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Mark the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

0.100% STANDARD - MUST READ BETWEEN 0.095% AND 0.105% INCLUSIVE
 0.080% STANDARD - MUST READ BETWEEN 0.076% AND 0.084% INCLUSIVE
 0.040% STANDARD - MUST READ BETWEEN 0.038% AND 0.042% INCLUSIVE

TEST 1 • .098	TEST 2 • .100	TEST 3 • .100
---------------	---------------	---------------

PERFORM R.F.I. TEST (PRINTOUT ATTACHED)

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:
(DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS	0	(0-.04)	0	(.05-.09)	0	(.10-.14)	0	(.15-.19)	0	OVER .19	0
----------	---	---------	---	-----------	---	-----------	---	-----------	---	----------	---

LIST ANY NEW PARTS AND DESCRIBE ANY ALTERATION OR MODIFICATION THAT WAS MADE TO RESTORE THE INSTRUMENT TO OPERATE SATISFACTORILY AND WITHIN ESTABLISHED LIMITS (USE OTHER SIDE IF NECESSARY).

INSPECTING OFFICER	
SIGNATURE <i>Reid T Fessler</i>	PRINT FULL NAME REID T FESSLER
TYPE II PERMIT NUMBER/EXPIRATION DATE 240018 01/28/2016	TELEPHONE NUMBER (660) 385-2132

RETURN COMPLETED REPORT TO THE: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
2875 James Blvd.
Poplar Bluff, MO 63901



GUTH LABORATORIES, INC.

690 NORTH 67th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-564-5470

CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number 13280 of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on October 18, 2013, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain 0.1217% (w/vol) ethyl alcohol. The expiration date for this lot number is October 16, 2015 at 11:59 PM.

When used in a calibrated Simulator, operating at 34°C +/- .2°C, this solution will give a breath alcohol analysis instrument reading of 0.100 g/210L +/- 3%.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President
GUTH LABORATORIES, INC.

NIST Traceability:

Testing was conducted using Cerilliant Reference Standard lot number FN122211-02 whose values are traceable to NIST.

All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

2

PERMIT
TYPE II
REID T FESSLER

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

DATAMASTER

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 1/28/2014

NUMBER 240018

EXPIRES 1/28/2016

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MO 580-0771 (6-10)

LAB-4 (R6-10)

STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator FESSLER, REID
Permit No 240018
Date Issued 1/28/2014 Date Expires 1/28/2016

Face This Side Down - This Edge In First

BAC DataMaster Evidence Ticket

PLISSIMUM

WWW. UNIFORMS...
...
...

...

COPY...

PRELIMINARY...

MEMPHIS...
SPECIAL...

HOW...

TYPE...
MAY...

DELETED...

FILE...

...

...

...

...

...

Signature

Face This Side Down - This Edge In First

BAC DataMaster Evidence Ticket

PLISSIMUM

WWW. UNIFORMS...
...
...

...

COPY...

PRELIMINARY...

MEMPHIS...
SPECIAL...

HOW...

TYPE...
MAY...

DELETED...

FILE...

...

...

...

...

...

...

Operator Signature

Face This Side Down - This Edge In First

BAC DataMaster Evidence Ticket

MISSOURI COURT REPORTERS ASSOCIATION

3000 DATAPoint Building, Suite 100
St. Louis, MO 63103

Client Name: MISSOURI COURT REPORTERS ASSOCIATION
Case Name: MISSOURI COURT REPORTERS ASSOCIATION
Case No: MISSOURI COURT REPORTERS ASSOCIATION
Case Description: MISSOURI COURT REPORTERS ASSOCIATION
Case Address: MISSOURI COURT REPORTERS ASSOCIATION
Case City: MISSOURI COURT REPORTERS ASSOCIATION
Case State: MISSOURI COURT REPORTERS ASSOCIATION
Case Zip: MISSOURI COURT REPORTERS ASSOCIATION

Case No: MISSOURI COURT REPORTERS ASSOCIATION

Case No:	MISSOURI COURT REPORTERS ASSOCIATION	3000 DATAPoint Building, Suite 100
Case Name:	MISSOURI COURT REPORTERS ASSOCIATION	St. Louis, MO 63103
Case Address:	MISSOURI COURT REPORTERS ASSOCIATION	St. Louis, MO 63103
Case City:	MISSOURI COURT REPORTERS ASSOCIATION	St. Louis, MO 63103
Case State:	MISSOURI COURT REPORTERS ASSOCIATION	St. Louis, MO 63103
Case Zip:	MISSOURI COURT REPORTERS ASSOCIATION	St. Louis, MO 63103

Case No: MISSOURI COURT REPORTERS ASSOCIATION
Case Name: MISSOURI COURT REPORTERS ASSOCIATION
Case Address: MISSOURI COURT REPORTERS ASSOCIATION

Operator Signature

