



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
 STATE PUBLIC HEALTH LABORATORY
 BREATH ALCOHOL PROGRAM
DATAMASTER MAINTENANCE REPORT

RECEIVED 1/28/14-REPORT #6

Complete this report at the time of the regular monthly preventive maintenance check (not to exceed 35 days).
 Complete this report whenever the instrument is serviced or repaired and whenever it is placed into service.
 Retain the original and send a copy within 15 days to the Breath Alcohol Program, DHSS.

REVIEWED
 By Carol Day at 10:05 am, Mar 14, 2014

DATAMASTER SN 204105	NAME OF AGENCY MISSOURI STATE HIGHWAY PATROL	DATE OF INSPECTION 01/21/2014
LOCATION OF INSTRUMENT (STREET AND CITY) SULLIVAN COUNTY SHERIFF'S OFFICE, 109 N. MAIN, MILAN, MO		TIME OF INSPECTION 8:32 am

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

<input checked="" type="checkbox"/> DIAGNOSTIC CHECK (PRINTOUT ATTACHED)	DATE AND TIME (from printout) 01/21/2014 0832 HOURS
<input checked="" type="checkbox"/> COMPUTER	<input checked="" type="checkbox"/> DETECTOR
<input checked="" type="checkbox"/> PROGRAM	<input checked="" type="checkbox"/> FILTERS
<input checked="" type="checkbox"/> HEATERS SAMPLE CHAMBER 49°C	<input checked="" type="checkbox"/> QUARTZ STANDARD
<input checked="" type="checkbox"/> FLOW DETECTOR	<input checked="" type="checkbox"/> CALIBRATION
<input checked="" type="checkbox"/> PUMP HIGH SPEED	<input checked="" type="checkbox"/> PRINTER

INDICATOR LIGHTS

SIMULATOR SOLUTION SUPPLIER REPCO MARKETING, INC LOT # 13001 EXP. DATE 03/07/2015

SIMULATOR TEMP (34°C ± 0.2°C) 34 °C SIMULATOR SN G11077 EXP. DATE 10/28/2014

CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Mark the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

0.100% STANDARD - MUST READ BETWEEN 0.095% AND 0.105% INCLUSIVE
 0.080% STANDARD - MUST READ BETWEEN 0.076% AND 0.084% INCLUSIVE
 0.040% STANDARD - MUST READ BETWEEN 0.038% AND 0.042% INCLUSIVE

TEST 1 .101	TEST 2 .100	TEST 3 .099
-------------	-------------	-------------

PERFORM R.F.I. TEST (PRINTOUT ATTACHED)

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:
 (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS	0	(0-.04)	1	(.05-.09)	1	(.10-.14)	1	(.15-.19)	0	OVER .19	0
----------	---	---------	---	-----------	---	-----------	---	-----------	---	----------	---

LIST ANY NEW PARTS AND DESCRIBE ANY ALTERATION OR MODIFICATION THAT WAS MADE TO RESTORE THE INSTRUMENT TO OPERATE SATISFACTORILY AND WITHIN ESTABLISHED LIMITS (USE OTHER SIDE IF NECESSARY).

MEETS DEPARTMENT OF HEALTH AND SENIOR SERVICES STANDARDS

INSPECTING OFFICER	
SIGNATURE 	PRINT FULL NAME REID T. FESSLER
TYPE II PERMIT NUMBER/EXPIRATION DATE 220016 02/03/2014	TELEPHONE NUMBER (660) 385-2132

RETURN COMPLETED REPORT TO THE: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
 2875 James Blvd.
 Poplar Bluff, MO 63901

CERTIFICATE OF ANALYSIS

MANUFACTURER AND SUPPLIER: RepCo Marketing, Inc.

LOT NUMBER: 13001

EXPIRATION DATE: March 7, 2015 at 11:59 p.m.

RepCo Marketing, Inc. certifies the following:


RepCo Marketing, Inc. manufactured, tested and supplied Lot Number 13001 of Alcohol Certified Solution for simulators. Random samples of said lot number were analyzed by an independent laboratory utilizing a gas chromatograph and found to contain .1215 gms/dl +/-0.003 gms/dl wt/vol ethanol (95% Confidence).

The alcohol and distilled water used in the solution were found to be free of any interfering substance.

This solution will produce a vapor alcohol value of .100 +/-3% gms/210L Breath when heated to 34 Degrees Celsius +/-0.2 Degrees Celsius in a simulator (95% Confidence).

The date of manufacture for this lot number is March 8, 2013
The expiration date for this lot number is March 7, 2015 at
11:59 p.m.

This document is a true representation of the original Certificate of Analysis.



Cecil B. Garner, President
RepCo Marketing, Inc.

State of MISSOURI
DEPARTMENT OF HEALTH



P E R M I T
TYPE II



REID T FESSLER

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field repairs, and operate the following breath analyzer(s):

DATAMASTER

for the determination of the alcoholic content of blood from a sample of expired (alveolar) air. Issued under the provisions of sections 577.020 through 577.041, RSMo 1986.

Date 02/03/2012

Number 220016

Expires 02/03/2014

MO 580-0771 (7-88)

Director of State Public Health Laboratory

Director, Department of Health

Lab. 4 (R7-88)

Face This Side Down - This Edge In First

BAC DataMaster Evidence Ticket

MISSOURI STATE HIGHWAY PATROL

BAC DATAMASTER SERIAL NUMBER 204105
01/21/14

TESTING OFFICER:

FESSLER/R/T

OFFICER I.D.: 992

PERMIT NUMBER: 220016

EXPIRATION DATE: 02/03/14

MISCELLANEOUS DATA:

.10 TEST

--- SUPERVISOR MODE ---

BLANK TEST	.000	08:35
INTERNAL STANDARD	VERIFIED	08:35
EXTERNAL STANDARD	.101	08:39
BLANK TEST	.000	08:39
EXTERNAL STANDARD	.100	08:41
BLANK TEST	.000	08:41
EXTERNAL STANDARD	.099	08:42
BLANK TEST	.000	08:43

n = 3
S.M. = .1
AVG. = .1

Operator Signature

Face This Side Down - This Edge In First

BAC DataMaster Evidence Ticket

MISSOURI STATE HIGHWAY PATROL

BAC DATAMASTER SERIAL NUMBER 204105
01/21/14

ARREST TIME: 08:00

SUBJECT NAME:

TEST

DOB: 11/11/11 SEX: M

STATE/D.L.L.: MO/TEST

ARRESTING OFFICER:

FESSLER/R/T

OFFICER I.D.: 992

TESTING OFFICER:

FESSLER/R/T

OFFICER I.D.: 992

PERMIT NUMBER: 220016

EXPIRATION DATE: 02/03/14

MISCELLANEOUS DATA:

--- BREATH ANALYSIS ---

RADIO INTERFERENCE

Operator Signature

Face This Side Down - This Edge In First

BAC DataMaster Evidence Ticket

MISSOURI STATE HIGHWAY PATROL

BAC DATAMASTER SERIAL NUMBER 284105

01/21/14

08:32

--- DIAGNOSTIC CHECK ---

COMPUTER:	OKAY
PROBAP (04-07-2009):	OKAY
HEATERS:	
SAMPLE CHAMBER:	49c
FLOW DETECTOR:	OKAY
PUMP	
HIGH SPEED:	OKAY
DETECTOR:	OKAY
FILTERS:	OKAY
SURFIZ STANDARD:	OKAY
CALIBRATION:	OKAY

PRINTER TEST

!@#\$%^&*()~+,-./0123456789:;<=>?@A[B]C[D]E[F]G
H[I]J[K]L[M]N[O]P[Q]R[S]T[U]V[W]X[Y]Z[\]^_`{|}~!@#\$%^&*()~+,-./0123456789:;<=>?@A[B]C[D]E[F]G
H[I]J[K]L[M]N[O]P[Q]R[S]T[U]V[W]X[Y]Z[\]^_`{|}~!@#\$%^&*()~+,-./0123456789:;<=>?@A[B]C[D]E[F]G

Operator Signature

